

7470A


LB94018

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36594



A= 0.0000e+000

B= 2.7427e-004

C= 3.8692e-003

Rho= 0.9999514

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.051	0.051	172	0.000	172				
0.2	0.200	0.206	0.006	736	0.0 %	736				
2.5	2.500	2.450	-0.050	8917	0.0 %	8917				
5.0	5.000	4.975	-0.025	18124	0.0 %	18124				
7.5	7.500	7.478	-0.022	27252	0.0 %	27252				
10.0	10.000	10.041	0.041	36594	0.0 %	36594				

LB94018 INSTRUMENT ID :CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Mar 2018 10:23:14

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	172	-	0.0000	7470A	PPB	20 Mar 2018 10:30:28	S	Std
0.2 - 1	S0.2	736	-	0.2000	7470A	PPB	20 Mar 2018 10:32:36	S	Std
2.5 - 1	S2.5	8917	-	2.5000	7470A	PPB	20 Mar 2018 10:34:44	S	Std
5.0 - 1	S5	18124	-	5.0000	7470A	PPB	20 Mar 2018 10:36:53	S	Std
7.5 - 1	S7.5	27252	-	7.5000	7470A	PPB	20 Mar 2018 10:39:01	S	Std
10.0 - 1	S10	36594	-	10.0000	7470A	PPB	20 Mar 2018 10:41:09	S	Std
ICV54 - 1	ICV54	13654	3.7488	-	7470A	PPB	20 Mar 2018 10:43:56	U	SMPL
ICB54 - 1	ICB54	50	0.0176	-	7470A	PPB	20 Mar 2018 10:46:02	U	SMPL
CCV05 - 1	CCV05	18803	5.1610	-	7470A	PPB	20 Mar 2018 10:48:09	U	SMPL
CCB05 - 1	CCB05	51	0.0179	-	7470A	PPB	20 Mar 2018 10:50:16	U	SMPL
CRA - 1	CRA	723	0.2022	-	7470A	PPB	20 Mar 2018 10:52:23	U	SMPL
HighStd - 1	HighStd	37080	10.1739	-	7470A	PPB	20 Mar 2018 10:54:30	U	SMPL
ChkStd - 1	ChkStd	24829	6.8138	-	7470A	PPB	20 Mar 2018 10:56:37	U	SMPL
PB107455BL - 1	PBW	-241	-0.0622	-	7470A	PPB	20 Mar 2018 10:58:46	U	SMPL
PB107455BS - 1	LCSW	13265	3.6421	-	7470A	PPB	20 Mar 2018 11:00:54	U	SMPL
PB107455TB - 1	PB107434TB	-107	-0.0255	-	7470A	PPB	20 Mar 2018 11:03:02	U	SMPL
J1755-04 - 1	OK-01-031618	-50	-0.0098	-	7470A	PPB	20 Mar 2018 11:05:10	U	SMPL
J1755-04DUP - 1	OK-01-031618DUP	-38	-0.0066	-	7470A	PPB	20 Mar 2018 11:07:19	U	SMPL
J1755-04MS - 1	OK-01-031618MS	13208	3.6265	-	7470A	PPB	20 Mar 2018 11:09:27	U	SMPL
J1755-04MSD - 1	OK-01-031618MSD	12867	3.5329	-	7470A	PPB	20 Mar 2018 11:11:34	U	SMPL
CCV06 - 1	CCV06	18834	5.1695	-	7470A	PPB	20 Mar 2018 11:13:40	U	SMPL
CCB06 - 1	CCB06	54	0.0187	-	7470A	PPB	20 Mar 2018 11:15:47	U	SMPL
J1963-04 - 1	1-4-RO	-63	-0.0134	-	7470A	PPB	20 Mar 2018 11:17:54	U	SMPL
PB107456BL - 1	PBW	-40	-0.0071	-	7470A	PPB	20 Mar 2018 11:20:01	U	SMPL
PB107456BS - 1	LCSW	13355	3.6668	-	7470A	PPB	20 Mar 2018 11:22:07	U	SMPL
PB107456TB - 1	PB107436TB	-76	-0.0170	-	7470A	PPB	20 Mar 2018 11:24:14	U	SMPL
J1960-02 - 1	BB	-10	0.0011	-	7470A	PPB	20 Mar 2018 11:26:22	U	SMPL
J1960-02DUP - 1	BBDUP	-9	0.0014	-	7470A	PPB	20 Mar 2018 11:28:30	U	SMPL
J1960-02MS - 1	BBMS	13032	3.5782	-	7470A	PPB	20 Mar 2018 11:30:38	U	SMPL
J1960-02MSD - 1	BBMSD	13038	3.5798	-	7470A	PPB	20 Mar 2018 11:32:46	U	SMPL
J1966-08 - 1	FRAC-1	-117	-0.0282	-	7470A	PPB	20 Mar 2018 11:34:54	U	SMPL
J1966-10 - 1	FRAC-2	-49	-0.0096	-	7470A	PPB	20 Mar 2018 11:37:02	U	SMPL
CCV07 - 1	CCV07	18822	5.1662	-	7470A	PPB	20 Mar 2018 11:39:09	U	SMPL
CCB07 - 1	CCB07	64	0.0214	-	7470A	PPB	20 Mar 2018 11:41:16	U	SMPL
PB107457BL - 1	PBW	-75	-0.0167	-	7470A	PPB	20 Mar 2018 11:43:25	U	SMPL
PB107457BS - 1	LCSW	13239	3.6350	-	7470A	PPB	20 Mar 2018 11:45:32	U	SMPL
J1960-01 - 1	BB	18163	4.9855	-	7470A	PPB	20 Mar 2018 11:47:39	U	SMPL
J1966-07 - 1	FRAC-1	623	0.1747	-	7470A	PPB	20 Mar 2018 11:49:46	U	SMPL
J1966-09 - 1	FRAC-2	337	0.0963	-	7470A	PPB	20 Mar 2018 11:51:55	U	SMPL
J1966-09DUP - 1	FRAC-2DUP	355	0.1012	-	7470A	PPB	20 Mar 2018 11:54:03	U	SMPL
J1966-09MS - 1	FRAC-2MS	13671	3.7534	-	7470A	PPB	20 Mar 2018 11:56:10	U	SMPL
J1966-09MSD - 1	FRAC-2MSD	14322	3.9320	-	7470A	PPB	20 Mar 2018 11:58:19	U	SMPL
CCV08 - 1	CCV08	18464	5.0680	-	7470A	PPB	20 Mar 2018 12:00:27	U	SMPL
CCB08 - 1	CCB08	-34	-0.0055	-	7470A	PPB	20 Mar 2018 12:02:34	U	SMPL
J1755-04LX5 - 1		50	0.0176	-	7470A	PPB	20 Mar 2018 12:04:42	U	SMPL
J1755-04A - 1		13488	3.7032	-	7470A	PPB	20 Mar 2018 12:06:49	U	SMPL
J1960-02LX5 - 1		-12	0.0006	-	7470A	PPB	20 Mar 2018 12:08:56	U	SMPL
J1960-02A - 1		13305	3.6531	-	7470A	PPB	20 Mar 2018 12:11:04	U	SMPL
J1966-09LX5 - 1		59	0.0201	-	7470A	PPB	20 Mar 2018 12:13:11	U	SMPL
J1966-09A - 1		14311	3.9290	-	7470A	PPB	20 Mar 2018 12:15:18	U	SMPL
CCV09 - 1	CCV09	18325	5.0299	-	7470A	PPB	20 Mar 2018 12:17:25	U	SMPL
CCB09 - 1	CCB09	-32	-0.0049	-	7470A	PPB	20 Mar 2018 12:19:33	U	SMPL