


ISM02.4

LB 94022
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45641



A= 0.0000e+000

B= 2.1974e-004

C= -4.1933e-002

Rho= 0.9999793

Accept=Accepted

slope
4.1933e-002

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/b
0.00	0.000	-0.030	-0.030	53	0.000	53					0/00
0.2	0.200	0.196	-0.004	1085	0.0 %	1085					-2
2.5	2.500	2.527	0.027	11690	0.0 %	11690					1
5.0	5.000	5.036	0.036	23109	0.0 %	23109					1
7.5	7.500	7.484	-0.016	34248	0.0 %	34248					8
10.0	10.000	9.987	-0.013	45641	0.0 %	45641					

263
3/20/18

LB94022
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	53	-		0 ISM02.4	PPB	3/20/2018 12:35	Std
	0.2 S0.2	1085	-		0.2 ISM02.4	PPB	3/20/2018 12:37	Std
	2.5 S2.5	11690	-		2.5 ISM02.4	PPB	3/20/2018 12:40	Std
	5 S5	23109	-		5 ISM02.4	PPB	3/20/2018 12:42	Std
	7.5 S7.5	34248	-		7.5 ISM02.4	PPB	3/20/2018 12:44	Std
	10 S10	45641	-		10 ISM02.4	PPB	3/20/2018 12:46	Std
ICV	ICV	17541	3.8125 -		ISM02.4	PPB	3/20/2018 12:49	SMPL
ICB	ICB	-112	-0.0665 -		ISM02.4	PPB	3/20/2018 12:52	SMPL
CCV010	CCV010	23155	5.0461 -		ISM02.4	PPB	3/20/2018 12:54	SMPL
CCB010	CCB010	-137	-0.072 -		ISM02.4	PPB	3/20/2018 12:56	SMPL
PB107486BL	PBS004	42	-0.0327 -		ISM02.4	PPB	3/20/2018 12:58	SMPL
J1961-01	MC1005	797	0.1332 -		ISM02.4	PPB	3/20/2018 13:01	SMPL
J1961-02	MC1006	748	0.1224 -		ISM02.4	PPB	3/20/2018 13:03	SMPL
J1961-03	MC1007	341	0.033 -		ISM02.4	PPB	3/20/2018 13:05	SMPL
J1961-04	MC1008	392	0.0442 -		ISM02.4	PPB	3/20/2018 13:07	SMPL
J1961-05	MC1010	152	-0.0085 -		ISM02.4	PPB	3/20/2018 13:10	SMPL
J1961-06	MC1011	203	0.0027 -		ISM02.4	PPB	3/20/2018 13:12	SMPL
J1961-07	MC1012	375	0.0405 -		ISM02.4	PPB	3/20/2018 13:14	SMPL
J1961-08	MC1033	826	0.1396 -		ISM02.4	PPB	3/20/2018 13:17	SMPL
J1961-09	MC1034	465	0.0602 -		ISM02.4	PPB	3/20/2018 13:19	SMPL
J1961-12	MC1001	3501	0.7274 -		ISM02.4	PPB	3/20/2018 13:21	SMPL
J1961-13	MC1002	1143	0.2092 -		ISM02.4	PPB	3/20/2018 13:23	SMPL
J1961-14	MC1002D	1145	0.2097 -		ISM02.4	PPB	3/20/2018 13:26	SMPL
J1961-15	MC1002S	12383	2.6791 -		ISM02.4	PPB	3/20/2018 13:28	SMPL
J1961-16	MC1004	860	0.147 -		ISM02.4	PPB	3/20/2018 13:30	SMPL
PB107487BL	PBS004	107	-0.0184 -		ISM02.4	PPB	3/20/2018 13:32	SMPL
J1968-02	MC0AA2	601	0.0901 -		ISM02.4	PPB	3/20/2018 13:35	SMPL
J1968-03	MC0AA3	767	0.1266 -		ISM02.4	PPB	3/20/2018 13:37	SMPL
J1968-04	MC0AA4	332	0.031 -		ISM02.4	PPB	3/20/2018 13:39	SMPL
CCV011	CCV011	23099	5.0338 -		ISM02.4	PPB	3/20/2018 13:42	SMPL
CCB011	CCB011	-124	-0.0692 -		ISM02.4	PPB	3/20/2018 13:44	SMPL
J1968-05	MC0AA5	1091	0.1978 -		ISM02.4	PPB	3/20/2018 13:46	SMPL
J1968-06	MC0AA5D	1103	0.2004 -		ISM02.4	PPB	3/20/2018 13:54	SMPL
J1968-07	MC0AA5S	12161	2.6303 -		ISM02.4	PPB	3/20/2018 13:56	SMPL
J1968-08	MC0AA6	1791	0.3516 -		ISM02.4	PPB	3/20/2018 13:58	SMPL
J1968-09	MC0AA7	916	0.1593 -		ISM02.4	PPB	3/20/2018 14:00	SMPL
J1968-10	MC0AA8	437	0.0541 -		ISM02.4	PPB	3/20/2018 14:03	SMPL
J1968-11	MC0AA9	456	0.0583 -		ISM02.4	PPB	3/20/2018 14:05	SMPL
J1968-12	MC0AB0	619	0.0941 -		ISM02.4	PPB	3/20/2018 14:07	SMPL
J1968-13	MC0AC2	327	0.0299 -		ISM02.4	PPB	3/20/2018 14:10	SMPL
CCV012	CCV012	23040	5.0209 -		ISM02.4	PPB	3/20/2018 14:12	SMPL
CCB012	CCB012	-115	-0.0672 -		ISM02.4	PPB	3/20/2018 14:14	SMPL