


7471B

LB94034
INSTRUMENT ID: CV-1

Linear

 μ Abs.:

42445



A= 0.0000e+000

B= 2.3426e-004

C= 1.8248e-002

Rho= 0.9998934

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.020	0.020	9	0.000	9				
0.2	0.200	0.221	0.021	866	0.0 %	866				
2.5	2.500	2.448	-0.052	10372	0.0 %	10372				
5.0	5.000	4.950	-0.050	21051	0.0 %	21051				
7.5	7.500	7.599	0.099	32361	0.0 %	32361				
10	10.000	9.962	-0.038	42445	0.0 %	42445				

LB94034 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Mar 2018 14:36:00

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	SO	9	-	0.0000	7471B	PPB	20 Mar 2018 14:42:16	S	Std
0.2 - 1	SO-2	866	-	0.2000	7471B	PPB	20 Mar 2018 14:44:33	S	Std
2.5 - 1	SO-2.5	10372	-	2.5000	7471B	PPB	20 Mar 2018 14:46:51	S	Std
5.0 - 1	SO-5	21051	-	5.0000	7471B	PPB	20 Mar 2018 14:49:08	S	Std
7.5 - 1	SO-7.5	32361	-	7.5000	7471B	PPB	20 Mar 2018 14:51:25	S	Std
10 - 1	SO-10	42445	-	10.0000	7471B	PPB	20 Mar 2018 14:53:42	S	Std
ICV55 - 1	ICV55	16114	3.7932	-	7471B	PPB	20 Mar 2018 14:56:26	U	SMPL
ICB55 - 1	ICB55	-98	-0.0047	-	7471B	PPB	20 Mar 2018 14:58:41	U	SMPL
CCV13 - 1	CCV13	21315	5.0116	-	7471B	PPB	20 Mar 2018 15:00:57	U	SMPL
CCB13 - 1	CCB13	-115	-0.0087	-	7471B	PPB	20 Mar 2018 15:03:13	U	SMPL
CRA - 1	CRA	876	0.2235	-	7471B	PPB	20 Mar 2018 15:05:29	U	SMPL
HighStd - 1	HighStd	43009	10.0937	-	7471B	PPB	20 Mar 2018 15:07:45	U	SMPL
ChkStd - 1	ChkStd	29867	7.0150	-	7471B	PPB	20 Mar 2018 15:10:01	U	SMPL
PB107485BL - 1	PBS	-152	-0.0174	-	7471B	PPB	20 Mar 2018 15:12:19	U	SMPL
PB107485BS - 1	LCSS	15499	3.6491	-	7471B	PPB	20 Mar 2018 15:14:38	U	SMPL
J1760-03 - 1	SU-03-031918	732	0.1897	-	7471B	PPB	20 Mar 2018 15:16:54	U	SMPL
J1971-01 - 1	ES-6-SA-1-WW@2	2773	0.6679	-	7471B	PPB	20 Mar 2018 15:19:12	U	SMPL
J1971-02 - 1	ES-6-SA-1-EW@2	1979	0.4819	-	7471B	PPB	20 Mar 2018 15:21:29	U	SMPL
J1971-03 - 1	ES-6-SA-2-WW@3	2126	0.5163	-	7471B	PPB	20 Mar 2018 15:23:47	U	SMPL
J1971-04 - 1	ES-6-SA-2-EW@3	2451	0.5924	-	7471B	PPB	20 Mar 2018 15:26:03	U	SMPL
CCV14 - 1	CCV14	21164	4.9762	-	7471B	PPB	20 Mar 2018 15:28:19	U	SMPL
CCB14 - 1	CCB14	-94	-0.0038	-	7471B	PPB	20 Mar 2018 15:30:35	U	SMPL
J1971-05 - 1	ES-6-SA-3-WW@4.5	1782	0.4357	-	7471B	PPB	20 Mar 2018 15:32:51	U	SMPL
J1971-06 - 1	ES-6-SA-3-EW@4.5	1150	0.2877	-	7471B	PPB	20 Mar 2018 15:35:06	U	SMPL
J1971-07 - 1	WEST-GARAGE-6-ST-LATERAL-NW@4	3295	0.7901	-	7471B	PPB	20 Mar 2018 15:37:22	U	SMPL
J1971-08 - 1	WEST-GARAGE-6-ST-LATERAL-SW@4	2533	0.6116	-	7471B	PPB	20 Mar 2018 15:39:38	U	SMPL
J1971-09 - 1	WEST-GARAGE-4-SA-LATERAL-NW@3	1518	0.3739	-	7471B	PPB	20 Mar 2018 15:41:55	U	SMPL
J1971-10 - 1	WEST-GARAGE-4-SA-LATERAL-SW@3	2828	0.6807	-	7471B	PPB	20 Mar 2018 15:44:12	U	SMPL
J1972-02 - 1	S-1 COMP	8381	1.9816	-	7471B	PPB	20 Mar 2018 15:46:29	U	SMPL
J1972-04 - 1	S-2 COMP	19755	4.6461	-	7471B	PPB	20 Mar 2018 15:48:47	U	SMPL
J1972-06 - 1	S-3 COMP	21973	5.1657	-	7471B	PPB	20 Mar 2018 15:51:04	U	SMPL
J1972-08 - 1	S-4 COMP	12930	3.0473	-	7471B	PPB	20 Mar 2018 15:53:20	U	SMPL
CCV15 - 1	CCV15	21253	4.9970	-	7471B	PPB	20 Mar 2018 15:55:38	U	SMPL
CCB15 - 1	CCB15	-89	-0.0026	-	7471B	PPB	20 Mar 2018 15:57:54	U	SMPL
J1973-01 - 1	COMP-1	9333	2.2046	-	7471B	PPB	20 Mar 2018 16:00:10	U	SMPL
J1973-01DUP - 1	COMP-1DUP	9906	2.3389	-	7471B	PPB	20 Mar 2018 16:02:27	U	SMPL
J1973-01MS - 1	COMP-1MS	24796	5.8270	-	7471B	PPB	20 Mar 2018 16:04:43	U	SMPL
J1973-01MSD - 1	COMP-1MSD	23661	5.5611	-	7471B	PPB	20 Mar 2018 16:07:00	U	SMPL
J1973-02 - 1	COMP-2	129665	30.3940	-	7471B	PPB	20 Mar 2018 16:09:17	U	SMPL
J1973-03 - 1	COMP-3	14401	3.3919	-	7471B	PPB	20 Mar 2018 16:11:35	U	SMPL
PB107488BL - 1	PBS	-125	-0.0110	-	7471B	PPB	20 Mar 2018 16:13:57	U	SMPL
PB107488BS - 1	PB107488BS	14054	3.3106	-	7471B	PPB	20 Mar 2018 16:16:15	U	SMPL
J1974-01 - 1	LCSS	1338	0.3317	-	7471B	PPB	20 Mar 2018 16:18:33	U	SMPL
J1974-01DUP - 1	YR18020063DUP	1233	0.3071	-	7471B	PPB	20 Mar 2018 16:20:49	U	SMPL
CCV16 - 1	CCV16	21664	5.0933	-	7471B	PPB	20 Mar 2018 16:23:05	U	SMPL
CCB16 - 1	CCB16	-115	-0.0087	-	7471B	PPB	20 Mar 2018 16:25:21	U	SMPL
J1974-01MS - 1	YR18020063MS	17134	4.0321	-	7471B	PPB	20 Mar 2018 16:27:37	U	SMPL
J1974-01MSD - 1	YR18020063MSD	16705	3.9316	-	7471B	PPB	20 Mar 2018 16:29:53	U	SMPL
J1975-05 - 1	FS-SOIL	893	0.2274	-	7471B	PPB	20 Mar 2018 16:32:09	U	SMPL
J1977-02 - 1	SP	6971	1.6513	-	7471B	PPB	20 Mar 2018 16:34:25	U	SMPL
J1979-01 - 1	SP-1	29333	6.8899	-	7471B	PPB	20 Mar 2018 16:36:42	U	SMPL
J1981-01 - 1	TILCON-MT-HOPE-1	-33	0.0105	-	7471B	PPB	20 Mar 2018 16:38:59	U	SMPL
J1981-02 - 1	TILCON-MT-HOPE-2	112	0.0445	-	7471B	PPB	20 Mar 2018 16:41:17	U	SMPL
J1983-01 - 1	EPE-L-4DS(10-12)	343	0.0986	-	7471B	PPB	20 Mar 2018 16:43:35	U	SMPL
J1983-02 - 1	EPE-L-4DS(16-18)	453	0.1244	-	7471B	PPB	20 Mar 2018 16:45:53	U	SMPL
CCV17 - 1	CCV17	21248	4.9959	-	7471B	PPB	20 Mar 2018 16:48:09	U	SMPL
CCB17 - 1	CCB17	-108	-0.0071	-	7471B	PPB	20 Mar 2018 16:50:25	U	SMPL
J1973-01LX5 - 1		1762	0.4310	-	7471B	PPB	20 Mar 2018 16:52:42	U	SMPL
J1973-01A - 1		26421	6.2077	-	7471B	PPB	20 Mar 2018 16:54:58	U	SMPL
J1974-01LX5 - 1		101	0.0419	-	7471B	PPB	20 Mar 2018 16:57:14	U	SMPL
J1974-01A - 1		17840	4.1975	-	7471B	PPB	20 Mar 2018 16:59:31	U	SMPL
J1973-02DLX5 - 1		25649	6.0269	-	7471B	PPB	20 Mar 2018 17:01:47	U	SMPL
CCV18 - 1	CCV18	21412	5.0343	-	7471B	PPB	20 Mar 2018 17:04:04	U	SMPL
CCB18 - 1	CCB18	-98	-0.0047	-	7471B	PPB	20 Mar 2018 17:06:22	U	SMPL