

245.1


LB94038

INSTRUMENT ID: CV1

Linear

 μ Abs.:

33863



A= 0.0000e+000

B= 2.9052e-004

C= 1.3600e-003

Rho= 0.9996201

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.007	-0.007	-30	0.000	-30				
0.05	0.050	0.053	0.003	179	0.0 %	179				
0.2	0.200	0.188	-0.012	644	0.0 %	644				
2.5	2.500	2.407	-0.093	8279	0.0 %	8279				
5.0	5.000	5.101	0.101	17554	0.0 %	17554				
7.5	7.500	7.668	0.168	26390	0.0 %	26390				
10.0	10.000	9.839	-0.161	33863	0.0 %	33863				

LB94038 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 20 Mar 2018 17:25:11

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-30	-	0.0000	245.1	PPB	20 Mar 2018 17:27:07	S	Std
0.05 - 1	50.05	179	-	0.0500	245.1	PPB	20 Mar 2018 17:29:15	S	Std
0.2 - 1	50.2	644	-	0.2000	245.1	PPB	20 Mar 2018 17:31:23	S	Std
2.5 - 1	52.5	8279	-	2.5000	245.1	PPB	20 Mar 2018 17:33:32	S	Std
5.0 - 1	55	17554	-	5.0000	245.1	PPB	20 Mar 2018 17:35:40	S	Std
7.5 - 1	57.5	26390	-	7.5000	245.1	PPB	20 Mar 2018 17:37:47	S	Std
10.0 - 1	510	33863	-	10.0000	245.1	PPB	20 Mar 2018 17:45:47	S	Std
ICV56 - 1	ICV56	13428	3.9025	-245.1		PPB	20 Mar 2018 17:48:28	U	SMPL
ICB56 - 1	ICB56	-140	-0.0393	-245.1		PPB	20 Mar 2018 17:50:35	U	SMPL
CCV19 - 1	CCV19	17023	4.9470	-245.1		PPB	20 Mar 2018 17:52:41	U	SMPL
CCB19 - 1	CCB19	-162	-0.0457	-245.1		PPB	20 Mar 2018 17:54:48	U	SMPL
CRA - 1	CRA	176	0.0525	-245.1		PPB	20 Mar 2018 17:59:33	U	SMPL
HighStd - 1	HighStd	33559	9.7511	-245.1		PPB	20 Mar 2018 18:01:39	U	SMPL
ChkStd - 1	ChkStd	24425	7.0974	-245.1		PPB	20 Mar 2018 18:03:45	U	SMPL
PB107460BL - 1	PBW	-43	-0.0111	-245.1		PPB	20 Mar 2018 18:08:47	U	SMPL
PB107460BS - 1	LCSW	13334	3.8752	-245.1		PPB	20 Mar 2018 18:10:53	U	SMPL
J1953-01 - 1	031418-TIDO-F-D-B	-180	-0.0509	-245.1		PPB	20 Mar 2018 18:13:01	U	SMPL
J1953-02 - 1	031418-TIDO-F-D-S	-111	-0.0309	-245.1		PPB	20 Mar 2018 18:15:09	U	SMPL
J1953-03 - 1	031418-TIDO-F-U-B	-133	-0.0373	-245.1		PPB	20 Mar 2018 18:17:17	U	SMPL
J1953-04 - 1	031418-TIDO-F-U-S	-99	-0.0274	-245.1		PPB	20 Mar 2018 18:19:26	U	SMPL
J1953-05 - 1	031418-JETT-E-D-B	-85	-0.0233	-245.1		PPB	20 Mar 2018 18:21:33	U	SMPL
CCV20 - 1	CCV20	16765	4.8720	-245.1		PPB	20 Mar 2018 18:23:39	U	SMPL
CCB20 - 1	CCB20	-155	-0.0437	-245.1		PPB	20 Mar 2018 18:25:46	U	SMPL
J1953-06 - 1	031418-JETT-E-D-M	-123	-0.0344	-245.1		PPB	20 Mar 2018 18:27:54	U	SMPL
J1953-07 - 1	031418-JETT-E-D-S	-103	-0.0286	-245.1		PPB	20 Mar 2018 18:30:00	U	SMPL
J1953-08 - 1	031418-JETT-E-U-B	-115	-0.0321	-245.1		PPB	20 Mar 2018 18:32:07	U	SMPL
J1953-09 - 1	031418-JETT-E-U-M	-93	-0.0257	-245.1		PPB	20 Mar 2018 18:34:14	U	SMPL
J1953-10 - 1	031418-JETT-E-U-S	-116	-0.0323	-245.1		PPB	20 Mar 2018 18:36:21	U	SMPL
J1954-01 - 1	031518-JETT-F-D-B	-91	-0.0251	-245.1		PPB	20 Mar 2018 18:38:29	U	SMPL
J1954-02 - 1	031518-JETT-F-D-M	-110	-0.0306	-245.1		PPB	20 Mar 2018 18:40:37	U	SMPL
J1954-03 - 1	031518-JETT-F-D-S	-88	-0.0242	-245.1		PPB	20 Mar 2018 18:42:45	U	SMPL
J1954-04 - 1	031518-JETT-F-U-B	-56	-0.0149	-245.1		PPB	20 Mar 2018 18:44:54	U	SMPL
J1954-05 - 1	031518-JETT-F-U-M	-101	-0.0280	-245.1		PPB	20 Mar 2018 18:47:02	U	SMPL
CCV21 - 1	CCV21	16786	4.8781	-245.1		PPB	20 Mar 2018 18:49:09	U	SMPL
CCB21 - 1	CCB21	-151	-0.0425	-245.1		PPB	20 Mar 2018 18:51:16	U	SMPL
J1954-06 - 1	031518-JETT-F-U-S	-120	-0.0335	-245.1		PPB	20 Mar 2018 18:53:24	U	SMPL
J1954-07 - 1	031518-TIDO-F-D-B	-116	-0.0323	-245.1		PPB	20 Mar 2018 18:55:31	U	SMPL
J1954-08 - 1	031518-TIDO-F-D-S	-104	-0.0289	-245.1		PPB	20 Mar 2018 18:57:38	U	SMPL
J1954-09 - 1	031518-TIDO-F-U-B	-101	-0.0280	-245.1		PPB	20 Mar 2018 18:59:45	U	SMPL
J1954-10 - 1	031518-TIDO-F-U-S	-101	-0.0280	-245.1		PPB	20 Mar 2018 19:01:52	U	SMPL
J1954-10DUP - 1	031518-TIDO-F-U-SDUP	-93	-0.0257	-245.1		PPB	20 Mar 2018 19:04:00	U	SMPL
J1954-10MS - 1	031518-TIDO-F-U-SMS	15195	4.4159	-245.1		PPB	20 Mar 2018 19:06:08	U	SMPL
J1954-10MSD - 1	031518-TIDO-F-U-SMSD	14775	4.2939	-245.1		PPB	20 Mar 2018 19:08:16	U	SMPL
PB107493BL - 1	PBW	-102	-0.0283	-245.1		PPB	20 Mar 2018 19:10:24	U	SMPL
PB107493BS - 1	LCSW	13408	3.8967	-245.1		PPB	20 Mar 2018 19:12:31	U	SMPL
CCV22 - 1	CCV22	16577	4.8174	-245.1		PPB	20 Mar 2018 19:14:38	U	SMPL
CCB22 - 1	CCB22	-117	-0.0326	-245.1		PPB	20 Mar 2018 19:16:46	U	SMPL
J1889-01 - 1	EFFWW	-16	-0.0033	-245.1		PPB	20 Mar 2018 19:18:54	U	SMPL
J1889-01DUP - 1	EFFWWDUP	-65	-0.0175	-245.1		PPB	20 Mar 2018 19:21:01	U	SMPL
J1889-01MS - 1	EFFWWMS	13180	3.8305	-245.1		PPB	20 Mar 2018 19:26:14	U	SMPL
J1889-01MSD - 1	EFFWWMSD	13110	3.8101	-245.1		PPB	20 Mar 2018 19:28:21	U	SMPL
J1891-08 - 1	PT-HG-WP	45298	13.1615	-245.1		PPB	20 Mar 2018 19:30:29	U	SMPL
J1976-01 - 1	031618-TIDO-F-D-B	-187	-0.0530	-245.1		PPB	20 Mar 2018 19:35:07	U	SMPL
J1976-02 - 1	031618-TIDO-F-D-S	-155	-0.0437	-245.1		PPB	20 Mar 2018 19:37:15	U	SMPL
J1976-03 - 1	031618-TIDO-F-U-B	-162	-0.0457	-245.1		PPB	20 Mar 2018 19:39:23	U	SMPL
J1976-04 - 1	031618-TIDO-F-U-S	-137	-0.0384	-245.1		PPB	20 Mar 2018 19:41:32	U	SMPL
J1978-01 - 1	031718-JETT-F-D-B	-92	-0.0254	-245.1		PPB	20 Mar 2018 19:43:39	U	SMPL
CCV23 - 1	CCV23	17452	5.0716	-245.1		PPB	20 Mar 2018 19:45:46	U	SMPL
CCB23 - 1	CCB23	-75	-0.0204	-245.1		PPB	20 Mar 2018 19:47:53	U	SMPL
J1978-02 - 1	031718-JETT-F-D-M	-103	-0.0286	-245.1		PPB	20 Mar 2018 19:50:01	U	SMPL
J1978-03 - 1	031718-JETT-F-D-S	-173	-0.0489	-245.1		PPB	20 Mar 2018 19:52:09	U	SMPL
J1978-04 - 1	031718-JETT-F-U-B	-166	-0.0469	-245.1		PPB	20 Mar 2018 19:54:16	U	SMPL
J1978-05 - 1	031718-JETT-F-U-M	-185	-0.0524	-245.1		PPB	20 Mar 2018 19:56:23	U	SMPL
J1978-06 - 1	031718-JETT-F-U-S	-167	-0.0472	-245.1		PPB	20 Mar 2018 19:58:31	U	SMPL
I1954-10LX5 - 1		-80	-0.0219	-245.1		PPB	20 Mar 2018 20:09:37	U	SMPL
I1954-10A - 1		15633	4.5431	-245.1		PPB	20 Mar 2018 20:11:45	U	SMPL
I1889-01LX5 - 1		-174	-0.0492	-245.1		PPB	20 Mar 2018 20:13:52	U	SMPL
I1889-01A - 1		13120	3.8130	-245.1		PPB	20 Mar 2018 20:16:00	U	SMPL
CCV24 - 1	CCV24	16766	4.8723	-245.1		PPB	20 Mar 2018 20:18:07	U	SMPL
CCB24 - 1	CCB24	-144	-0.0405	-245.1		PPB	20 Mar 2018 20:20:15	U	SMPL
I1891-08DLX2 - 1	PT-HG-WP	22953	6.6698	-245.1		PPB	20 Mar 2018 20:24:31	U	SMPL
CCV25 - 1	CCV25	16838	4.8932	-245.1		PPB	20 Mar 2018 20:30:58	U	SMPL
CCB25 - 1	CCB25	-138	-0.0387	-245.1		PPB	20 Mar 2018 20:33:08	U	SMPL