



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94041

Temp.: 16.7 °C

SLOPE.: 97.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/20/2018	18:00
2	CAL2	1	Water			6.99	03/20/2018	18:01
3	CAL3	1	Water			10.02	03/20/2018	18:02
4	ICV	1	Water			7.00	03/20/2018	18:03
5	CCV1	1	Water			2.01	03/20/2018	18:05
6	J1990-02	1	Water			6.94	03/20/2018	18:06
7	J1990-02DUP	1	Water			6.95	03/20/2018	18:07
8	J1992-02	1	Water			6.93	03/20/2018	18:08
9	J1993-05	1	Water			6.83	03/20/2018	18:09
10	J2002-04	1	Water			6.95	03/20/2018	18:11
11	J2002-05	1	Water			6.95	03/20/2018	18:11
12	CCV2	1	Water			12.02	03/20/2018	18:12

WORKLIST(Hardcopy Internal Chain)

2394041

WorkList Name : corrosivity j2002

WorkList ID : 109820

Date : 3/20/2018 5:49:41 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/20/2018	Water	J1990-02-1A	Corrosivity	Cool 4 deg C	PSEG01	J13	WELL-A	03/20/2018	9040C
03/20/2018	Water	J1992-02-1B	Corrosivity	Cool 4 deg C	PSEG01	J13	DRUM-1-4-A	03/20/2018	9040C
03/19/2018	Water	J1993-05-1A	Corrosivity	Cool 4 deg C	PSEG01	J21	TP-6-MH-A	03/19/2018	9040C
03/20/2018	Water	J2002-04-1A	Corrosivity	Cool 4 deg C	PSEG01	J22	TP-4-MH-B	03/20/2018	9040C
03/20/2018	Water	J2002-05-1A	Corrosivity	Cool 4 deg C	PSEG01	J22	TP-4-MH-B	03/20/2018	9040C

Date/Time 03/20/18 5:50 PM

Received by: [Signature]

Relinquished by: [Signature]

Date/Time 03/20/18 7:30 PM

Received by: [Signature]

Relinquished by: [Signature]