


7470A

LB94044
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35094



A= 0.0000e+000

B= 2.8593e-004

C= -5.2956e-003

Rho= 0.9998320

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.014	-0.014	-31	0.000	-31				
0.2	0.200	0.182	-0.018	656	0.0 %	656				
2.5	2.500	2.513	0.013	8807	0.0 %	8807				
5.0	5.000	5.109	0.109	17885	0.0 %	17885				
7.5	7.500	7.381	-0.119	25833	0.0 %	25833				
10.0	10.000	10.029	0.029	35094	0.0 %	35094				

LB94044 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 20 Mar 2018 20:49:58					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-31	-	0.0000	7470A	PPB	20 Mar 2018 20:51:18	S	Std
0.2 - 1	50.2	656	-	0.2000	7470A	PPB	20 Mar 2018 20:53:26	S	Std
2.5 - 1	52.5	8807	-	2.5000	7470A	PPB	20 Mar 2018 20:55:34	S	Std
5.0 - 1	55	17885	-	5.0000	7470A	PPB	20 Mar 2018 20:57:43	S	Std
7.5 - 1	57.5	25833	-	7.5000	7470A	PPB	20 Mar 2018 20:59:50	S	Std
10.0 - 1	510	35094	-	10.0000	7470A	PPB	20 Mar 2018 21:01:58	S	Std
ICV57 - 1	ICV57	13509	3.8574	-	7470A	PPB	20 Mar 2018 21:04:38	U	SMPL
ICB57 - 1	ICB57	-164	-0.0522	-	7470A	PPB	20 Mar 2018 21:06:44	U	SMPL
CCV26 - 1	CCV26	18124	5.1769	-	7470A	PPB	20 Mar 2018 21:08:51	U	SMPL
CCB26 - 1	CCB26	-178	-0.0562	-	7470A	PPB	20 Mar 2018 21:10:57	U	SMPL
CRA - 1	CRA	652	0.1811	-	7470A	PPB	20 Mar 2018 21:13:04	U	SMPL
HighStd - 1	HighStd	35848	10.2448	-	7470A	PPB	20 Mar 2018 21:15:11	U	SMPL
ChkStd - 1	ChkStd	23946	6.8416	-	7470A	PPB	20 Mar 2018 21:17:18	U	SMPL
PB107517BL - 1	PBW	-229	-0.0708	-	7470A	PPB	20 Mar 2018 21:19:27	U	SMPL
PB107517BS - 1	LCSW	13088	3.7370	-	7470A	PPB	20 Mar 2018 21:21:36	U	SMPL
J1892-08 - 1	PT-HG-WP	47416	13.5525	-	7470A	PPB	20 Mar 2018 21:23:43	U	SMPL
J1979-05 - 1	A3960	4706	1.3403	-	7470A	PPB	20 Mar 2018 21:30:32	U	SMPL
J1979-05DUP - 1	A3960DUP	4539	1.2925	-	7470A	PPB	20 Mar 2018 21:32:41	U	SMPL
J1979-05MS - 1	A3960MS	18049	5.1555	-	7470A	PPB	20 Mar 2018 21:34:49	U	SMPL
J1979-05MSD - 1	A3960MSD	17969	5.1326	-	7470A	PPB	20 Mar 2018 21:36:56	U	SMPL
CCV27 - 1	CCV27	17892	5.1106	-	7470A	PPB	20 Mar 2018 21:39:03	U	SMPL
CCB27 - 1	CCB27	-134	-0.0436	-	7470A	PPB	20 Mar 2018 21:41:11	U	SMPL
PB107531BL - 1	PBW	-30	-0.0139	-	7470A	PPB	20 Mar 2018 21:43:18	U	SMPL
PB107531BS - 1	LCSW	13337	3.8082	-	7470A	PPB	20 Mar 2018 21:45:25	U	SMPL
PB107531TB - 1	PB107481TB	-90	-0.0310	-	7470A	PPB	20 Mar 2018 21:47:32	U	SMPL
J1760-04 - 1	SU-03-031918	-20	-0.0110	-	7470A	PPB	20 Mar 2018 21:49:39	U	SMPL
J1760-04DUP - 1	SU-03-031918DUP	-14	-0.0093	-	7470A	PPB	20 Mar 2018 21:51:47	U	SMPL
J1760-04MS - 1	SU-03-031918MS	12905	3.6847	-	7470A	PPB	20 Mar 2018 21:53:54	U	SMPL
J1760-04MSD - 1	SU-03-031918MSD	12176	3.4762	-	7470A	PPB	20 Mar 2018 21:56:02	U	SMPL
J1974-02 - 1	YR18020063	-128	-0.0419	-	7470A	PPB	20 Mar 2018 21:58:11	U	SMPL
J1975-07 - 1	FS-SOIL	0	-0.0053	-	7470A	PPB	20 Mar 2018 22:00:19	U	SMPL
J1977-04 - 1	SP	-12	-0.0087	-	7470A	PPB	20 Mar 2018 22:02:26	U	SMPL
CCV28 - 1	CCV28	17763	5.0737	-	7470A	PPB	20 Mar 2018 22:04:32	U	SMPL
CCB28 - 1	CCB28	-169	-0.0536	-	7470A	PPB	20 Mar 2018 22:06:39	U	SMPL
J1979-04 - 1	SP-1	-10	-0.0082	-	7470A	PPB	20 Mar 2018 22:08:47	U	SMPL
PB107537BL - 1	PBW	58	0.0113	-	7470A	PPB	20 Mar 2018 22:10:55	U	SMPL
PB107537BS - 1	LCSW	12669	3.6172	-	7470A	PPB	20 Mar 2018 22:13:02	U	SMPL
PB107537TB - 1	PB107489TB	-90	-0.0310	-	7470A	PPB	20 Mar 2018 22:15:09	U	SMPL
J1979-06 - 1	A3960	-39	-0.0164	-	7470A	PPB	20 Mar 2018 22:17:17	U	SMPL
J1979-06DUP - 1	A3960DUP	-4	-0.0064	-	7470A	PPB	20 Mar 2018 22:19:25	U	SMPL
J1979-06MS - 1	A3960MS	13009	3.7144	-	7470A	PPB	20 Mar 2018 22:21:33	U	SMPL
J1979-06MSD - 1	A3960MSD	12863	3.6726	-	7470A	PPB	20 Mar 2018 22:23:41	U	SMPL
CCV29 - 1	CCV29	17435	4.9799	-	7470A	PPB	20 Mar 2018 22:25:48	U	SMPL
CCB29 - 1	CCB29	-198	-0.0619	-	7470A	PPB	20 Mar 2018 22:27:56	U	SMPL
J1979-05LX5 - 1		765	0.2134	-	7470A	PPB	20 Mar 2018 22:30:04	U	SMPL
J1979-05A - 1		20215	5.7748	-	7470A	PPB	20 Mar 2018 22:32:12	U	SMPL
J1760-04LX5 - 1		-170	-0.0539	-	7470A	PPB	20 Mar 2018 22:34:20	U	SMPL
J1760-04A - 1		13585	3.8791	-	7470A	PPB	20 Mar 2018 22:36:28	U	SMPL
J1979-06LX5 - 1		-101	-0.0342	-	7470A	PPB	20 Mar 2018 22:38:36	U	SMPL
J1979-06A - 1		13500	3.8548	-	7470A	PPB	20 Mar 2018 22:40:44	U	SMPL
J1892-08DLX2 - 1	PT-HG-WP	22761	6.5028	-	7470A	PPB	20 Mar 2018 22:42:52	U	SMPL
CCV30 - 1	CCV30	17465	4.9885	-	7470A	PPB	20 Mar 2018 22:44:59	U	SMPL
CCB30 - 1	CCB30	-212	-0.0659	-	7470A	PPB	20 Mar 2018 22:47:07	U	SMPL
							20 Mar 2018 23:01:34	U	SMPL