



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : Sweta

Run Number: LB94052

Temp.: 20.1 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/22/2018	07:50
2	CAL2	1	Water			7.01	03/22/2018	07:51
3	CAL3	1	Water			10.02	03/22/2018	07:53
4	ICV	1	Water			6.99	03/22/2018	08:00
5	CCV1	1	Water			2.01	03/22/2018	08:01
6	J2009-01	1	Solid	20.02	20	6.68	03/22/2018	08:03
7	J2009-01DUP	1	Solid	20.02	20	6.69	03/22/2018	08:05
8	J2009-02	1	Solid	20.03	20	5.99	03/22/2018	08:06
9	J2009-03	1	Solid	20.05	20	7.18	03/22/2018	08:09
10	CCV2	1	Water			12.01	03/22/2018	08:10

WORKLIST(Hardcopy Internal Chain)

2304052

WorkList Name : PH J2009

WorkList ID : 109839

Date : 3/22/2018 7:29:37 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/20/2018	Solid	J2009-01-F	pH	Cool 4 deg C	AREC01	K12	S-1	03/20/2018	9045C
03/20/2018	Solid	J2009-02-F	pH	Cool 4 deg C	AREC01	K12	S-2	03/20/2018	9045C
03/20/2018	Solid	J2009-03-F	pH	Cool 4 deg C	AREC01	K12	S-3	03/20/2018	9045C

Date/Time 03/22/18 7:30 AM
 Received by: JG
 Relinquished by: U

Date/Time 03/22/18 9:30 AM
 Received by: U
 Relinquished by: JG