


7470A

LB94066
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35594



A= 0.0000e+000

B= 2.8099e-004

C= 1.2691e-002

Rho= 0.9999915

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.011	0.011	-7	0.000	-7				
0.2	0.200	0.197	-0.003	655	0.0 %	655				
2.5	2.500	2.511	0.011	8892	0.0 %	8892				
5.0	5.000	4.970	-0.030	17641	0.0 %	17641				
7.5	7.500	7.497	-0.003	26636	0.0 %	26636				
10.0	10.000	10.014	0.014	35594	0.0 %	35594				

LB94066 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 22 Mar 2018 14:33:13					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-7	-	0.0000	7470A	PPB	22 Mar 2018 14:41:00	S	Std
0.2 - 1	50.2	655	-	0.2000	7470A	PPB	22 Mar 2018 14:43:08	S	Std
2.5 - 1	52.5	8892	-	2.5000	7470A	PPB	22 Mar 2018 14:45:17	S	Std
5.0 - 1	55	17641	-	5.0000	7470A	PPB	22 Mar 2018 14:47:25	S	Std
7.5 - 1	57.5	26636	-	7.5000	7470A	PPB	22 Mar 2018 14:49:33	S	Std
10.0 - 1	510	35594	-	10.0000	7470A	PPB	22 Mar 2018 14:51:41	S	Std
ICV58 - 1	ICV58	13676	3.8555	-	7470A	PPB	22 Mar 2018 14:54:11	U	SMPL
ICB58 - 1	ICB58	-114	-0.0193	-	7470A	PPB	22 Mar 2018 14:56:17	U	SMPL
CCV31 - 1	CCV31	17581	4.9528	-	7470A	PPB	22 Mar 2018 14:58:24	U	SMPL
CCB31 - 1	CCB31	-117	-0.0202	-	7470A	PPB	22 Mar 2018 15:00:31	U	SMPL
CRA - 1	CRA	651	0.1956	-	7470A	PPB	22 Mar 2018 15:02:38	U	SMPL
HighStd - 1	HighStd	35678	10.0380	-	7470A	PPB	22 Mar 2018 15:04:45	U	SMPL
ChkStd - 1	ChkStd	24758	6.9695	-	7470A	PPB	22 Mar 2018 15:06:52	U	SMPL
PB107532BL - 1	PBW	-220	-0.0491	-	7470A	PPB	22 Mar 2018 15:09:01	U	SMPL
PB107532BS - 1	LCSW	12635	3.5630	-	7470A	PPB	22 Mar 2018 15:11:10	U	SMPL
PB107532TB - 1	PB107532TB	-190	-0.0407	-	7470A	PPB	22 Mar 2018 15:13:18	U	SMPL
J1973-04 - 1	COMP-1	389	0.1220	-	7470A	PPB	22 Mar 2018 15:15:26	U	SMPL
J1973-05 - 1	COMP-2	-41	0.0012	-	7470A	PPB	22 Mar 2018 15:17:34	U	SMPL
J1973-05DUP - 1	COMP-2DUP	-50	-0.0014	-	7470A	PPB	22 Mar 2018 15:19:42	U	SMPL
J1973-05MS - 1	COMP-2MS	13549	3.8199	-	7470A	PPB	22 Mar 2018 15:21:49	U	SMPL
CCV32 - 1	CCV32	17796	5.0132	-	7470A	PPB	22 Mar 2018 15:23:56	U	SMPL
CCB32 - 1	CCB32	-135	-0.0252	-	7470A	PPB	22 Mar 2018 15:26:03	U	SMPL
J1973-05MSD - 1	COMP-2MSD	13414	3.7819	-	7470A	PPB	22 Mar 2018 15:33:18	U	SMPL
J1973-06 - 1	COMP-3	572	0.1734	-	7470A	PPB	22 Mar 2018 15:35:24	U	SMPL
PB107562BL - 1	PBW	-152	-0.0300	-	7470A	PPB	22 Mar 2018 15:37:31	U	SMPL
PB107562BS - 1	LCSW	12109	3.4152	-	7470A	PPB	22 Mar 2018 15:39:38	U	SMPL
PB107562TB - 1	PB107533TB	-202	-0.0441	-	7470A	PPB	22 Mar 2018 15:41:46	U	SMPL
J1756-08 - 1	OR-03-032018-A	-109	-0.0179	-	7470A	PPB	22 Mar 2018 15:43:53	U	SMPL
J1756-10 - 1	OR-03-032018-B	-108	-0.0177	-	7470A	PPB	22 Mar 2018 15:46:01	U	SMPL
J1870-03 - 1	WC-06-08-12	-122	-0.0216	-	7470A	PPB	22 Mar 2018 15:48:10	U	SMPL
J1876-03 - 1	WC-04-04-10	-106	-0.0171	-	7470A	PPB	22 Mar 2018 15:50:18	U	SMPL
J1876-06 - 1	WC-05-01-03	-96	-0.0143	-	7470A	PPB	22 Mar 2018 15:52:25	U	SMPL
CCV33 - 1	CCV33	17421	4.9079	-	7470A	PPB	22 Mar 2018 15:54:32	U	SMPL
CCB33 - 1	CCB33	-110	-0.0182	-	7470A	PPB	22 Mar 2018 15:56:39	U	SMPL
J1984-01 - 1	OILY-DEBRIS	-131	-0.0241	-	7470A	PPB	22 Mar 2018 15:58:47	U	SMPL
J1991-03 - 1	WES-0-6	-107	-0.0174	-	7470A	PPB	22 Mar 2018 16:00:53	U	SMPL
J1991-04 - 1	WES-6-12	-87	-0.0118	-	7470A	PPB	22 Mar 2018 16:03:00	U	SMPL
J1993-03 - 1	TP-6-MH-A	-16	0.0082	-	7470A	PPB	22 Mar 2018 16:05:07	U	SMPL
J1995-04 - 1	PIT-A-0-5	-114	-0.0193	-	7470A	PPB	22 Mar 2018 16:07:15	U	SMPL
J1995-08 - 1	PIT-A-0-10	-76	-0.0087	-	7470A	PPB	22 Mar 2018 16:09:22	U	SMPL
J1997-08 - 1	DGA-TOP-LAYER-1-A	-71	-0.0073	-	7470A	PPB	22 Mar 2018 16:11:31	U	SMPL
J1997-16 - 1	DGA-TOP-LAYER-1-B	-91	-0.0129	-	7470A	PPB	22 Mar 2018 16:13:39	U	SMPL
J1997-24 - 1	SLAB-CONCRETE-A	-84	-0.0109	-	7470A	PPB	22 Mar 2018 16:15:48	U	SMPL
J1997-32 - 1	SLAB-CONCRETE-B	-61	-0.0044	-	7470A	PPB	22 Mar 2018 16:17:55	U	SMPL
CCV34 - 1	CCV34	17307	4.8758	-	7470A	PPB	22 Mar 2018 16:20:01	U	SMPL
CCB34 - 1	CCB34	-125	-0.0224	-	7470A	PPB	22 Mar 2018 16:22:08	U	SMPL
J1997-40 - 1	CINDER-BLOCK	-13	0.0090	-	7470A	PPB	22 Mar 2018 16:24:16	U	SMPL
J1997-41 - 1	BLACK-MASTIC	-92	-0.0132	-	7470A	PPB	22 Mar 2018 16:26:24	U	SMPL
J1997-41DUP - 1	BLACK-MASTICDUP	-74	-0.0081	-	7470A	PPB	22 Mar 2018 16:28:31	U	SMPL
J1997-41MS - 1	BLACK-MASTICMS	12850	3.6234	-	7470A	PPB	22 Mar 2018 16:30:38	U	SMPL
J1997-41MSD - 1	BLACK-MASTICMSD	12686	3.5774	-	7470A	PPB	22 Mar 2018 16:32:46	U	SMPL
J1997-43 - 1	PAINT-CHIPS	-178	-0.0373	-	7470A	PPB	22 Mar 2018 16:34:53	U	SMPL
J2002-03 - 1	TP-4-MH-B	262	0.0863	-	7470A	PPB	22 Mar 2018 16:37:02	U	SMPL
PB107563BL - 1	PBW	-105	-0.0168	-	7470A	PPB	22 Mar 2018 16:39:10	U	SMPL
PB107563BS - 1	LCSW	12366	3.4874	-	7470A	PPB	22 Mar 2018 16:41:19	U	SMPL
PB107563TB - 1	PB107535TB	-197	-0.0427	-	7470A	PPB	22 Mar 2018 16:43:26	U	SMPL
CCV35 - 1	CCV35	18078	5.0925	-	7470A	PPB	22 Mar 2018 16:45:33	U	SMPL
CCB35 - 1	CCB35	-138	-0.0261	-	7470A	PPB	22 Mar 2018 16:47:41	U	SMPL
J1990-02 - 1	WELL-A	-94	-0.0137	-	7470A	PPB	22 Mar 2018 16:49:49	U	SMPL
J1990-02DUP - 1	WELL-ADUP	-99	-0.0151	-	7470A	PPB	22 Mar 2018 16:51:56	U	SMPL
J1990-02MS - 1	WELL-AMS	12300	3.4689	-	7470A	PPB	22 Mar 2018 16:54:03	U	SMPL
J1990-02MSD - 1	WELL-AMSD	12512	3.5285	-	7470A	PPB	22 Mar 2018 16:56:10	U	SMPL
J1990-03 - 1	WELL-B	-213	-0.0472	-	7470A	PPB	22 Mar 2018 16:58:18	U	SMPL
J1992-02 - 1	DRUM-1-4-A	-113	-0.0191	-	7470A	PPB	22 Mar 2018 17:00:25	U	SMPL
J1992-03 - 1	DRUM-1-4-B	-99	-0.0151	-	7470A	PPB	22 Mar 2018 17:02:34	U	SMPL
J1993-05 - 1	TP-6-MH-A	-111	-0.0185	-	7470A	PPB	22 Mar 2018 17:04:42	U	SMPL
J2002-05 - 1	TP-4-MH-B	-106	-0.0171	-	7470A	PPB	22 Mar 2018 17:06:50	U	SMPL
PB107565BL - 1	PBW	-117	-0.0202	-	7470A	PPB	22 Mar 2018 17:08:58	U	SMPL
CCV36 - 1	CCV36	17978	5.0644	-	7470A	PPB	22 Mar 2018 17:11:05	U	SMPL
CCB36 - 1	CCB36	-130	-0.0238	-	7470A	PPB	22 Mar 2018 17:13:13	U	SMPL
PB107565BS - 1	LCSW	12548	3.5386	-	7470A	PPB	22 Mar 2018 17:15:21	U	SMPL
J1990-01 - 1	WELL-A	224	0.0756	-	7470A	PPB	22 Mar 2018 17:17:28	U	SMPL
J1992-01 - 1	DRUM-1-4-A	72	0.0329	-	7470A	PPB	22 Mar 2018 17:19:36	U	SMPL
J1993-04 - 1	TP-6-MH-A	1073	0.3142	-	7470A	PPB	22 Mar 2018 17:21:43	U	SMPL
J1999-01 - 1	AM-WWTP-001	-182	-0.0384	-	7470A	PPB	22 Mar 2018 17:23:51	U	SMPL
J2002-04 - 1	TP-4-MH-B	7099	2.0075	-	7470A	PPB	22 Mar 2018 17:25:59	U	SMPL
J2002-04DUP - 1	TP-4-MH-BDUP	7154	2.0229	-	7470A	PPB	22 Mar 2018 17:28:06	U	SMPL
J2002-04MS - 1	TP-4-MH-BMS	22005	6.1959	-	7470A	PPB	22 Mar 2018 17:30:14	U	SMPL
J2002-04MSD - 1	TP-4-MH-BMSD	21303	5.9987	-	7470A	PPB	22 Mar 2018 17:32:23	U	SMPL
CCV37 - 1	CCV37	17433	4.9112	-	7470A	PPB	22 Mar 2018 17:34:32	U	SMPL
CCB37 - 1	CCB37	-102	-0.0160	-	7470A	PPB	22 Mar 2018 17:36:40	U	SMPL

LB94066 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 22 Mar 2018 14:33:13					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
J1973-05LX5 - 1		-49	-0.0011		-7470A	PPB	22 Mar 2018 17:38:49	U	SMPL
J1973-05A - 1		13528	3.8140		-7470A	PPB	22 Mar 2018 17:40:56	U	SMPL
J1997-41LX5 - 1		-125	-0.0224		-7470A	PPB	22 Mar 2018 17:43:04	U	SMPL
J1997-41A - 1		13488	3.8027		-7470A	PPB	22 Mar 2018 17:45:12	U	SMPL
J1990-02LX5 - 1		-156	-0.0311		-7470A	PPB	22 Mar 2018 17:47:20	U	SMPL
J1990-02A - 1		13539	3.8171		-7470A	PPB	22 Mar 2018 17:49:28	U	SMPL
J2002-04LX5 - 1		1161	0.3389		-7470A	PPB	22 Mar 2018 17:51:35	U	SMPL
J2002-04A - 1		20813	5.8610		-7470A	PPB	22 Mar 2018 17:53:43	U	SMPL
CCV38 - 1	CCV38	17265	4.8640		-7470A	PPB	22 Mar 2018 17:55:52	U	SMPL
CCB38 - 1	CCB38	-113	-0.0191		-7470A	PPB	22 Mar 2018 17:58:01	U	SMPL