

1394068

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

3/22/2018 14:17

Test: Reactive
Total CN Mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.945	0.0	0.094	
ICB1	-0.645	0.0	0.004	
CCV1	245.199	0.0	0.234	
CCB1	-0.592	0.0	0.004	
PB107566BL	-0.779	0.0	0.004	
J1990-02	-0.686	0.0	0.004	
J1990-02DUP	-0.623	0.0	0.004	
J1992-02	-0.714	0.0	0.004	
J1993-05	-0.669	0.0	0.004	
J2002-04	-0.461	0.0	0.004	
J2002-05	-0.659	0.0	0.004	
CCV2	259.618	0.0	0.247	
CCB2	-0.441	0.0	0.004	

N	13
Mean	45.730
SD	95.5454
CV%	208.93

AS

Aquakem v. 7.2AQ1

Results from time period:

Thu Mar 22 14:14:04 2018

Thu Mar 22 14:17:28 2018

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.9718	µg/l	3/22/2018 11:38:18	
5.OPPBCN	A	Reactive CN	P	4.2761	µg/l	3/22/2018 11:38:19	
10PPBCN	A	Reactive CN	P	9.0175	µg/l	3/22/2018 11:38:20	
50PPBCN	A	Reactive CN	P	50.7048	µg/l	3/22/2018 11:38:21	
100PPBCN	A	Reactive CN	P	99.5956	µg/l	3/22/2018 11:38:22	
250PPBCN	A	Reactive CN	P	254.6812	µg/l	3/22/2018 11:38:23	
500PPBCN	A	Reactive CN	P	497.6967	µg/l	3/22/2018 11:38:24	
ICV1	S	Reactive CN	P	95.9449	µg/l	3/22/2018 14:14:04	
ICB1	S	Reactive CN	P	-0.6448	µg/l	3/22/2018 14:14:05	
CCV1	S	Reactive CN	P	245.1988	µg/l	3/22/2018 14:14:06	
CCB1	S	Reactive CN	P	-0.5916	µg/l	3/22/2018 14:14:07	
PB107566BL	S	Reactive CN	P	-0.7788	µg/l	3/22/2018 14:14:08	
J1990-02	S	Reactive CN	P	-0.6863	µg/l	3/22/2018 14:14:09	
J1990-02DUP	S	Reactive CN	P	-0.6229	µg/l	3/22/2018 14:14:10	
J1992-02	S	Reactive CN	P	-0.7144	µg/l	3/22/2018 14:14:11	
J1993-05	S	Reactive CN	P	-0.6689	µg/l	3/22/2018 14:14:12	
J2002-04	S	Reactive CN	P	-0.4609	µg/l	3/22/2018 14:14:13	
J2002-05	S	Reactive CN	P	-0.6595	µg/l	3/22/2018 14:14:14	
CCV2	S	Reactive CN	P	259.6181	µg/l	3/22/2018 14:17:27	
CCB2	S	Reactive CN	P	-0.441	µg/l	3/22/2018 14:17:28	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

3/22/2018 11:39


Test Total CN

Accepted 3/22/2018 11:39

Factor 1069
Bias 0.004

Coeff. of det. 0.999853

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.9718	0.0000	
2	5.0PPBCN	0.008	4.2761	5.0000	
3	10PPBCN	0.013	9.0175	10.0000	
4	50PPBCN	0.052	50.7048	50.0000	
5	100PPBCN	0.098	99.5956	100.0000	
6	250PPBCN	0.243	254.6812	250.0000	
7	500PPBCN	0.470	497.6967	500.0000	