



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94069

Temp.: 22.8 °C

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/22/2018	14:15
2	CAL2	1	Water			7.00	03/22/2018	14:16
3	CAL3	1	Water			10.02	03/22/2018	14:18
4	ICV	1	Water			7.00	03/22/2018	14:20
5	CCV1	1	Water			2.01	03/22/2018	14:21
6	J1900-20	1	Water			6.30	03/22/2018	14:22
7	J2020-01	1	Water			4.32	03/22/2018	14:23
8	J2020-01DUP	1	Water			4.33	03/22/2018	14:25
9	CCV2	1	Water			12.01	03/22/2018	14:26

1394069

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph j2020

WorkList ID : 109883

Date : 3/22/2018 2:00:18 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/21/2018	Water	J2020-01-17	pH	Cool 4 deg C	STER04	K22	DEIONIZED-WATER	03/21/2018	SM 4500-PH B

Date/Time 03/22/18 2:00 PM
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 03/22/18 3:20 PM
Received by: [Signature]
Relinquished by: [Signature]