



Analytical Summary Report

Analysis Method: 9041A

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94071

Temp.: NA °C

SLOPE.: NA

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267
3.0-5.5 pH paper refill	W1934
6.0-8.0 pH paper refill	W2106
Test Paper,PH Short Range 9.0/10.0	W1939
6.0-8.0 pH paper refill	W2106
0.0-3.0 pH paper refill	W1931
10.0-12.0 pH paper refill	W1940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.0	03/22/2018	14:15
2	CAL2	1	Water			7.0	03/22/2018	14:16
3	CAL3	1	Water			10.0	03/22/2018	14:18
4	ICV	1	Water			7.0	03/22/2018	14:20
5	CCV1	1	Water			2.0	03/22/2018	14:21
6	J1900-22	1	Water			6.4	03/22/2018	14:22
7	CCV2	1	Water			12.0	03/22/2018	14:26