



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB94072

Ambient Barometric Pressure (mmHg): 755

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	03/22/2018	09:10
2	J1990-02		1	>100.0	>212.0	03/22/2018	09:35
3	J1990-02DUP		1	>100.0	>212.0	03/22/2018	10:02
4	J1992-02		1	>100.0	>212.0	03/22/2018	10:35
5	J1993-05		1	>100.0	>212.0	03/22/2018	10:55
6	J2002-04		1	>100.0	>212.0	03/22/2018	11:20
7	J2002-05		1	>100.0	>212.0	03/22/2018	11:45

Result = (Celsius * 1.8) + 32

LB 94072

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP32218

WorkList ID : 109855

Date : 3/22/2018 8:22:40 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/23/2018	Water	J1990-02	Flash Point	Cool 4 deg C	PSEG01	J13	WELL-A	03/20/2018	1010A
03/23/2018	Water	J1992-02	Flash Point	Cool 4 deg C	PSEG01	J13	DRUM-1-4-A	03/20/2018	1010A
03/23/2018	Water	J1993-05	Flash Point	Cool 4 deg C	PSEG01	J21	TP-6-MH-A	03/19/2018	1010A
03/23/2018	Water	J2002-04	Flash Point	Cool 4 deg C	PSEG01	J22	TP-4-MH-B	03/20/2018	1010A
03/23/2018	Water	J2002-05	Flash Point	Cool 4 deg C	PSEG01	J22	TP-4-MH-B	03/20/2018	1010A

Date/Time 3/22/18 8:30 AM

Received by: KS

Relinquished by: CP

Date/Time

Received by:

Relinquished by:

3/22/18 11:50 AM

CP

KS