

# TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 03/23/2018

DATE & TIME IN: 03/23/2018 09:00

OVEN TEMP: 103 °C

Run Number: LB94091

DATE & TIME OUT: 03/23/2018 13:00

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

| Dish # | Lab ID      | Client ID            | Empty Dish Weight (g) | Final Empty Dish Weight (g) | Sample Volume (ml) | 1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g) | 2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g) | Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g) | Weight (g) | Result mg/L |
|--------|-------------|----------------------|-----------------------|-----------------------------|--------------------|-----------------------------------------------------------------|-----------------------------------------------------------------|-------------------------------------------------------------------|------------|-------------|
| 1      | LB94091BL   | LB94091BL            | 1.8354                | 1.8354                      | 100                | 1.8354                                                          | 1.8354                                                          | 1.8354                                                            | 0          | 0           |
| 2      | LB94091BS   | LB94091BS            | 1.8374                | 1.8374                      | 100                | 1.887                                                           | 1.887                                                           | 1.887                                                             | 0.0496     | 496         |
| 3      | J2012-01    | 031918-TIDO-F-D-B    | 1.3929                | 1.3929                      | 200                | 1.4076                                                          | 1.4076                                                          | 1.4076                                                            | 0.0147     | 73.5        |
| 4      | J2012-01DUP | 031918-TIDO-F-D-BDUP | 1.4042                | 1.4042                      | 200                | 1.4189                                                          | 1.4189                                                          | 1.4189                                                            | 0.0147     | 73.5        |
| 5      | J2012-02    | 031918-TIDO-F-D-S    | 1.4112                | 1.4112                      | 200                | 1.4249                                                          | 1.4249                                                          | 1.4249                                                            | 0.0137     | 68.5        |
| 6      | J2012-03    | 031918-TIDO-F-U-B    | 1.4128                | 1.4128                      | 200                | 1.4268                                                          | 1.4268                                                          | 1.4268                                                            | 0.014      | 70          |
| 7      | J2012-04    | 031918-TIDO-F-U-S    | 1.4036                | 1.4036                      | 200                | 1.4168                                                          | 1.4168                                                          | 1.4168                                                            | 0.0132     | 66          |
| 8      | J2014-01    | 032018-TIDO-F-D-B    | 1.3951                | 1.3951                      | 200                | 1.4115                                                          | 1.4115                                                          | 1.4115                                                            | 0.0164     | 82          |
| 9      | J2014-02    | 032018-TIDO-F-D-S    | 1.3909                | 1.3909                      | 200                | 1.4056                                                          | 1.4056                                                          | 1.4056                                                            | 0.0147     | 73.5        |
| 10     | J2014-03    | 032018-TIDO-F-U-B    | 1.3974                | 1.3974                      | 200                | 1.4103                                                          | 1.4103                                                          | 1.4103                                                            | 0.0129     | 64.5        |
| 11     | J2014-04    | 032018-TIDO-F-U-S    | 1.4064                | 1.4064                      | 200                | 1.4185                                                          | 1.4185                                                          | 1.4185                                                            | 0.0121     | 60.5        |
| 12     | J2014-05    | 032018-JETT-F-D-M    | 1.4171                | 1.4171                      | 200                | 1.4261                                                          | 1.4261                                                          | 1.4261                                                            | 0.009      | 45          |
| 13     | J2014-06    | 032018-JETT-F-U-M    | 1.4071                | 1.4071                      | 400                | 1.4204                                                          | 1.4204                                                          | 1.4204                                                            | 0.0133     | 33.2        |
| 14     | J2015-01    | SPDES-1              | 1.3979                | 1.3979                      | 900                | 1.4015                                                          | 1.4015                                                          | 1.4015                                                            | 0.0036     | 4           |
| 15     | J2017-01    | COMP                 | 1.3988                | 1.3988                      | 50                 | 1.4146                                                          | 1.4146                                                          | 1.4146                                                            | 0.0158     | 316         |



# WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS032318

WorkList ID : 109895

Date : 3/23/2018 7:28:36 AM

21394091

| Due Date   | Matrix | Sample      | Test | Preservative | Customer | Storage Location | Customer Sample   | Collect Date | Method   |
|------------|--------|-------------|------|--------------|----------|------------------|-------------------|--------------|----------|
| 03/26/2018 | Water  | J2012-01 C  | TSS  | Cool 4 deg C | HDR103   | K12              | 031918-TIDO-F-D-B | 03/19/2018   | SM2540 D |
| 03/26/2018 | Water  | J2012-02 C  | TSS  | Cool 4 deg C | HDR103   | K12              | 031918-TIDO-F-D-S | 03/19/2018   | SM2540 D |
| 03/26/2018 | Water  | J2012-03 C  | TSS  | Cool 4 deg C | HDR103   | K12              | 031918-TIDO-F-U-B | 03/19/2018   | SM2540 D |
| 03/26/2018 | Water  | J2012-04 C  | TSS  | Cool 4 deg C | HDR103   | K12              | 031918-TIDO-F-U-S | 03/19/2018   | SM2540 D |
| 03/27/2018 | Water  | J2014-01 C  | TSS  | Cool 4 deg C | HDR103   | K12              | 032018-TIDO-F-D-B | 03/20/2018   | SM2540 D |
| 03/27/2018 | Water  | J2014-02 C  | TSS  | Cool 4 deg C | HDR103   | K12              | 032018-TIDO-F-D-S | 03/20/2018   | SM2540 D |
| 03/27/2018 | Water  | J2014-03 C  | TSS  | Cool 4 deg C | HDR103   | K12              | 032018-TIDO-F-U-B | 03/20/2018   | SM2540 D |
| 03/27/2018 | Water  | J2014-04 C  | TSS  | Cool 4 deg C | HDR103   | K12              | 032018-TIDO-F-U-S | 03/20/2018   | SM2540 D |
| 03/27/2018 | Water  | J2014-05 C  | TSS  | Cool 4 deg C | HDR103   | K12              | 032018-JETT-F-D-M | 03/20/2018   | SM2540 D |
| 03/27/2018 | Water  | J2014-06 C  | TSS  | Cool 4 deg C | HDR103   | K12              | 032018-JETT-F-U-M | 03/20/2018   | SM2540 D |
| 03/23/2018 | Water  | J2015-01 D  | TSS  | Cool 4 deg C | SKAN06   | K12              | SPDES-1           | 03/22/2018   | SM2540 D |
| 03/29/2018 | Water  | J2017-01 AD | TSS  | Cool 4 deg C | ARAM01   | K12              | COMP              | 03/22/2018   | SM2540 D |

Date/Time 03/23/18, 08:00  
 Received by: AS  
 Relinquished by: CP

Date/Time 03/23/18, 09:30  
 Received by: CP  
 Relinquished by: AS