



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB94097

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP62976
Turbidity std, 40NTU	WP62968
turbidity std 5 NTU	WP62971
turbidity std 20 NTU	WP62969
Turbidity std 1 NTU	WP62972
turbidity std 0 NTU	WP62974
turbidity std 10 NTU	WP62970
10 NTU Standard 500 ml	W2242
TURBIDITY LOD STD, 0.5NTU	WP62975

Intercept: 0.0788

Slope: 0.9499

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.182	03/23/2018	10:39
2	CAL2	1	1	1.090	03/23/2018	10:41
3	CAL3	5	1	4.720	03/23/2018	10:44
4	CAL4	10	1	9.670	03/23/2018	10:46
5	CAL5	20	1	18.800	03/23/2018	10:48
6	CAL6	40	1	38.200	03/23/2018	10:49



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB94097

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.100	10.550	03/23/2018	10:50
2	ICB		1	0.159	0.084	03/23/2018	10:53
3	CCV1	10	1	9.560	9.981	03/23/2018	10:59
4	CCB1		1	0.169	0.095	03/23/2018	11:02
5	LB94097BL		1	0.180	0.107	03/23/2018	11:06
6	J1161-10		1	0.723	0.678	03/23/2018	11:10
7	J1161-12		1	0.728	0.683	03/23/2018	11:13
8	J1900-37		1	20.200	21.182	03/23/2018	11:15
9	CCV2	10	1	9.480	9.897	03/23/2018	11:18
10	CCB2		1	0.165	0.091	03/23/2018	11:20