



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB94098

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP62976
Turbidity std, 40NTU	WP62968
turbidity std 5 NTU	WP62971
turbidity std 20 NTU	WP62969
Turbidity std 1 NTU	WP62972
turbidity std 0 NTU	WP62974
turbidity std 10 NTU	WP62970
10 NTU Standard 500 ml	W2242
TURBIDITY LOD STD, 0.5NTU	WP62975

Intercept: 0.0788

Slope: 0.9499

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.182	03/23/2018	10:39
2	CAL2	1	1	1.090	03/23/2018	10:41
3	CAL3	5	1	4.720	03/23/2018	10:44
4	CAL4	10	1	9.670	03/23/2018	10:46
5	CAL5	20	1	18.800	03/23/2018	10:48
6	CAL6	40	1	38.200	03/23/2018	10:49



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB94098

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.100	10.550	03/23/2018	10:50
2	ICB		1	0.159	0.084	03/23/2018	10:53
3	CCV1	10	1	9.480	9.897	03/23/2018	11:18
4	CCB1		1	0.165	0.091	03/23/2018	11:20
5	LB94098BL		1	0.179	0.105	03/23/2018	11:22
6	J1161-07		1	0.720	0.675	03/23/2018	11:24
7	J1161-09		1	0.726	0.681	03/23/2018	11:27
8	J1900-38		1	20.100	21.077	03/23/2018	11:30
9	CCV2	10	1	9.670	10.097	03/23/2018	11:32
10	CCB2		1	0.156	0.081	03/23/2018	11:35