



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB94106

Temp.: 21.3 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/23/2018	13:15
2	CAL2	1	Water			7.00	03/23/2018	13:17
3	CAL3	1	Water			10.02	03/23/2018	13:19
4	ICV	1	Water			7.00	03/23/2018	13:21
5	CCV1	1	Water			2.00	03/23/2018	13:23
6	J2025-01	1	Water			4.40	03/23/2018	13:25
7	J2025-01DUP	1	Water			4.41	03/23/2018	13:27
8	J2025-02	1	Water			4.46	03/23/2018	13:29
9	J2025-03	1	Water			4.95	03/23/2018	13:31
10	J2026-01	1	Water			4.82	03/23/2018	13:33
11	CCV2	1	Water			12.00	03/23/2018	13:35

WORKLIST(Hardcopy Internal Chain)

LB94106

Worklist Name : Ph032318

Worklist ID : 109927

Date : 3/23/2018 12:59:47 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/19/2018	Water	J2025-01	pH	Cool 4 deg C	STER07	K21	CELL-A	03/19/2018	SM 4500-PH B
03/19/2018	Water	J2025-02	pH	Cool 4 deg C	STER07	K21	CELLS-B-C	03/19/2018	SM 4500-PH B
03/19/2018	Water	J2025-03	pH	Cool 4 deg C	STER07	K21	CELL-D	03/19/2018	SM 4500-PH B
03/16/2018	Water	J2026-01	pH	Cool 4 deg C	STER15	K53	GILROY-OMEGA-CELL	03/16/2018	SM 4500-PH B

Date/Time 03/23/18 1:10 PM

Received by: SP

Relinquished by: CP

Date/Time 03/23/18 1:45 PM

Received by: CP

Relinquished by: SP