



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94119

Temp.: 17.2 °C

SLOPE.: 97.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/23/2018	18:10
2	CAL2	1	Water			6.99	03/23/2018	18:11
3	CAL3	1	Water			10.01	03/23/2018	18:12
4	ICV	1	Water			7.00	03/23/2018	18:14
5	CCV1	1	Water			2.01	03/23/2018	18:15
6	J2050-08	1	Water			6.87	03/23/2018	18:16
7	J2050-08DUP	1	Water			6.88	03/23/2018	18:17
8	J2052-06	1	Water			6.52	03/23/2018	18:19
9	CCV2	1	Water			12.02	03/23/2018	18:20

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j2052

WorkList ID : 109950

Date : 3/23/2018 5:48:44 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/23/2018	Water	J2050-08	Corrosivity	Cool 4 deg C	PSEG01	J23	TP-2-MH-C	03/23/2018	9040C
03/23/2018	Water	J2052-06	Corrosivity	Cool 4 deg C	PSEG01	J21	TP-B-GW	03/23/2018	9040C

Date/Time 03/23/18 5:50 PM
 Received by: JH
 Relinquished by: JH

Date/Time 03/23/18 7:10 PM
 Received by: JH
 Relinquished by: JH