


245.1

LB 94124
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35005



A= 0.0000e+000

B= 2.8485e-004

C= -7.6873e-003

Rho= 0.9999113

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	-40	0.000	-40				
0.05	0.050	0.055	0.005	220	0.0 %	220				
0.2	0.200	0.185	-0.015	677	0.0 %	677				
2.5	2.500	2.469	-0.031	8696	0.0 %	8696				
5.0	5.000	5.117	0.117	17992	0.0 %	17992				
7.5	7.500	7.479	-0.021	26281	0.0 %	26281				
10.0	10.000	9.964	-0.036	35005	0.0 %	35005				

LB94124 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 23 Mar 2018 19:34:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-40	-	0.0000	245.1	PPB	23 Mar 2018 19:51:23	S	Std
0.05 - 1	50.05	220	-	0.0500	245.1	PPB	23 Mar 2018 19:56:02	S	Std
0.2 - 1	50.2	677	-	0.2000	245.1	PPB	23 Mar 2018 19:58:11	S	Std
2.5 - 1	52.5	8696	-	2.5000	245.1	PPB	23 Mar 2018 20:00:19	S	Std
5.0 - 1	55	17992	-	5.0000	245.1	PPB	23 Mar 2018 20:02:27	S	Std
7.5 - 1	57.5	26281	-	7.5000	245.1	PPB	23 Mar 2018 20:04:34	S	Std
10.0 - 1	510	35005	-	10.0000	245.1	PPB	23 Mar 2018 20:06:42	S	Std
ICV60 - 1	ICV60	14039	3.9914	-245.1		PPB	23 Mar 2018 20:11:43	U	SMPL
ICB60 - 1	ICB60	-159	-0.0530	-245.1		PPB	23 Mar 2018 20:13:49	U	SMPL
CCV58 - 1	CCV58	17834	5.0724	-245.1		PPB	23 Mar 2018 20:15:55	U	SMPL
CCB58 - 1	CCB58	-184	-0.0601	-245.1		PPB	23 Mar 2018 20:18:02	U	SMPL
CRA - 1	CRA	200	0.0493	-245.1		PPB	23 Mar 2018 20:20:10	U	SMPL
HighStd - 1	HighStd	34165	9.7243	-245.1		PPB	23 Mar 2018 20:22:16	U	SMPL
ChkStd - 1	ChkStd	24353	6.9293	-245.1		PPB	23 Mar 2018 20:24:23	U	SMPL
PB107615BL - 1	PBW	-263	-0.0826	-245.1		PPB	23 Mar 2018 20:26:32	U	SMPL
PB107615BS - 1	LCSW	12478	3.5467	-245.1		PPB	23 Mar 2018 20:28:41	U	SMPL
J1949-01 - 1	32220008-3	-209	-0.0672	-245.1		PPB	23 Mar 2018 20:30:48	U	SMPL
J1949-01DUP - 1	32220008-3DUP	-110	-0.0390	-245.1		PPB	23 Mar 2018 20:32:56	U	SMPL
J1949-01MS - 1	32220008-3MS	12987	3.6917	-245.1		PPB	23 Mar 2018 20:35:04	U	SMPL
J1949-01MSD - 1	32220008-3MSD	12484	3.5484	-245.1		PPB	23 Mar 2018 20:37:13	U	SMPL
J2012-01 - 1	031918-TIDO-F-D-B	-187	-0.0610	-245.1		PPB	23 Mar 2018 20:39:20	U	SMPL
CCV59 - 1	CCV59	17488	4.9738	-245.1		PPB	23 Mar 2018 20:41:27	U	SMPL
CCB59 - 1	CCB59	-193	-0.0627	-245.1		PPB	23 Mar 2018 20:43:33	U	SMPL
J2012-02 - 1	031918-TIDO-F-D-S	-96	-0.0350	-245.1		PPB	23 Mar 2018 20:45:41	U	SMPL
J2012-03 - 1	031918-TIDO-F-U-B	-81	-0.0308	-245.1		PPB	23 Mar 2018 20:47:47	U	SMPL
J2012-04 - 1	031918-TIDO-F-U-S	-66	-0.0265	-245.1		PPB	23 Mar 2018 20:49:53	U	SMPL
J2014-01 - 1	032018-TIDO-F-D-B	-33	-0.0171	-245.1		PPB	23 Mar 2018 20:52:00	U	SMPL
J2014-02 - 1	032018-TIDO-F-D-S	-28	-0.0157	-245.1		PPB	23 Mar 2018 20:54:08	U	SMPL
J2014-03 - 1	032018-TIDO-F-U-B	-71	-0.0279	-245.1		PPB	23 Mar 2018 20:56:15	U	SMPL
J2014-04 - 1	032018-TIDO-F-U-S	-78	-0.0299	-245.1		PPB	23 Mar 2018 20:58:24	U	SMPL
J2014-05 - 1	032018-JETT-F-D-M	-63	-0.0256	-245.1		PPB	23 Mar 2018 21:00:32	U	SMPL
J2014-06 - 1	032018-JETT-F-U-M	-66	-0.0265	-245.1		PPB	23 Mar 2018 21:02:41	U	SMPL
CCV60 - 1	CCV60	17607	5.0077	-245.1		PPB	23 Mar 2018 21:04:48	U	SMPL
CCB60 - 1	CCB60	-174	-0.0573	-245.1		PPB	23 Mar 2018 21:06:55	U	SMPL
J1949-01LX5 - 1		-88	-0.0328	-245.1		PPB	23 Mar 2018 21:09:02	U	SMPL
J1949-01A - 1		13586	3.8623	-245.1		PPB	23 Mar 2018 21:11:09	U	SMPL
CCV61 - 1	CCV61	17702	5.0348	-245.1		PPB	23 Mar 2018 21:13:16	U	SMPL
CCB61 - 1	CCB61	-185	-0.0604	-245.1		PPB	23 Mar 2018 21:15:23	U	SMPL