



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : Sweta

Run Number: LB94126

Temp.: 21.3 °C

BalanceID: WC SC-4

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

True Value of CCV3 = 2.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/23/2018	17:25
2	CAL2	1	Water			7.00	03/23/2018	17:26
3	CAL3	1	Water			10.03	03/23/2018	17:28
4	ICV	1	Water			7.01	03/23/2018	17:29
5	CCV1	1	Water			2.01	03/23/2018	17:30
6	J1758-02	1	Solid	20.02	20	7.72	03/23/2018	17:31
7	J1758-02DUP	1	Solid	20.03	20	7.73	03/23/2018	17:32
8	J1758-04	1	Solid	20.05	20	7.86	03/23/2018	17:33
9	J1758-06	1	Solid	20.02	20	7.47	03/23/2018	17:35
10	J1758-08	1	Solid	20.01	20	7.92	03/23/2018	17:36
11	J1760-06	1	Solid	20.03	20	8.31	03/23/2018	17:38
12	J1760-08	1	Solid	20.02	20	8.72	03/23/2018	17:40
13	J2044-03	1	Solid	20.02	20	8.85	03/23/2018	17:41
14	J2044-04	1	Solid	20.05	20	9.15	03/23/2018	17:42
15	J2050-03	1	Solid	20.02	20	7.37	03/23/2018	17:43
16	CCV2	1	Water			12.01	03/23/2018	17:44
17	J2050-06	1	Solid	20.03	20	9.14	03/23/2018	17:46
18	J2052-02	1	Solid	20.06	20	8.87	03/23/2018	17:48
19	J2052-04	1	Solid	20.02	20	7.71	03/23/2018	17:50
20	CCV3	1	Water			2.01	03/23/2018	17:51

WORKLIST(Hardcopy Internal Chain)

1394126

WorkList Name : corrosivity j2044

WorkList ID : 109949

Date : 3/23/2018 5:48:08 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/23/2018	Solid	J1758-02	Corrosivity	Cool 4 deg C	PSEG05	J12	PL-02-032318-A	03/23/2018	9045C
03/23/2018	Solid	J1758-04	Corrosivity	Cool 4 deg C	PSEG05	J12	PL-02-032318-B	03/23/2018	9045C
03/23/2018	Solid	J1758-06	Corrosivity	Cool 4 deg C	PSEG05	J12	PL-02-032318-C	03/23/2018	9045C
03/23/2018	Solid	J1758-08	Corrosivity	Cool 4 deg C	PSEG05	J12	PL-02-032318-D	03/23/2018	9045C
03/23/2018	Solid	J1760-06	Corrosivity	Cool 4 deg C	PSEG05		SU-04-032318-A	03/23/2018	9045C
03/23/2018	Solid	J1760-08	Corrosivity	Cool 4 deg C	PSEG05		SU-04-032318-B	03/23/2018	9045C
03/23/2018	Solid	J2044-03	Corrosivity	Cool 4 deg C	TULL02	K51	TEST-PIT-LOC#1	03/23/2018	9045C
03/23/2018	Solid	J2044-04	Corrosivity	Cool 4 deg C	TULL02	K51	TEST-PIT-LOC#2	03/23/2018	9045C
03/23/2018	Solid	J2050-03	Corrosivity	Cool 4 deg C	PSEG01	J23	TP-2-MH-C	03/23/2018	9045C
03/23/2018	Solid	J2050-06	Corrosivity	Cool 4 deg C	PSEG01	J23	TP-3-DB	03/23/2018	9045C
03/23/2018	Solid	J2052-02	Corrosivity	Cool 4 deg C	PSEG01	J21	TP-BS	03/23/2018	9045C
03/23/2018	Solid	J2052-04	Corrosivity	Cool 4 deg C	PSEG01	J21	TP-BD	03/23/2018	9045C

Date/Time 03-23-18 5:48 PM
 Received by: JH
 Relinquished by: [Signature]

Date/Time 03-23-18 7:25 PM
 Received by: [Signature]
 Relinquished by: JH