



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94130

Temp.: 22.7 °C

SLOPE.: 99.9

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/26/2018	08:35
2	CAL2	1	Water			7.00	03/26/2018	08:37
3	CAL3	1	Water			10.01	03/26/2018	08:39
4	ICV	1	Water			7.01	03/26/2018	09:41
5	CCV1	1	Water			2.00	03/26/2018	08:43
6	J2056-04	1	Water			4.91	03/26/2018	08:45
7	J2056-04DUP	1	Water			4.92	03/26/2018	08:47
8	J2056-05	1	Water			4.66	03/26/2018	08:49
9	J2056-07	1	Water			5.04	03/26/2018	08:51
10	J2056-08	1	Water			5.18	03/26/2018	08:53
11	J2056-09	1	Water			6.13	03/26/2018	08:55
12	J2056-10	1	Water			5.43	03/26/2018	08:57
13	J2056-11	1	Water			6.78	03/26/2018	08:59
14	J2056-12	1	Water			5.74	03/26/2018	09:01
15	J2056-13	1	Water			6.14	03/26/2018	09:03
16	CCV2	1	Water			12.00	03/26/2018	09:05
17	J2056-15	1	Water			5.89	03/26/2018	09:07
18	J2056-16	1	Water			5.28	03/26/2018	09:09
19	J2056-17	1	Water			4.50	03/26/2018	09:11
20	J2056-18	1	Water			6.26	03/26/2018	09:13

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	J2056-19	1	Water			5.40	03/26/2018	09:15
22	J2056-20	1	Water			6.96	03/26/2018	09:17
23	J2056-21	1	Water			5.11	03/26/2018	09:19
24	CCV3	1	Water			2.00	03/26/2018	09:22

WORKLIST(Hardcopy Internal Chain)

LB 94130

Worklist Name : Ph032618

Worklist ID : 109973

Date : 3/26/2018 8:24:52 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/23/2018	Water	J2056-04	pH	Cool 4 deg C	PSEG01	J23	22146	03/23/2018	9040C
03/23/2018	Water	J2056-05	pH	Cool 4 deg C	PSEG01	J23	22148	03/23/2018	9040C
03/23/2018	Water	J2056-07	pH	Cool 4 deg C	PSEG01	J23	22136	03/23/2018	9040C
03/23/2018	Water	J2056-08	pH	Cool 4 deg C	PSEG01	J23	22151	03/23/2018	9040C
03/23/2018	Water	J2056-09	pH	Cool 4 deg C	PSEG01	J23	22152	03/23/2018	9040C
03/23/2018	Water	J2056-10	pH	Cool 4 deg C	PSEG01	J23	22155	03/23/2018	9040C
03/23/2018	Water	J2056-11	pH	Cool 4 deg C	PSEG01	J23	22156	03/23/2018	9040C
03/23/2018	Water	J2056-12	pH	Cool 4 deg C	PSEG01	J23	22157	03/23/2018	9040C
03/23/2018	Water	J2056-13	pH	Cool 4 deg C	PSEG01	J23	22159	03/23/2018	9040C
03/23/2018	Water	J2056-15	pH	Cool 4 deg C	PSEG01	J23	22162	03/23/2018	9040C
03/23/2018	Water	J2056-16	pH	Cool 4 deg C	PSEG01	J23	22164	03/23/2018	9040C
03/23/2018	Water	J2056-17	pH	Cool 4 deg C	PSEG01	J23	22165	03/23/2018	9040C
03/23/2018	Water	J2056-18	pH	Cool 4 deg C	PSEG01	J23	22167	03/23/2018	9040C
03/23/2018	Water	J2056-19	pH	Cool 4 deg C	PSEG01	J23	22169	03/23/2018	9040C
03/23/2018	Water	J2056-20	pH	Cool 4 deg C	PSEG01	J23	22171	03/23/2018	9040C
03/23/2018	Water	J2056-21	pH	Cool 4 deg C	PSEG01	J23	22172	03/23/2018	9040C

Date/Time

03/26/18 8:30 AM

Received by:

Relinquished by:

SP

Date/Time

03/26/18 9:25 AM

Received by:

Relinquished by:

SP