

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 03/26/2018

DATE & TIME IN: 03/26/2018 09:30

OVEN TEMP: 103 °C

Run Number: LB94131

DATE & TIME OUT: 03/26/2018 14:00

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB94131BL	LB94131BL	1.4171	1.4171	100	1.4171	1.4171	1.4171	0	0
2	LB94131BS	LB94131BS	1.3975	1.3975	100	1.4521	1.4521	1.4521	0.0546	546
3	J2038-01	032218-JETB-F-D-M	1.4181	1.4181	200	1.4338	1.4338	1.4338	0.0157	78.5
4	J2038-01DUP	032218-JETB-F-D-MDUP	1.3987	1.3987	200	1.4145	1.4145	1.4145	0.0158	79
5	J2038-02	032218-JETB-F-U-M	1.392	1.392	400	1.412	1.412	1.412	0.02	50
6	J2038-03	032218-TIDO-F-D-B	1.3941	1.3941	200	1.4159	1.4159	1.4159	0.0218	109
7	J2038-04	032218-TIDO-F-D-S	1.3982	1.3982	200	1.4268	1.4268	1.4268	0.0286	143
8	J2038-05	032218-TIDO-F-U-B	1.4075	1.4075	200	1.4093	1.4093	1.4093	0.0018	9
9	J2038-06	032218-TIDO-F-U-S	1.3915	1.3915	200	1.4092	1.4092	1.4092	0.0177	88.5
10	J2046-02	TW-02	1.4067	1.4067	700	1.4095	1.4095	1.4095	0.0028	4
11	J2050-07	TP-2-MH-C	1.3942	1.3942	100	1.4427	1.4427	1.4427	0.0485	485
12	J2052-05	TP-B-GW	1.4008	1.4008	20	1.4633	1.4633	1.4633	0.0625	3125
13	J2054-01	MW-02-20180323-DEP	1.4133	1.4133	400	1.4224	1.4224	1.4224	0.0091	22.8

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS032618

WorkList ID : 109957

Date : 3/26/2018 7:23:53 AM

1394131

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/29/2018	Water	J2038-01 C	TSS	Cool 4 deg C	HDR103	K31	032218-JETB-F-D-M	03/22/2018	SM2540 D
03/29/2018	Water	J2038-02 C	TSS	Cool 4 deg C	HDR103	K31	032218-JETB-F-U-M	03/22/2018	SM2540 D
03/29/2018	Water	J2038-03 C	TSS	Cool 4 deg C	HDR103	K31	032218-TIDO-F-D-B	03/22/2018	SM2540 D
03/29/2018	Water	J2038-04 C	TSS	Cool 4 deg C	HDR103	K31	032218-TIDO-F-D-S	03/22/2018	SM2540 D
03/29/2018	Water	J2038-05 C	TSS	Cool 4 deg C	HDR103	K31	032218-TIDO-F-U-B	03/22/2018	SM2540 D
03/29/2018	Water	J2038-06 C	TSS	Cool 4 deg C	HDR103	K31	032218-TIDO-F-U-S	03/22/2018	SM2540 D
03/30/2018	Water	J2046-02 D	TSS	Cool 4 deg C	ENTA04	K21	TW-02	03/23/2018	SM2540 D
03/28/2018	Water	J2050-07 F	TSS	Cool 4 deg C	PSEG01	J23	TP-2-MH-C	03/23/2018	SM2540 D
03/28/2018	Water	J2052-05 G	TSS	Cool 4 deg C	PSEG01	J21	TP-B-GW	03/23/2018	SM2540 D
03/30/2018	Water	J2054-01 F	TSS	Cool 4 deg C	GZAG04	K61	MW-02-20180323-DEP	03/23/2018	SM2540 D

Date/Time 03/26/18, 08:30

Received by: AS

Relinquished by: Cp

Date/Time 03/26/18, 09:30

Received by: AS

Relinquished by: AS