



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: AparnaDate: 03/27/2018OVEN TEMP1: 103 °C TEMP: 104 °CDATE & TIME IN: 03/26/2018 15:00 OUT: 03/27/2018 07:45OVEN TEMP2: 103 °C TEMP: 105 °CDATE & TIME IN: 03/27/2018 09:00 OUT: 03/27/2018 10:30Run Number: LB94132BalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB94132BL	LB94132BL	85.4885	85.4885	100	85.4885	85.4885	85.4885	0	0
2	J2046-02	TW-02	91.026	91.026	100	91.1718	91.1718	91.1718	0.1458	1458
3	J2054-01	MW-02-20180323-DEP	84.6651	84.6651	100	84.8976	84.8976	84.8976	0.2325	2325
4	J2054-01DUP	MW-02-20180323-DEPDUP	82.7971	82.7971	100	83.0265	83.0265	83.0265	0.2294	2294

A = Sample Volume (ml)
B = Empty Dish Weight (g)
C = Final Empty Dish + Sample weight after drying @105°C(g)
D = Weight (g)

Weight (g) = C - B

$$\text{Result mg/L} = \frac{D}{A} * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS032618

WorkList ID : 109984

Date : 3/26/2018 9:30:34 AM

LB9H132

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/30/2018	Water	J2046-02	TS	Cool 4 deg C	ENTA04	K21	TW-02	03/23/2018	SM2540 B
03/30/2018	Water	J2054-01	TS	Cool 4 deg C	GZAG04	K61	MW-02-20180323-DEP	03/23/2018	SM2540 B

Date/Time 03/26/18 14:00
 Received by: AS
 Relinquished by: CP

Date/Time 03/26/18 15:20
 Received by: CP
 Relinquished by: AS