

Chemtech Consulting Group


Analytical Review Report

Date Printed : 3/27/18

Analyst : 

Data File : LB94133.csv

Approved By : 

Approved Date : 

Worksheet # :

Lab Sample ID	Client ID	Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1
Parameter		PPB	Final Conc		%Rec	Method	Method	LCL	UCL	RPD	Line 2
Chloride											
STD1	STD1			W	3/19/18						
Chloride		PASS	0.000	0.000							mg/L
STD2	STD2			W	3/19/18						
Chloride		PASS	0.255	0.255							mg/L
STD3	STD3			W	3/19/18						
Chloride		PASS	0.637	0.637							mg/L
STD4	STD4			W	3/19/18						
Chloride		PASS	1.197	1.20							mg/L
STD5	STD5			W	3/19/18						
Chloride		PASS	2.856	2.86							mg/L
STD6	STD6			W	3/19/18						
Chloride		PASS	5.825	5.82							mg/L
STD7	STD7			W	3/19/18						
Chloride		PASS	7.680	7.68							mg/L
ICV1	ICV1			W	3/19/18						
Chloride		PASS	2.908	2.91	97.0	90	110				mg/L
ICB1	ICB1			W	3/19/18						
Chloride		PASS	0.000	0.000				+/-0.0750			mg/L
CCV1	CCV1			W	3/26/18						
Chloride		PASS	2.846	2.85	95.0	90	110				mg/L
CCB1	CCB1			W	3/26/18						
Chloride		PASS	0.000	0.000				+/-0.0750			mg/L
LB94133BLS	LB94133BLS			S	3/26/18						
Chloride		PASS	0.000	0.000				+/-0.0320			mg/Kg
LB94133BSS	LB94133BSS			S	3/26/18						
Chloride		PASS	2.890	57.70	96.0	90.00	110.00				mg/Kg
J2047-01	C1-CARVER-HALL-RCA	1		S	3/26/18						
Chloride	PASS		3.596	78.0							mg/Kg
J2047-01D	C1-CARVER-HALL-RCAD	1		S	3/26/18						
Chloride	PASS		3.529	76.70				1.7	20		mg/Kg
J2047-01S	C1-CARVER-HALL-RCAS	1		S	3/26/18						
Chloride	FAIL		7.725	168.0	138.0	80	120				mg/Kg
J2047-01SD	C1-CARVER-HALL-RCASD	1		S	3/26/18						
Chloride	FAIL		7.981	174.0	147.0	80	120	3.5	15		mg/Kg
CCV2	CCV2			W	3/26/18						
Chloride	PASS		2.788	2.79	93.0	90	110				mg/L
CCB2	CCB2			W	3/26/18						
Chloride	PASS		0.000	0.000				+/-0.0750			mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 3/27/18

Analyst : SV

Data File : LB94133.csv

Approved By : [Signature]

Approved Date : 3/27/18



Worksheet # : 275015

Lab Sample ID	Client ID	Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method	Line 1
Parameter			Final Conc	%Rec		LCL	UCL	Line 2
Nitrate								
STD1	STD1			W	3/19/18			
Nitrate		PASS	0.000	0.000				mg/L
STD2	STD2			W	3/19/18			
Nitrate		PASS	0.230	0.230				mg/L
STD3	STD3			W	3/19/18			
Nitrate		PASS	0.545	0.545				mg/L
STD4	STD4			W	3/19/18			
Nitrate		PASS	0.979	0.979				mg/L
STD5	STD5			W	3/19/18			
Nitrate		PASS	2.326	2.33				mg/L
STD6	STD6			W	3/19/18			
Nitrate		PASS	4.937	4.94				mg/L
STD7	STD7			W	3/19/18			
Nitrate		PASS	6.358	6.36				mg/L
ICV1	ICV1			W	3/19/18			
Nitrate		PASS	2.388	2.39	96.0	90	110	mg/L
ICB1	ICB1			W	3/19/18			
Nitrate		PASS	0.000	0.000			+/-0.0270	mg/L
CCV1	CCV1			W	3/26/18			
Nitrate		PASS	2.275	2.28	91.0	90	110	mg/L
CCB1	CCB1			W	3/26/18			
Nitrate		PASS	0.000	0.000			+/-0.0270	mg/L
LB94133BLS	LB94133BLS			S	3/26/18			
Nitrate		PASS	0.000	0.000			+/-0.0110	mg/Kg
LB94133BSS	LB94133BSS			S	3/26/18			
Nitrate		PASS	2.377	47.40	95.0	90.00	110.00	mg/Kg
J2047-01	C1-CARVER-HALL-RCA		1	S	3/26/18			
Nitrate		PASS	0.247	5.36				mg/Kg
J2047-01D	C1-CARVER-HALL-RCAD		1	S	3/26/18			
Nitrate		PASS	0.247	5.37			0.2	20 mg/Kg
J2047-01S	C1-CARVER-HALL-RCAS		1	S	3/26/18			
Nitrate		PASS	2.525	54.9	91.0	80	120	mg/Kg
J2047-01SD	C1-CARVER-HALL-RCASD		1	S	3/26/18			
Nitrate		PASS	2.476	54.0	89.0	80	120	1.7 15 mg/Kg
CCV2	CCV2			W	3/26/18			
Nitrate		PASS	2.294	2.29	92.0	90	110	mg/L
CCB2	CCB2			W	3/26/18			
Nitrate		PASS	0.000	0.000			+/-0.0270	mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 3/27/18
 Analyst : 
 Data File : LB94133.csv

Approved By : 
 Approved Date : 3/27/18
 Worksheet # : 

Lab Sample ID	Client ID	Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method	Line 1
Parameter			Final Conc	%Rec		LCL	UCL	Line 2
Nitrite								
STD1	STD1			W	3/19/18			
Nitrite		PASS	0.000	0.000				mg/L
STD2	STD2			W	3/19/18			
Nitrite		PASS	0.237	0.237				mg/L
STD3	STD3			W	3/19/18			
Nitrite		PASS	0.641	0.641				mg/L
STD4	STD4			W	3/19/18			
Nitrite		PASS	1.178	1.18				mg/L
STD5	STD5			W	3/19/18			
Nitrite		PASS	2.919	2.92				mg/L
STD6	STD6			W	3/19/18			
Nitrite		PASS	5.783	5.78				mg/L
STD7	STD7			W	3/19/18			
Nitrite		PASS	7.691	7.69				mg/L
ICV1	ICV1			W	3/19/18			
Nitrite		PASS	2.929	2.93	98.0	90	110	mg/L
ICB1	ICB1			W	3/19/18			
Nitrite		PASS	0.000	0.000			+/-0.0220	mg/L
CCV1	CCV1			W	3/26/18			
Nitrite		PASS	2.799	2.80	93.0	90	110	mg/L
CCB1	CCB1			W	3/26/18			
Nitrite		PASS	0.000	0.000			+/-0.0220	mg/L
LB94133BLS	LB94133BLS			S	3/26/18			
Nitrite		PASS	0.000	0.000			+/-0.0100	mg/Kg
LB94133BSS	LB94133BSS			S	3/26/18			
Nitrite		PASS	2.890	57.70	96.0	90.00	110.00	mg/Kg
J2047-01	C1-CARVER-HALL-RCA		1	S	3/26/18			
Nitrite		PASS	0.000	0.000				mg/Kg
J2047-01D	C1-CARVER-HALL-RCAD		1	S	3/26/18			
Nitrite		PASS	0.000	0.00			0	20 mg/Kg
J2047-01S	C1-CARVER-HALL-RCAS		1	S	3/26/18			
Nitrite		PASS	2.923	63.5	97.0	80	120	mg/Kg
J2047-01SD	C1-CARVER-HALL-RCASD		1	S	3/26/18			
Nitrite		PASS	2.967	64.7	99.0	80	120	1.9 15 mg/Kg
CCV2	CCV2			W	3/26/18			
Nitrite		PASS	2.763	2.76	92.0	90	110	mg/L
CCB2	CCB2			W	3/26/18			
Nitrite		PASS	0.000	0.000			+/-0.0220	mg/L

Chemtech Consulting Group

Analytical Review Report

Date Printed : 3/27/18
 Analyst : 
 Data File : LB94133.csv

Approved By : _____
 Approved Date : _____
 Worksheet # : _____

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				
Parameter			PPB		Final Conc	%Rec	Method	Method			Line 1	
							LCL	UCL	RPD	Max RPD	Units	Line 2
Sulfate												
STD1	STD1				W	3/19/18						
Sulfate		PASS	0.000		0.000						mg/L	
STD2	STD2				W	3/19/18						
Sulfate		PASS	1.311		1.31						mg/L	
STD3	STD3				W	3/19/18						
Sulfate		PASS	3.197		3.20						mg/L	
STD4	STD4				W	3/19/18						
Sulfate		PASS	5.899		5.90						mg/L	
STD5	STD5				W	3/19/18						
Sulfate		PASS	14.149		14.1						mg/L	
STD6	STD6				W	3/19/18						
Sulfate		PASS	29.550		29.6						mg/L	
STD7	STD7				W	3/19/18						
Sulfate		PASS	38.144		38.1						mg/L	
ICV1	ICV1				W	3/19/18						
Sulfate		PASS	15.206		15.20	101.0	90	110			mg/L	
ICB1	ICB1				W	3/19/18						
Sulfate		PASS	0.000		0.000			+/-0.1320			mg/L	
CCV1	CCV1				W	3/26/18						
Sulfate		PASS	13.829		13.80	92.0	90	110			mg/L	
CCB1	CCB1				W	3/26/18						
Sulfate		PASS	0.000		0.000			+/-0.1320			mg/L	
LB94133BLS	LB94133BLS				S	3/26/18						
Sulfate		PASS	0.000		0.000			+/-0.0960			mg/Kg	
LB94133BSS	LB94133BSS				S	3/26/18						
Sulfate		PASS	15.184		303.00	101.0	90.00	110.00			mg/Kg	
J2047-01	C1-CARVER-HALL-RCA	1			S	3/26/18						
Sulfate	OR		56.280		1220						mg/Kg	
J2047-01D	C1-CARVER-HALL-RCAD	1			S	3/26/18						
Sulfate	PASS		54.947		1190.00				2.5	20	mg/Kg	
J2047-01S	C1-CARVER-HALL-RCAS	1			S	3/26/18						
Sulfate	PASS		68.332		1480.0	80.0	80	120			mg/Kg	
J2047-01SD	C1-CARVER-HALL-RCASD	1			S	3/26/18						
Sulfate	FAIL		67.030		1460.0	73.0	80	120	1.4	15	mg/Kg	
J2047-01DL	C1-CARVER-HALL-RCADL	2			S	3/26/18						
Sulfate	PASS		25.370		1100						mg/Kg	
CCV2	CCV2				W	3/26/18						
Sulfate		PASS	13.935		13.90	93.0	90	110			mg/L	
CCB2	CCB2				W	3/26/18						
Sulfate		PASS	0.000		0.000			+/-0.1320			mg/L	