



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94136

Temp.: 17.2 °C

SLOPE.: 97.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/23/2018	18:10
2	CAL2	1	Water			6.99	03/23/2018	18:11
3	CAL3	1	Water			10.01	03/23/2018	18:12
4	ICV	1	Water			7.00	03/23/2018	18:14
5	CCV1	1	Water			2.02	03/23/2018	18:25
6	J2046-02	1	Water			6.95	03/23/2018	18:26
7	J2046-02DUP	1	Water			6.96	03/23/2018	18:27
8	CCV2	1	Water			12.02	03/23/2018	18:29

23 94136

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph j2046

WorkList ID : 109994

Date : 3/23/2018 6:25:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/23/2018	Water	J2046-02 -17	pH	Cool 4 deg C	ENTA04	K21	TW-02	03/23/2018	SM 4500-PH B

Date/Time 03/23/18 6:25 PM
Received by: JP
Relinquished by: CP

Date/Time 03/23/18 7:04 PM
Received by: JP
Relinquished by: JP