



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94154

Temp.: 15.4 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/26/2018	15:35
2	CAL2	1	Water			7.01	03/26/2018	15:36
3	CAL3	1	Water			10.03	03/26/2018	15:37
4	ICV	1	Water			6.99	03/26/2018	15:38
5	CCV1	1	Water			2.01	03/26/2018	15:39
6	J2067-02	1	Water			6.51	03/26/2018	15:40
7	J2067-02DUP	1	Water			6.52	03/26/2018	15:41
8	J2071-11	1	Water			6.85	03/26/2018	15:42
9	CCV2	1	Water			12.02	03/26/2018	15:44

13 94154

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j2062

WorkList ID : 110013

Date : 3/26/2018 3:12:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/26/2018	Water	J2067-02	Corrosivity	Cool 4 deg C	PSEG01	J33	RO-CV-RB47185RT	03/26/2018	9040C
03/26/2018	Water	J2071-11	Corrosivity	Cool 4 deg C	PSEG01	J43	TP-7-DB	03/26/2018	9040C

Date/Time 03/26/18 3:20 PM
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 03/26/18 5:30 PM
Received by: [Signature]
Relinquished by: [Signature]