



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB94155

Temp.: 15.4 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

True Value of CCV3 = 2.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/26/2018	15:35
2	CAL2	1	Water			7.01	03/26/2018	15:36
3	CAL3	1	Water			10.03	03/26/2018	15:37
4	ICV	1	Water			6.99	03/26/2018	15:38
5	CCV1	1	Water			2.01	03/26/2018	15:40
6	J2062-06	1	Water			5.02	03/26/2018	15:42
7	J2062-06DUP	1	Water			5.03	03/26/2018	15:44
8	J2062-07	1	Water			6.41	03/26/2018	15:45
9	J2062-08	1	Water			5.65	03/26/2018	15:46
10	J2062-09	1	Water			6.17	03/26/2018	15:48
11	J2062-10	1	Water			6.91	03/26/2018	15:50
12	J2062-11	1	Water			6.72	03/26/2018	15:51
13	J2067-01	1	Water			6.51	03/26/2018	16:00
14	J2071-10	1	Water			6.85	03/26/2018	16:01
15	J2072-01	1	Water			8.57	03/26/2018	16:03
16	CCV2	1	Water			12.01	03/26/2018	16:05
17	J2072-02	1	Water			12.86	03/26/2018	16:06
18	J2072-03	1	Water			12.38	03/26/2018	16:09
19	J2072-04	1	Water			7.34	03/26/2018	16:10
20	J2072-05	1	Water			12.22	03/26/2018	16:11

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water			2.01	03/26/2018	16:12

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph J2062

WorkList ID : 110012

Date : 3/26/2018 3:11:58 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/26/2018	Water	J2062-06 -1A	pH	Cool 4 deg C	PSEG01	J31	19626	03/26/2018	9040C
03/26/2018	Water	J2062-07 -1A	pH	Cool 4 deg C	PSEG01	J31	19627	03/26/2018	9040C
03/26/2018	Water	J2062-08 -1A	pH	Cool 4 deg C	PSEG01	J31	19628	03/26/2018	9040C
03/26/2018	Water	J2062-09 -1A	pH	Cool 4 deg C	PSEG01	J31	19674	03/26/2018	9040C
03/26/2018	Water	J2062-10 -1A	pH	Cool 4 deg C	PSEG01	J31	19643	03/26/2018	9040C
03/26/2018	Water	J2062-11 -1A	pH	Cool 4 deg C	PSEG01	J31	19673	03/26/2018	9040C
03/26/2018	Water	J2067-01 -1A	pH	Cool 4 deg C	PSEG01	J33	RO-CV-RB47185RT	03/26/2018	9040C
03/26/2018	Water	J2071-10 -1A	pH	Cool 4 deg C	PSEG01	J43	TP-7-DB	03/26/2018	9040C
03/22/2018	Water	J2072-01 -1A	pH	Cool 4 deg C	WEST04	K63	COAA1	03/22/2018	9040C
03/22/2018	Water	J2072-02 -1A	pH	Cool 4 deg C	WEST04	K63	COAB1	03/22/2018	9040C
03/22/2018	Water	J2072-03 -1A	pH	Cool 4 deg C	WEST04	K63	COAB2	03/22/2018	9040C
03/22/2018	Water	J2072-04 -1A	pH	Cool 4 deg C	WEST04	K63	COAB3	03/22/2018	9040C
03/22/2018	Water	J2072-05 -1A	pH	Cool 4 deg C	WEST04	K63	COAB4	03/22/2018	9040C

Date/Time 03/26/18 3:15 PM
 Received by: JD
 Relinquished by: CD

Date/Time 03/26/18 5:25 PM
 Received by: CD
 Relinquished by: JP