



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94156

Temp.: 22.6 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/26/2018	17:15
2	CAL2	1	Water			7.00	03/26/2018	17:16
3	CAL3	1	Water			10.01	03/26/2018	17:18
4	ICV	1	Water			7.01	03/26/2018	17:20
5	CCV1	1	Water			2.02	03/26/2018	17:21
6	J1749-06	1	Solid	20.02	20	7.39	03/26/2018	17:22
7	J1749-06DUP	1	Solid	20.03	20	7.40	03/26/2018	17:23
8	J2063-03	1	Solid	20.04	20	5.30	03/26/2018	17:25
9	J2064-03	1	Solid	20.05	20	6.48	03/26/2018	17:26
10	J2065-03	1	Solid	20.02	20	6.59	03/26/2018	17:28
11	J2066-03	1	Solid	20.01	20	5.92	03/26/2018	17:30
12	J2068-03	1	Solid	20.03	20	8.07	03/26/2018	17:31
13	J2071-03	1	Solid	20.06	20	9.30	03/26/2018	17:33
14	J2071-06	1	Solid	20.03	20	8.96	03/26/2018	17:34
15	J2071-09	1	Solid	20.03	20	7.84	03/26/2018	17:35
16	CCV2	1	Water			12.02	03/26/2018	17:36

WORKLIST(Hardcopy Internal Chain)

1294156

WorkList Name : corrosion j2069

WorkList ID : 110014

Date : 3/26/2018 3:13:28 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/26/2018	Solid	J1749-06	Corrosivity	Cool 4 deg C	PSEG05		BU-03-032618	03/26/2018	9045C
03/26/2018	Solid	J2063-03	Corrosivity	Cool 4 deg C	PSEG01	J31	RO-CV-2533	03/26/2018	9045C
03/26/2018	Solid	J2064-03	Corrosivity	Cool 4 deg C	PSEG01	J31	RO-CV-2532	03/26/2018	9045C
03/26/2018	Solid	J2065-03	Corrosivity	Cool 4 deg C	PSEG01	J33	RO-CV-2536	03/26/2018	9045C
03/26/2018	Solid	J2066-03	Corrosivity	Cool 4 deg C	PSEG01	J33	RO-CV-2036	03/26/2018	9045C
03/26/2018	Solid	J2068-03	Corrosivity	Cool 4 deg C	PSEG01	J41	RO-CV-2043	03/26/2018	9045C
03/26/2018	Solid	J2071-03	Corrosivity	Cool 4 deg C	PSEG01	J43	TP-1-MH-D-S	03/26/2018	9045C
03/26/2018	Solid	J2071-06	Corrosivity	Cool 4 deg C	PSEG01	J43	TP-1-MH-D-D	03/26/2018	9045C
03/26/2018	Solid	J2071-09	Corrosivity	Cool 4 deg C	PSEG01	J43	TP-7-DB	03/26/2018	9045C

Date/Time 03/26/18 3:25 PM
 Received by: JS
 Relinquished by: JS

Date/Time 03/26/18 6:15 PM
 Received by: CP
 Relinquished by: JS