


7471B

LB 94159
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42439



A= 0.0000e+000

B= 2.3597e-004

C= 3.4295e-002

Rho= 0.9999563

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	-23	0.000	-23				
0.2	0.200	0.223	0.023	801	0.0 %	801				
2.5	2.500	2.466	-0.034	10307	0.0 %	10307				
5.0	5.000	4.964	-0.036	20891	0.0 %	20891				
7.5	7.500	7.469	-0.031	31506	0.0 %	31506				
10	10.000	10.049	0.049	42439	0.0 %	42439				

LB94159 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 26 Mar 2018 19:16:42					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-23	-	0.0000	7471B	PPB	26 Mar 2018 19:20:55	S	Std
0.2 - 1	50.2	801	-	0.2000	7471B	PPB	26 Mar 2018 19:23:13	S	Std
2.5 - 1	52.5	10307	-	2.5000	7471B	PPB	26 Mar 2018 19:25:30	S	Std
5.0 - 1	55	20891	-	5.0000	7471B	PPB	26 Mar 2018 19:27:48	S	Std
7.5 - 1	57.5	31506	-	7.5000	7471B	PPB	26 Mar 2018 19:30:05	S	Std
10 - 1	510	42439	-	10.0000	7471B	PPB	26 Mar 2018 19:32:22	S	Std
ICV62 - 1	ICV62	16207	3.8587	-	7471B	PPB	26 Mar 2018 19:35:05	U	SMPL
ICB62 - 1	ICB62	-208	-0.0148	-	7471B	PPB	26 Mar 2018 19:37:20	U	SMPL
CCV70 - 1	CCV70	21035	4.9979	-	7471B	PPB	26 Mar 2018 19:39:36	U	SMPL
CCB70 - 1	CCB70	-173	-0.0065	-	7471B	PPB	26 Mar 2018 19:41:51	U	SMPL
CRA - 1	CRA	795	0.2219	-	7471B	PPB	26 Mar 2018 19:44:08	U	SMPL
HighStd - 1	HighStd	42189	9.9897	-	7471B	PPB	26 Mar 2018 19:46:24	U	SMPL
ChkStd - 1	ChkStd	29158	6.9147	-	7471B	PPB	26 Mar 2018 19:48:40	U	SMPL
PB107623BL - 1	PBS	-235	-0.0212	-	7471B	PPB	26 Mar 2018 19:50:58	U	SMPL
PB107623BS - 1	LCSS	15287	3.6416	-	7471B	PPB	26 Mar 2018 19:53:16	U	SMPL
J1748-07 - 1	AU-05-032218-A	2066	0.5218	-	7471B	PPB	26 Mar 2018 19:55:33	U	SMPL
J1748-09 - 1	AU-05-032218-B	2168	0.5459	-	7471B	PPB	26 Mar 2018 19:57:50	U	SMPL
J1748-11 - 1	AU-05-032218-C	1070	0.2868	-	7471B	PPB	26 Mar 2018 20:00:07	U	SMPL
J1748-13 - 1	AU-05-032218-D	1425	0.3706	-	7471B	PPB	26 Mar 2018 20:02:25	U	SMPL
J1760-05 - 1	SU-04-032318-A	635	0.1841	-	7471B	PPB	26 Mar 2018 20:04:41	U	SMPL
CCV71 - 1	CCV71	21115	5.0168	-	7471B	PPB	26 Mar 2018 20:06:56	U	SMPL
CCB71 - 1	CCB71	-134	0.0027	-	7471B	PPB	26 Mar 2018 20:09:12	U	SMPL
J1760-07 - 1	SU-04-032318-B	1958	0.4963	-	7471B	PPB	26 Mar 2018 20:11:28	U	SMPL
J2044-01 - 1	TEST-PIT-LOC#1	16425	3.9101	-	7471B	PPB	26 Mar 2018 20:13:44	U	SMPL
J2044-02 - 1	TEST-PIT-LOC#2	1607	0.4135	-	7471B	PPB	26 Mar 2018 20:16:00	U	SMPL
J2046-01 - 1	PF-01	315	0.1086	-	7471B	PPB	26 Mar 2018 20:18:17	U	SMPL
J2047-01 - 1	C1-CARVER-HALL-RCA	13760	3.2812	-	7471B	PPB	26 Mar 2018 20:20:34	U	SMPL
J2047-01DUP - 1	C1-CARVER-HALL-RCADUP	13120	3.1302	-	7471B	PPB	26 Mar 2018 20:22:50	U	SMPL
J2047-01MS - 1	C1-CARVER-HALL-RCAMS	29723	7.0480	-	7471B	PPB	26 Mar 2018 20:25:08	U	SMPL
J2047-01MSD - 1	C1-CARVER-HALL-RCAMSD	29205	6.9258	-	7471B	PPB	26 Mar 2018 20:27:25	U	SMPL
J2048-01 - 1	SB-108(4-5)	4055	0.9912	-	7471B	PPB	26 Mar 2018 20:29:44	U	SMPL
J2048-02 - 1	SB-108(12-13)	65	0.0496	-	7471B	PPB	26 Mar 2018 20:32:01	U	SMPL
CCV72 - 1	CCV72	21320	5.0652	-	7471B	PPB	26 Mar 2018 20:34:17	U	SMPL
CCB72 - 1	CCB72	-114	0.0074	-	7471B	PPB	26 Mar 2018 20:36:34	U	SMPL
J2048-03 - 1	SB-106(2-3)	252	0.0938	-	7471B	PPB	26 Mar 2018 20:38:50	U	SMPL
J2048-04 - 1	SB-106(20-21)	117	0.0619	-	7471B	PPB	26 Mar 2018 20:41:06	U	SMPL
J2048-05 - 1	DUP-032318	8818	2.1151	-	7471B	PPB	26 Mar 2018 20:43:22	U	SMPL
J2050-01 - 1	TP-2-MH-C	9891	2.3683	-	7471B	PPB	26 Mar 2018 20:45:39	U	SMPL
J2050-04 - 1	TP-3-DB	166537	39.3321	-	7471B	PPB	26 Mar 2018 20:47:56	U	SMPL
J2052-01 - 1	TP-BS	814	0.2264	-	7471B	PPB	26 Mar 2018 20:50:13	U	SMPL
J2052-03 - 1	TP-BD	1423	0.3701	-	7471B	PPB	26 Mar 2018 20:52:40	U	SMPL
PB107639BL - 1	PBS	-30	0.0272	-	7471B	PPB	26 Mar 2018 20:54:57	U	SMPL
PB107639BS - 1	LCSS	15302	3.6451	-	7471B	PPB	26 Mar 2018 20:57:15	U	SMPL
J1758-01 - 1	PL-02-032318-A	1403	0.3654	-	7471B	PPB	26 Mar 2018 20:59:31	U	SMPL
CCV73 - 1	CCV73	20973	4.9833	-	7471B	PPB	26 Mar 2018 21:01:48	U	SMPL
CCB73 - 1	CCB73	-95	0.0119	-	7471B	PPB	26 Mar 2018 21:04:04	U	SMPL
J1758-01DUP - 1	PL-02-032318-ADUP	1364	0.3562	-	7471B	PPB	26 Mar 2018 21:06:21	U	SMPL
J1758-01MS - 1	PL-02-032318-AMS	17117	4.0734	-	7471B	PPB	26 Mar 2018 21:08:37	U	SMPL
J1758-01MSD - 1	PL-02-032318-AMSD	17231	4.1003	-	7471B	PPB	26 Mar 2018 21:10:53	U	SMPL
J1758-03 - 1	PL-02-032318-B	1167	0.3097	-	7471B	PPB	26 Mar 2018 21:13:10	U	SMPL
J1758-05 - 1	PL-02-032318-C	1002	0.2707	-	7471B	PPB	26 Mar 2018 21:15:26	U	SMPL
J1758-07 - 1	PL-02-032318-D	1336	0.3496	-	7471B	PPB	26 Mar 2018 21:17:44	U	SMPL
J2047-01LX5 - 1		2654	0.6606	-	7471B	PPB	26 Mar 2018 21:20:01	U	SMPL
J2047-01A - 1		30392	7.2059	-	7471B	PPB	26 Mar 2018 21:22:18	U	SMPL
J1758-01LX5 - 1		114	0.0612	-	7471B	PPB	26 Mar 2018 21:24:36	U	SMPL
J1758-01A - 1		18816	4.4743	-	7471B	PPB	26 Mar 2018 21:26:53	U	SMPL
CCV74 - 1	CCV74	20705	4.9201	-	7471B	PPB	26 Mar 2018 21:29:09	U	SMPL
CCB75 - 1	CCB75	-58	0.0206	-	7471B	PPB	26 Mar 2018 21:38:15	U	SMPL

3/27/18