



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB94166

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	80.6	03/26/2018	09:05
2	J2046-02		1	>100.0	>212.0	>212.0	03/26/2018	10:02
3	J2046-02DUP		1	>100.0	>212.0	>212.0	03/26/2018	10:20
4	J2050-08		1	>100.0	>212.0	>212.0	03/26/2018	10:45
5	J2052-06		1	>100.0	>212.0	>212.0	03/26/2018	11:10
6	J2054-01		1	>100.0	>212.0	>212.0	03/26/2018	11:35
7	J2056-04		1	>100.0	>212.0	>212.0	03/26/2018	11:50
8	J2056-05		1	>100.0	>212.0	>212.0	03/26/2018	12:10
9	J2056-07		1	>100.0	>212.0	>212.0	03/26/2018	12:32
10	J2056-08		1	>100.0	>212.0	>212.0	03/26/2018	12:55
11	J2056-09		1	>100.0	>212.0	>212.0	03/26/2018	13:15
12	J2056-10		1	>100.0	>212.0	>212.0	03/26/2018	13:45
13	J2056-11		1	>100.0	>212.0	>212.0	03/26/2018	14:15
14	J2056-12		1	>100.0	>212.0	>212.0	03/26/2018	14:42
15	J2056-13		1	>100.0	>212.0	>212.0	03/26/2018	15:02
16	J2056-15		1	>100.0	>212.0	>212.0	03/26/2018	15:32
17	J2056-16		1	>100.0	>212.0	>212.0	03/26/2018	15:58
18	J2056-17		1	>100.0	>212.0	>212.0	03/26/2018	16:20
19	J2056-18		1	>100.0	>212.0	>212.0	03/26/2018	16:45
20	J2056-19		1	>100.0	>212.0	>212.0	03/26/2018	17:15
21	J2056-20		1	>100.0	>212.0	>212.0	03/26/2018	17:46
22	J2056-21		1	>100.0	>212.0	>212.0	03/26/2018	18:02

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

WORKLIST(Hardcopy Internal Chain)

LB94166

WorkList Name : FP32618

WorkList ID : 109970

Date : 3/26/2018 8:19:52 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/30/2018	Water	J2046-02	Flash Point	Cool 4 deg C	ENTAG04	K21	TW-02	03/23/2018	1010A
03/28/2018	Water	J2050-08	Flash Point	Cool 4 deg C	PSEG01	J23	TP-2-MH-C	03/23/2018	1010A
03/28/2018	Water	J2052-06	Flash Point	Cool 4 deg C	PSEG01	J21	TP-B-GW	03/23/2018	1010A
04/06/2018	Water	J2054-01	Flash Point	Cool 4 deg C	GZAG04	K61	MW-02-20180323-DEP	03/23/2018	1010A
03/30/2018	Water	J2056-04	Flash Point	Cool 4 deg C	PSEG01	J23	22146	03/23/2018	1010A
03/30/2018	Water	J2056-05	Flash Point	Cool 4 deg C	PSEG01	J23	22148	03/23/2018	1010A
03/30/2018	Water	J2056-07	Flash Point	Cool 4 deg C	PSEG01	J23	22136	03/23/2018	1010A
03/30/2018	Water	J2056-08	Flash Point	Cool 4 deg C	PSEG01	J23	22151	03/23/2018	1010A
03/30/2018	Water	J2056-09	Flash Point	Cool 4 deg C	PSEG01	J23	22152	03/23/2018	1010A
03/30/2018	Water	J2056-10	Flash Point	Cool 4 deg C	PSEG01	J23	22155	03/23/2018	1010A
03/30/2018	Water	J2056-11	Flash Point	Cool 4 deg C	PSEG01	J23	22156	03/23/2018	1010A
03/30/2018	Water	J2056-12	Flash Point	Cool 4 deg C	PSEG01	J23	22157	03/23/2018	1010A
03/30/2018	Water	J2056-13	Flash Point	Cool 4 deg C	PSEG01	J23	22159	03/23/2018	1010A
03/30/2018	Water	J2056-15	Flash Point	Cool 4 deg C	PSEG01	J23	22162	03/23/2018	1010A
03/30/2018	Water	J2056-16	Flash Point	Cool 4 deg C	PSEG01	J23	22164	03/23/2018	1010A
03/30/2018	Water	J2056-17	Flash Point	Cool 4 deg C	PSEG01	J23	22165	03/23/2018	1010A
03/30/2018	Water	J2056-18	Flash Point	Cool 4 deg C	PSEG01	J23	22167	03/23/2018	1010A
03/30/2018	Water	J2056-19	Flash Point	Cool 4 deg C	PSEG01	J23	22169	03/23/2018	1010A
03/30/2018	Water	J2056-20	Flash Point	Cool 4 deg C	PSEG01	J23	22171	03/23/2018	1010A
03/30/2018	Water	J2056-21	Flash Point	Cool 4 deg C	PSEG01	J23	22172	03/23/2018	1010A

Date/Time 3/26/18 8:30 AM
 Received by: KS
 Relinquished by: CP

Date/Time 3/26/18 18:10
 Received by: CP
 Relinquished by: PS