

1394172

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

3/27/2018 12:19

Test: Phenolics-*mg/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.015	0.0	0.105	
ICB1	0.007	0.0	0.003	
CCV1	0.984	0.0	0.101	
CCB1	0.005	0.0	0.003	
PB107671BL	0.008	0.0	0.003	
PB107671BS	1.016	0.0	0.105	
J2054-01	0.025	0.0	0.005	
J2054-01DUP	0.024	0.0	0.005	
J2054-01MS	0.883	0.0	0.091	
J2054-01MSD	0.876	0.0	0.091	
CCV2	0.966	0.0	0.100	
CCB2	0.005	0.0	0.003	

N	12
Mean	0.485
SD	0.4950
CV%	102.15

AS

Aquakem v. 7.2AQ1

Results from time period:

Tue Mar 27 11:37:02 2018

Tue Mar 27 12:08:33 2018

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0525	mg/l	3/27/2018 11:37:02	
0.1PPM	A	Phenolics-	P	0.1072	mg/l	3/27/2018 11:37:03	
0.25PPM	A	Phenolics-	P	0.2527	mg/l	3/27/2018 11:37:04	
0.50PPM	A	Phenolics-	P	0.4997	mg/l	3/27/2018 11:37:05	
1.0PPM	A	Phenolics-	P	0.9771	mg/l	3/27/2018 11:37:06	
2.0PPM	A	Phenolics-	P	2.0108	mg/l	3/27/2018 11:37:07	
ICV1	S	Phenolics-	P	1.0152	mg/l	3/27/2018 12:05:45	
ICB1	S	Phenolics-	P	0.0068	mg/l	3/27/2018 12:05:46	
CCV1	S	Phenolics-	P	0.9838	mg/l	3/27/2018 12:05:47	
CCB1	S	Phenolics-	P	0.0054	mg/l	3/27/2018 12:05:48	
PB107671BL	S	Phenolics-	P	0.0085	mg/l	3/27/2018 12:05:49	
PB107671BS	S	Phenolics-	P	1.0159	mg/l	3/27/2018 12:05:50	
J2054-01	S	Phenolics-	P	0.0247	mg/l	3/27/2018 12:05:51	
J2054-01DUP	S	Phenolics-	P	0.0245	mg/l	3/27/2018 12:05:52	
J2054-01MS	S	Phenolics-	P	0.8828	mg/l	3/27/2018 12:05:53	
J2054-01MSD	S	Phenolics-	P	0.8765	mg/l	3/27/2018 12:05:54	
CCV2	S	Phenolics-	P	0.9658	mg/l	3/27/2018 12:05:55	
CCB2	S	Phenolics-	P	0.0047	mg/l	3/27/2018 12:05:56	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

3/27/2018 11:42


Test Phenolics-

Accepted 3/27/2018 11:42

Factor 9.936
 Bias 0.002

Coeff. of det. 0.999747

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0525	0.0500	
2	Phenol-2	0.013	0.1072	0.1000	
3	Phenol-2	0.028	0.2527	0.2500	
4	Phenol-2	0.053	0.4997	0.5000	
5	Phenol-2	0.101	0.9771	1.0000	
6	Phenol-2	0.205	2.0108	2.0000	