

245.1


LB94175

INSTRUMENT ID: CV2

Linear

 μ Abs.:

34015



A= 0.0000e+000

B= 2.9657e-004

C= -2.0496e-002

Rho= 0.9997861

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.004	-0.004	57	0.000	57				
0.05	0.050	0.042	-0.008	212	0.0 %	212				
0.2	0.200	0.195	-0.005	727	0.0 %	727				
2.5	2.500	2.518	0.018	8559	0.0 %	8559				
5.0	5.000	5.096	0.096	17252	0.0 %	17252				
7.5	7.500	7.335	-0.165	24803	0.0 %	24803				
10.0	10.000	10.067	0.067	34015	0.0 %	34015				

LB94175 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 27 Mar 2018 14:22:35

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	57	-	0.0000	245.1	PPB	27 Mar 2018 14:31:31	S	Std
0.05 - 1	50.05	212	-	0.0500	245.1	PPB	27 Mar 2018 14:33:40	S	Std
0.2 - 1	50.2	727	-	0.2000	245.1	PPB	27 Mar 2018 14:35:48	S	Std
2.5 - 1	52.5	8559	-	2.5000	245.1	PPB	27 Mar 2018 14:37:56	S	Std
5.0 - 1	55	17252	-	5.0000	245.1	PPB	27 Mar 2018 14:40:04	S	Std
7.5 - 1	57.5	24803	-	7.5000	245.1	PPB	27 Mar 2018 14:46:47	S	Std
10.0 - 1	510	34015	-	10.0000	245.1	PPB	27 Mar 2018 14:48:55	S	Std
ICV63 - 1	ICV63	13546	3.9968	-245.1		PPB	27 Mar 2018 14:54:19	U	SMPL
ICB63 - 1	ICB63	-117	-0.0552	-245.1		PPB	27 Mar 2018 14:56:25	U	SMPL
CCV75 - 1	CCV75	17643	5.2118	-245.1		PPB	27 Mar 2018 14:58:32	U	SMPL
CCB75 - 1	CCB75	-75	-0.0427	-245.1		PPB	27 Mar 2018 15:00:39	U	SMPL
CRA - 1	CRA	217	0.0439	-245.1		PPB	27 Mar 2018 15:02:47	U	SMPL
HighStd - 1	HighStd	34891	10.3270	-245.1		PPB	27 Mar 2018 15:04:53	U	SMPL
ChkStd - 1	ChkStd	23290	6.8865	-245.1		PPB	27 Mar 2018 15:16:17	U	SMPL
PB107665BL - 1	PBW	-153	-0.0659	-245.1		PPB	27 Mar 2018 15:18:24	U	SMPL
PB107665BS - 1	LCSW	13634	4.0229	-245.1		PPB	27 Mar 2018 15:20:33	U	SMPL
J2038-01 - 1	032218-JETB-F-D-M	-102	-0.0507	-245.1		PPB	27 Mar 2018 15:22:40	U	SMPL
J2038-02 - 1	032218-JETB-F-U-M	-18	-0.0258	-245.1		PPB	27 Mar 2018 15:24:49	U	SMPL
J2038-03 - 1	032218-TIDO-F-D-B	-2	-0.0211	-245.1		PPB	27 Mar 2018 15:26:57	U	SMPL
J2038-04 - 1	032218-TIDO-F-D-S	10	-0.0175	-245.1		PPB	27 Mar 2018 15:29:05	U	SMPL
J2038-05 - 1	032218-TIDO-F-U-B	31	-0.0113	-245.1		PPB	27 Mar 2018 15:31:12	U	SMPL
CCV76 - 1	CCV76	17770	5.2495	-245.1		PPB	27 Mar 2018 15:33:19	U	SMPL
CCB76 - 1	CCB76	-135	-0.0605	-245.1		PPB	27 Mar 2018 15:35:26	U	SMPL
J2038-06 - 1	032218-TIDO-F-U-S	-26	-0.0282	-245.1		PPB	27 Mar 2018 15:37:34	U	SMPL
J2053-01 - 1	MW-01-20180323	293	0.0664	-245.1		PPB	27 Mar 2018 15:39:41	U	SMPL
J2053-02 - 1	MW-02-20180323	2460	0.7091	-245.1		PPB	27 Mar 2018 15:41:47	U	SMPL
J2053-03 - 1	MW-03-20180323	239	0.0504	-245.1		PPB	27 Mar 2018 15:43:54	U	SMPL
J2053-04 - 1	MW-04-20180323	111	0.0124	-245.1		PPB	27 Mar 2018 15:46:02	U	SMPL
J2054-01 - 1	MW-02-20180323-DEP	1128	0.3140	-245.1		PPB	27 Mar 2018 15:52:47	U	SMPL
J2054-01DUP - 1	MW-02-20180323-DEPDUP	1118	0.3111	-245.1		PPB	27 Mar 2018 15:54:55	U	SMPL
J2054-01MS - 1	MW-02-20180323-DEPMS	15876	4.6878	-245.1		PPB	27 Mar 2018 15:57:03	U	SMPL
J2054-01MSD - 1	MW-02-20180323-DEPMSD	15488	4.5727	-245.1		PPB	27 Mar 2018 15:59:12	U	SMPL
J2061-01 - 1	032418-JETT-E-D-B	-15	-0.0249	-245.1		PPB	27 Mar 2018 16:01:19	U	SMPL
CCV77 - 1	CCV77	17597	5.1982	-245.1		PPB	27 Mar 2018 16:03:26	U	SMPL
CCB77 - 1	CCB77	-105	-0.0516	-245.1		PPB	27 Mar 2018 16:05:33	U	SMPL
J2061-02 - 1	032418-JETT-E-D-M	60	-0.0027	-245.1		PPB	27 Mar 2018 16:07:41	U	SMPL
J2061-03 - 1	032418-JETT-E-D-S	54	-0.0045	-245.1		PPB	27 Mar 2018 16:09:49	U	SMPL
J2061-04 - 1	032418-JETT-E-U-B	70	0.0003	-245.1		PPB	27 Mar 2018 16:11:56	U	SMPL
J2061-05 - 1	032418-JETT-E-U-M	86	0.0050	-245.1		PPB	27 Mar 2018 16:14:03	U	SMPL
J2061-06 - 1	032418-JETT-E-U-S	60	-0.0027	-245.1		PPB	27 Mar 2018 16:16:10	U	SMPL
J2054-01LX5 - 1		238	0.0501	-245.1		PPB	27 Mar 2018 16:18:18	U	SMPL
J2054-01A - 1		16949	5.0060	-245.1		PPB	27 Mar 2018 16:20:25	U	SMPL
CCV78 - 1	CCV78	17456	5.1564	-245.1		PPB	27 Mar 2018 16:22:34	U	SMPL
CCB78 - 1	CCB78	-83	-0.0451	-245.1		PPB	27 Mar 2018 16:24:42	U	SMPL