



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: AparnaDate: 03/28/2018DATE & TIME IN: 03/28/2018 08:00OVEN TEMP: 103 °CRun Number: LB94190DATE & TIME OUT: 03/28/2018 12:00OVEN TEMP: 105 °CBalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB94190BL	LB94190BL	1.3962	1.3962	100	1.3962	1.3962	1.3962	0	0
2	LB94190BS	LB94190BS	1.3998	1.3998	100	1.4495	1.4495	1.4495	0.0497	497
3	J2076-01	EFF-180326-05	1.3998	1.3998	900	1.4174	1.4174	1.4174	0.0176	19.6

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS032818

WorkList ID : 110074

Date : 3/28/2018 7:18:29 AM

LB94190

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/28/2018	Water	J2076-01	TSS	Cool 4 deg C	VIAS01	K23	EFF-180326-05	03/26/2018	SM2540 D

Date/Time 03/28/18, 07:40
 Received by: AS
 Relinquished by: CP

Date/Time 03/28/18, 08:10
 Received by: CP
 Relinquished by: AS