


7471B

LB 94197
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42760



A= 0.0000e+000

B= 2.3609e-004

C= -2.3154e-002

Rho= 0.9999214

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.017	0.017	170	0.000	170				
0.2	0.200	0.227	0.027	1061	0.0 %	1061				
2.5	2.500	2.488	-0.012	10637	0.0 %	10637				
5.0	5.000	4.970	-0.030	21150	0.0 %	21150				
7.5	7.500	7.425	-0.075	31549	0.0 %	31549				
10	10.000	10.072	0.072	42760	0.0 %	42760				

LB94197 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 28 Mar 2018 11:29:20				
Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type
0.0 - 1	50	170	-	0.0000	7471B	PPB	28 Mar 2018 11:43:58	S Std
0.2 - 1	50.2	1061	-	0.2000	7471B	PPB	28 Mar 2018 11:46:16	S Std
2.5 - 1	52.5	10637	-	2.5000	7471B	PPB	28 Mar 2018 11:48:33	S Std
5.0 - 1	55	21150	-	5.0000	7471B	PPB	28 Mar 2018 11:50:50	S Std
7.5 - 1	57.5	31549	-	7.5000	7471B	PPB	28 Mar 2018 11:53:07	S Std
10 - 1	510	42760	-	10.0000	7471B	PPB	28 Mar 2018 11:57:57	S Std
ICV64 - 1	ICV64	16696	3.9186	-	7471B	PPB	28 Mar 2018 12:02:16	U SMPL
ICB64 - 1	ICB64	42	-0.0132	-	7471B	PPB	28 Mar 2018 12:04:31	U SMPL
CCV79 - 1	CCV79	20310	4.7719	-	7471B	PPB	28 Mar 2018 12:06:47	U SMPL
CCB79 - 1	CCB79	23	-0.0177	-	7471B	PPB	28 Mar 2018 12:09:03	U SMPL
CRA - 1	CRA	1071	0.2297	-	7471B	PPB	28 Mar 2018 12:11:19	U SMPL
HighStd - 1	HighStd	43347	10.2107	-	7471B	PPB	28 Mar 2018 12:13:35	U SMPL
ChkStd - 1	ChkStd	29103	6.8478	-	7471B	PPB	28 Mar 2018 12:15:51	U SMPL
PB107719BL - 1	PBS	-60	-0.0373	-	7471B	PPB	28 Mar 2018 12:18:09	U SMPL
PB107719BS - 1	LCSS	15009	3.5203	-	7471B	PPB	28 Mar 2018 12:20:27	U SMPL
J1749-05 - 1	BU-03-032618	1811	0.4044	-	7471B	PPB	28 Mar 2018 12:30:23	U SMPL
J1749-05DUP - 1	BU-03-032618DUP	1566	0.3466	-	7471B	PPB	28 Mar 2018 12:32:39	U SMPL
J1749-05MS - 1	BU-03-032618MS	18801	4.4156	-	7471B	PPB	28 Mar 2018 12:34:57	U SMPL
J1749-05MSD - 1	BU-03-032618MSD	17476	4.1028	-	7471B	PPB	28 Mar 2018 12:37:14	U SMPL
J1754-11 - 1	CL-02-032718-A	2077	0.4672	-	7471B	PPB	28 Mar 2018 12:39:30	U SMPL
CCV80 - 1	CCV80	20837	4.8963	-	7471B	PPB	28 Mar 2018 12:41:46	U SMPL
CCB80 - 1	CCB80	61	-0.0088	-	7471B	PPB	28 Mar 2018 12:44:02	U SMPL
J1754-13 - 1	CL-02-032718-B	2664	0.6058	-	7471B	PPB	28 Mar 2018 12:46:18	U SMPL
J1754-15 - 1	CL-02-032718-C	2256	0.5095	-	7471B	PPB	28 Mar 2018 12:48:34	U SMPL
J1754-17 - 1	CL-02-032718-D	2708	0.6162	-	7471B	PPB	28 Mar 2018 12:50:50	U SMPL
J1754-19 - 1	CL-02-032718-E	5126	1.1870	-	7471B	PPB	28 Mar 2018 12:53:07	U SMPL
J1756-11 - 1	OR-01-032718-A	4217	0.9724	-	7471B	PPB	28 Mar 2018 12:55:23	U SMPL
J1756-13 - 1	OR-01-032718-B	1934	0.4334	-	7471B	PPB	28 Mar 2018 12:57:40	U SMPL
J2063-01 - 1	RO-CV-2533	41588	9.7954	-	7471B	PPB	28 Mar 2018 12:59:57	U SMPL
J2064-01 - 1	RO-CV-2532	4641	1.0725	-	7471B	PPB	28 Mar 2018 13:02:15	U SMPL
J2065-01 - 1	RO-CV-2536	1713	0.3813	-	7471B	PPB	28 Mar 2018 13:04:35	U SMPL
J2066-01 - 1	RO-CV-2036	15534	3.6443	-	7471B	PPB	28 Mar 2018 13:06:51	U SMPL
CCV81 - 1	CCV81	20892	4.9093	-	7471B	PPB	28 Mar 2018 13:09:07	U SMPL
CCB81 - 1	CCB81	65	-0.0078	-	7471B	PPB	28 Mar 2018 13:11:23	U SMPL
J2068-01 - 1	RO-CV-2043	8486	1.9803	-	7471B	PPB	28 Mar 2018 13:13:39	U SMPL
J2071-01 - 1	TP-1-MH-D-S	35046	8.2509	-	7471B	PPB	28 Mar 2018 13:15:55	U SMPL
J2071-04 - 1	TP-1-MH-D-D	33611	7.9121	-	7471B	PPB	28 Mar 2018 13:18:12	U SMPL
J2071-07 - 1	TP-7-DB	47680	11.2337	-	7471B	PPB	28 Mar 2018 13:20:29	U SMPL
J2075-01 - 1	SB-107(4-5)	1664	0.3697	-	7471B	PPB	28 Mar 2018 13:22:47	U SMPL
J2075-02 - 1	SB-107(12-13)	234	0.0321	-	7471B	PPB	28 Mar 2018 13:25:06	U SMPL
J2075-03 - 1	MW-103(2-3)	356	0.0609	-	7471B	PPB	28 Mar 2018 13:27:23	U SMPL
J2075-04 - 1	MW-103(14-15)	1485	0.3274	-	7471B	PPB	28 Mar 2018 13:29:41	U SMPL
PB107741BL - 1	PBS	49	-0.0116	-	7471B	PPB	28 Mar 2018 13:31:59	U SMPL
PB107741BS - 1	LCSS	15226	3.5716	-	7471B	PPB	28 Mar 2018 13:34:15	U SMPL
CCV82 - 1	CCV82	20936	4.9196	-	7471B	PPB	28 Mar 2018 13:36:31	U SMPL
CCB82 - 1	CCB82	60	-0.0090	-	7471B	PPB	28 Mar 2018 13:38:47	U SMPL
J2050-04DLX10 - 1	TP-3-DB	18301	4.2975	-	7471B	PPB	28 Mar 2018 13:41:03	U SMPL
J2085-01 - 1	WO-RO-1499	3671	0.8435	-	7471B	PPB	28 Mar 2018 13:43:20	U SMPL
J2086-01 - 1	RS-124023-124035	2641	0.6004	-	7471B	PPB	28 Mar 2018 13:45:37	U SMPL
J2087-01 - 1	JRS-RB201719RT-CV-2530	643	0.1287	-	7471B	PPB	28 Mar 2018 13:47:53	U SMPL
J2088-01 - 1	TP-2	180	0.0193	-	7471B	PPB	28 Mar 2018 13:50:10	U SMPL
J2088-04 - 1	TP-1	309	0.0498	-	7471B	PPB	28 Mar 2018 13:52:27	U SMPL
J2092-01 - 1	FK-BPM-01-B02-B02A-C02-C02A	1271	0.2769	-	7471B	PPB	28 Mar 2018 13:54:43	U SMPL
J2092-01DUP - 1	FK-BPM-01-B02-B02A-C02-C02ADUP	1321	0.2887	-	7471B	PPB	28 Mar 2018 13:57:01	U SMPL
J2092-01MS - 1	FK-BPM-01-B02-B02A-C02-C02AMS	16864	3.9583	-	7471B	PPB	28 Mar 2018 13:59:18	U SMPL
J2092-01MSD - 1	FK-BPM-01-B02-B02A-C02-C02AMSD	16992	3.9885	-	7471B	PPB	28 Mar 2018 14:01:34	U SMPL
CCV83 - 1	CCV83	20633	4.8481	-	7471B	PPB	28 Mar 2018 14:03:51	U SMPL
CCB83 - 1	CCB83	53	-0.0106	-	7471B	PPB	28 Mar 2018 14:06:07	U SMPL
J2093-01 - 1	FK-BPM-01-G0-G0A-G01-G01A	1334	0.2918	-	7471B	PPB	28 Mar 2018 14:08:23	U SMPL
J2096-01 - 1	CS-T2	15008	3.5201	-	7471B	PPB	28 Mar 2018 14:10:39	U SMPL
J1749-05LX5 - 1		342	0.0576	-	7471B	PPB	28 Mar 2018 14:12:56	U SMPL
J1749-05A - 1		18685	4.3882	-	7471B	PPB	28 Mar 2018 14:15:12	U SMPL
J2092-01LX5 - 1		275	0.0418	-	7471B	PPB	28 Mar 2018 14:17:29	U SMPL
J2092-01A - 1		18032	4.2340	-	7471B	PPB	28 Mar 2018 14:19:46	U SMPL
J2071-07DLX2 - 1		22643	5.3227	-	7471B	PPB	28 Mar 2018 14:22:04	U SMPL
CCV84 - 1	CCV84	20617	4.8443	-	7471B	PPB	28 Mar 2018 14:24:21	U SMPL
CCB84 - 1	CCB84	51	-0.0111	-	7471B	PPB	28 Mar 2018 14:26:39	U SMPL