

7470A


LB 94207

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36060



A= 0.0000e+000

B= 2.8008e-004

C= -2.2549e-003

Rho= 0.9997584

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	19	0.000	19				
0.2	0.200	0.228	0.028	822	0.0 %	822				
2.5	2.500	2.487	-0.013	8889	0.0 %	8889				
5.0	5.000	5.048	0.048	18030	0.0 %	18030				
7.5	7.500	7.336	-0.164	26202	0.0 %	26202				
10.0	10.000	10.097	0.097	36060	0.0 %	36060				

LB94207 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Mar 2018 15:43:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	19	-	0.00007470A	PPB		28 Mar 2018 15:45:52	S	Std
0.2 - 1	50.2	822	-	0.20007470A	PPB		28 Mar 2018 15:48:00	S	Std
2.5 - 1	52.5	8889	-	2.50007470A	PPB		28 Mar 2018 15:50:08	S	Std
5.0 - 1	55	18030	-	5.00007470A	PPB		28 Mar 2018 15:52:16	S	Std
7.5 - 1	57.5	26202	-	7.50007470A	PPB		28 Mar 2018 15:54:24	S	Std
10.0 - 1	510	36060	-	10.00007470A	PPB		28 Mar 2018 15:56:32	S	Std
ICV65 - 1	ICV65	13454	3.7660	-7470A	PPB		28 Mar 2018 15:59:21	U	SMPL
ICB65 - 1	ICB65	61	0.0148	-7470A	PPB		28 Mar 2018 16:01:27	U	SMPL
CCV85 - 1	CCV85	18090	5.0644	-7470A	PPB		28 Mar 2018 16:03:34	U	SMPL
CCB85 - 1	CCB85	45	0.0103	-7470A	PPB		28 Mar 2018 16:05:41	U	SMPL
CRA - 1	CRA	792	0.2196	-7470A	PPB		28 Mar 2018 16:07:49	U	SMPL
HighStd - 1	HighStd	36326	10.1720	-7470A	PPB		28 Mar 2018 16:09:55	U	SMPL
ChkStd - 1	ChkStd	24024	6.7264	-7470A	PPB		28 Mar 2018 16:12:02	U	SMPL
PB107716BL - 1	PBW	-82	-0.0252	-7470A	PPB		28 Mar 2018 16:14:11	U	SMPL
PB107716BS - 1	LCSW	14646	4.0998	-7470A	PPB		28 Mar 2018 16:16:20	U	SMPL
PB107716TB - 1	PB107673TB	18	0.0028	-7470A	PPB		28 Mar 2018 16:18:27	U	SMPL
J1749-06 - 1	BU-03-032618	77	0.0193	-7470A	PPB		28 Mar 2018 16:20:35	U	SMPL
J2056-01 - 1	22174	136	0.0358	-7470A	PPB		28 Mar 2018 16:22:43	U	SMPL
J2056-02 - 1	22250-51	87	0.0221	-7470A	PPB		28 Mar 2018 16:24:52	U	SMPL
J2062-01 - 1	19642	55	0.0131	-7470A	PPB		28 Mar 2018 16:26:58	U	SMPL
CCV86 - 1	CCV86	17920	5.0168	-7470A	PPB		28 Mar 2018 16:29:05	U	SMPL
CCB86 - 1	CCB86	48	0.0112	-7470A	PPB		28 Mar 2018 16:31:12	U	SMPL
J2062-02 - 1	14717	92	0.0235	-7470A	PPB		28 Mar 2018 16:33:20	U	SMPL
J2063-03 - 1	RO-CV-2533	139	0.0367	-7470A	PPB		28 Mar 2018 16:35:27	U	SMPL
J2064-03 - 1	RO-CV-2532	83	0.0210	-7470A	PPB		28 Mar 2018 16:37:33	U	SMPL
J2065-03 - 1	RO-CV-2536	245	0.0664	-7470A	PPB		28 Mar 2018 16:39:40	U	SMPL
J2066-03 - 1	RO-CV-2036	177	0.0473	-7470A	PPB		28 Mar 2018 16:41:47	U	SMPL
J2068-03 - 1	RO-CV-2043	185	0.0496	-7470A	PPB		28 Mar 2018 16:43:54	U	SMPL
J2071-03 - 1	TP-1-MH-D-S	149	0.0395	-7470A	PPB		28 Mar 2018 16:46:03	U	SMPL
J2071-06 - 1	TP-1-MH-D-D	34	0.0073	-7470A	PPB		28 Mar 2018 16:48:11	U	SMPL
J2071-09 - 1	TP-7-DB	193	0.0518	-7470A	PPB		28 Mar 2018 16:50:19	U	SMPL
J2071-09DUP - 1	TP-7-DBDUP	222	0.0599	-7470A	PPB		28 Mar 2018 16:52:26	U	SMPL
CCV87 - 1	CCV87	17917	5.0160	-7470A	PPB		28 Mar 2018 16:54:33	U	SMPL
CCB87 - 1	CCB87	-13	-0.0059	-7470A	PPB		28 Mar 2018 16:56:40	U	SMPL
J2071-09MS - 1	TP-7-DBMS	13835	3.8727	-7470A	PPB		28 Mar 2018 16:58:48	U	SMPL
J2071-09MSD - 1	TP-7-DBMSD	13497	3.7780	-7470A	PPB		28 Mar 2018 17:00:55	U	SMPL
PB107717BL - 1	PBW	-5	-0.0037	-7470A	PPB		28 Mar 2018 17:03:02	U	SMPL
PB107717BS - 1	LCSW	13373	3.7433	-7470A	PPB		28 Mar 2018 17:12:39	U	SMPL
PB107717TB - 1	PB107675TB	23	0.0042	-7470A	PPB		28 Mar 2018 17:14:46	U	SMPL
J2062-05 - 1	19626-7-8-74	655	0.1812	-7470A	PPB		28 Mar 2018 17:16:53	U	SMPL
J2062-10 - 1	19643	89	0.0227	-7470A	PPB		28 Mar 2018 17:19:01	U	SMPL
J2062-10DUP - 1	19643DUP	113	0.0294	-7470A	PPB		28 Mar 2018 17:21:10	U	SMPL
J2062-10MS - 1	19643MS	13104	3.6679	-7470A	PPB		28 Mar 2018 17:23:19	U	SMPL
J2062-10MSD - 1	19643MSD	13491	3.7763	-7470A	PPB		28 Mar 2018 17:25:26	U	SMPL
CCV88 - 1	CCV88	17724	4.9619	-7470A	PPB		28 Mar 2018 17:27:33	U	SMPL
CCB88 - 1	CCB88	-136	-0.0403	-7470A	PPB		28 Mar 2018 17:29:40	U	SMPL
J2062-11 - 1	19673	149	0.0395	-7470A	PPB		28 Mar 2018 17:31:48	U	SMPL
J2067-02 - 1	RO-CV-RB47185RT	118	0.0308	-7470A	PPB		28 Mar 2018 17:33:55	U	SMPL
J2071-11 - 1	TP-7-DB	102	0.0263	-7470A	PPB		28 Mar 2018 17:36:02	U	SMPL
PB107718BL - 1	PBW	143	0.0378	-7470A	PPB		28 Mar 2018 17:38:09	U	SMPL
PB107718BS - 1	LCSW	14276	3.9962	-7470A	PPB		28 Mar 2018 17:40:17	U	SMPL
J2067-01 - 1	RO-CV-RB47185RT	456	0.1255	-7470A	PPB		28 Mar 2018 17:42:25	U	SMPL
J2067-01DUP - 1	RO-CV-RB47185RTDUP	513	0.1414	-7470A	PPB		28 Mar 2018 17:44:32	U	SMPL
J2067-01MS - 1	RO-CV-RB47185RTMS	12227	3.4223	-7470A	PPB		28 Mar 2018 17:46:40	U	SMPL
J2067-01MSD - 1	RO-CV-RB47185RTMSD	12397	3.4699	-7470A	PPB		28 Mar 2018 17:48:49	U	SMPL
J2071-10 - 1	TP-7-DB	48586	13.6058	-7470A	PPB		28 Mar 2018 17:50:57	U	SMPL
CCV89 - 1	CCV89	17463	4.8888	-7470A	PPB		28 Mar 2018 17:53:05	U	SMPL
CCB89 - 1	CCB89	-101	-0.0305	-7470A	PPB		28 Mar 2018 17:55:14	U	SMPL
J2071-09LX5 - 1		20	0.0033	-7470A	PPB		28 Mar 2018 18:19:16	U	SMPL
J2071-09A - 1		13700	3.8349	-7470A	PPB		28 Mar 2018 18:21:23	U	SMPL
J2062-10LX5 - 1		-74	-0.0230	-7470A	PPB		28 Mar 2018 18:23:31	U	SMPL
J2062-10A - 1		13548	3.7923	-7470A	PPB		28 Mar 2018 18:25:39	U	SMPL
J2067-01LX5 - 1		-27	-0.0098	-7470A	PPB		28 Mar 2018 18:27:46	U	SMPL
J2067-01A - 1		11689	3.2716	-7470A	PPB		28 Mar 2018 18:29:53	U	SMPL
J2071-10DLX2 - 1		22273	6.2360	-7470A	PPB		28 Mar 2018 18:32:00	U	SMPL
J2071-10DLX2 - 1		17442	4.8829	-7470A	PPB		28 Mar 2018 18:36:19	U	SMPL
J2071-10DLX2 - 1		0	-0.0023	-7470A	PPB		28 Mar 2018 18:38:27	U	SMPL

203
3/28/18