



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB94212

BalanceID: WC SC-4

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP62779
COD calibration std. 100 ppm	WP62777
COD calibration std. 10 ppm	WP62776
COD calibration std. 150 ppm	WP62778
COD calibration std. 50 ppm	WP62781
COD calibration std. 0 ppm	WP62775
COD Digestion Vials Low Level 0-150Mg/L	W2291
COD ICV-LCS std, 50ppm	WP63237
COD CCV std, 50ppm	WP63238

Temp In (C): <u>150</u>	Date In: <u>04/02/2018</u>	Time In: <u>13:00</u>
Temp Out (C): <u>150</u>	Date Out: <u>04/02/2018</u>	Time Out: <u>15:01</u>

Intercept: 0.3627

Slope: 1.0071

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	03/16/2018	12:30
2	CAL2	10	1	Water	11.000	03/16/2018	12:30
3	CAL3	50	1	Water	51.000	03/16/2018	12:31
4	CAL4	100	1	Water	100.000	03/16/2018	12:31
5	CAL5	150	1	Water	152.000	03/16/2018	12:32



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB94212

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.287	03/16/2018	12:32
2	ICB				1	Water	0.000	-0.360	04/02/2018	12:33
3	CCV1	50			1	Water	51.000	50.280	04/02/2018	16:10
4	CCB1				1	Water	0.000	-0.360	04/02/2018	16:10
5	LB94212BL		1.01	50	1	Solid	0.000	-0.360	04/02/2018	16:11
6	LB94212BS	50	1.00	50	1	Solid	50.000	49.287	04/02/2018	16:11
7	J1161-01		1.03	50	1	Solid	5.000	4.605	04/02/2018	16:12
8	J1161-03		1.00	50	1	Solid	4.000	3.612	04/02/2018	16:12
9	CCV2	50			1	Water	51.000	50.280	04/02/2018	16:13
10	CCB2				1	Water	0.000	-0.360	04/03/2018	16:13