



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94218

Temp.: 21.3 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/28/2018	17:30
2	CAL2	1	Water			7.01	03/28/2018	17:31
3	CAL3	1	Water			10.02	03/28/2018	17:32
4	ICV	1	Water			6.99	03/28/2018	17:34
5	CCV1	1	Water			2.01	03/28/2018	17:35
6	J1749-08	1	Solid	20.02	20	7.10	03/28/2018	17:36
7	J1749-08DUP	1	Solid	20.02	20	7.11	03/28/2018	17:37
8	J1755-06	1	Solid	20.03	20	9.03	03/28/2018	17:39
9	J1761-04	1	Solid	20.01	20	8.55	03/28/2018	17:41
10	J2103-04	1	Solid	20.04	20	5.25	03/28/2018	17:43
11	J2103-08	1	Solid	20.02	20	5.35	03/28/2018	17:45
12	J2117-02	1	Solid	20.03	20	8.47	03/28/2018	17:47
13	J2117-04	1	Solid	20.05	20	7.76	03/28/2018	17:48
14	CCV2	1	Water			12.02	03/28/2018	17:50

WORKLIST(Hardcopy Internal Chain)

1894218

WorkList Name : corrosivity j2117			WorkList ID : 110131		Date : 3/28/2018 5:22:01 PM				
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/28/2018	Solid	J1749-08	Corrosivity	Cool 4 deg C	PSEG05		BU-02-032818	03/28/2018	9045C
03/28/2018	Solid	J1755-06	Corrosivity	Cool 4 deg C	PSEG05		OK-02-032818	03/28/2018	9045C
03/28/2018	Solid	J1761-04	Corrosivity	Cool 4 deg C	PSEG05		TR-04-032818	03/28/2018	9045C
03/28/2018	Solid	J2103-04	Corrosivity	Cool 4 deg C	PSEG01	J53	LS-SOIL-1	03/28/2018	9045C
03/28/2018	Solid	J2103-08	Corrosivity	Cool 4 deg C	PSEG01	J53	LS-SOIL-2	03/28/2018	9045C
03/28/2018	Solid	J2117-02	Corrosivity	Cool 4 deg C	PSEG01	J62	TPIAS	03/28/2018	9045C
03/28/2018	Solid	J2117-04	Corrosivity	Cool 4 deg C	PSEG01	J62	TPIAD	03/28/2018	9045C

Date/Time 03/28/18 5:00 PM
 Received by: JD
 Relinquished by: CQ

Date/Time 03/28/18 6:30 PM
 Received by: CQ
 Relinquished by: JP