



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB94226

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/29/2018	10:25
2	CAL2	1	Water			7.00	03/29/2018	10:27
3	CAL3	1	Water			10.03	03/29/2018	10:29
4	ICV	1	Water			7.00	03/29/2018	10:31
5	CCV1	1	Water			2.00	03/29/2018	10:33
6	J2121-21	1	Solid	20.01	20	7.84	03/29/2018	10:35
7	J2121-21DUP	1	Solid	20.03	20	7.85	03/29/2018	10:37
8	J2121-22	1	Solid	20.05	20	6.22	03/29/2018	10:39
9	J2121-23	1	Solid	20.01	20	7.94	03/29/2018	10:41
10	J2121-24	1	Solid	20.03	20	8.04	03/29/2018	10:43
11	J2121-25	1	Solid	20.01	20	7.60	03/29/2018	10:45
12	J2121-26	1	Solid	20	20	8.22	03/29/2018	10:47
13	J2121-27	1	Solid	20.01	20	8.12	03/29/2018	10:49
14	J2121-28	1	Solid	20.03	20	7.63	03/29/2018	10:51
15	J2121-29	1	Solid	20.05	20	8.54	03/29/2018	10:53
16	CCV2	1	Water			12.00	03/29/2018	10:55
17	J2121-30	1	Solid	20.02	20	5.60	03/29/2018	10:57
18	J2122-13	1	Solid	20	20	8.30	03/29/2018	10:59
19	J2122-14	1	Solid	20.03	20	8.60	03/29/2018	11:02
20	J2122-15	1	Solid	20.01	20	8.24	03/29/2018	11:04

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	J2122-16	1	Solid	20.05	20	7.80	03/29/2018	11:06
22	J2122-17	1	Solid	20.01	20	8.40	03/29/2018	11:08
23	J2122-18	1	Solid	20.02	20	7.90	03/29/2018	11:10
24	J2122-18DUP	1	Solid	20.01	20	7.91	03/29/2018	11:12
25	CCV3	1	Water			2.00	03/29/2018	11:15

WORKLIST(Hardcopy Internal Chain)

LB94226

WorkList Name : Corrosivity032918

WorkList ID : 110175

Date : 3/29/2018 10:14:59 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/26/2018	Solid	J2121-21	Corrosivity	Cool 4 deg C	LIR001	K23	PA-01-COMP	03/26/2018	9045C
03/26/2018	Solid	J2121-22	Corrosivity	Cool 4 deg C	LIR001	K23	PA-02-COMP	03/26/2018	9045C
03/26/2018	Solid	J2121-23	Corrosivity	Cool 4 deg C	LIR001	K23	PA-03-COMP	03/26/2018	9045C
03/26/2018	Solid	J2121-24	Corrosivity	Cool 4 deg C	LIR001	K23	PA-05-COMP	03/26/2018	9045C
03/26/2018	Solid	J2121-25	Corrosivity	Cool 4 deg C	LIR001	K23	PA-06-COMP	03/26/2018	9045C
03/26/2018	Solid	J2121-26	Corrosivity	Cool 4 deg C	LIR001	K23	PA-07-COMP	03/26/2018	9045C
03/26/2018	Solid	J2121-27	Corrosivity	Cool 4 deg C	LIR001	K23	PA-10-COMP	03/26/2018	9045C
03/27/2018	Solid	J2121-28	Corrosivity	Cool 4 deg C	LIR001	K23	PA-04-COMP	03/27/2018	9045C
03/27/2018	Solid	J2121-29	Corrosivity	Cool 4 deg C	LIR001	K23	PA-08-COMP	03/27/2018	9045C
03/27/2018	Solid	J2121-30	Corrosivity	Cool 4 deg C	LIR001	K23	PA-09-COMP	03/27/2018	9045C
03/27/2018	Solid	J2122-13	Corrosivity	Cool 4 deg C	LIR001	K23	PB-12-COMP	03/27/2018	9045C
03/27/2018	Solid	J2122-14	Corrosivity	Cool 4 deg C	LIR001	K23	PB-11-COMP	03/27/2018	9045C
03/27/2018	Solid	J2122-15	Corrosivity	Cool 4 deg C	LIR001	K23	PB-14-COMP	03/27/2018	9045C
03/27/2018	Solid	J2122-16	Corrosivity	Cool 4 deg C	LIR001	K23	PB-13-COMP	03/27/2018	9045C
03/27/2018	Solid	J2122-17	Corrosivity	Cool 4 deg C	LIR001	K23	PB-16-COMP	03/27/2018	9045C
03/27/2018	Solid	J2122-18	Corrosivity	Cool 4 deg C	LIR001	K23	PB-15-COMP	03/27/2018	9045C

Date/Time 03/29/18 10:20 AM

Received by: SP

Relinquished by: CP

Date/Time 03/29/18 11:20 AM

Received by: SP

Relinquished by: CP