


7471B

LB94239
INSTRUMENT ID: CV2

Linear

 μ Abs.:

42659



A= 0.0000e+000

B= 2.3283e-004

C= 3.1884e-002

Rho= 0.9998843

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.041	0.041	40	0.000	40				
0.2	0.200	0.215	0.015	785	0.0 %	785				
2.5	2.500	2.399	-0.101	10167	0.0 %	10167				
5.0	5.000	5.007	0.007	21369	0.0 %	21369				
7.5	7.500	7.574	0.074	32393	0.0 %	32393				
10	10.000	9.964	-0.036	42659	0.0 %	42659				

LB94239 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 Mar 2018 15:35:53

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	40	-	0.0000	7471B	PPB	29 Mar 2018 15:38:21	S	Std
0.2 - 1	50.2	785	-	0.2000	7471B	PPB	29 Mar 2018 15:40:38	S	Std
2.5 - 1	52.5	10167	-	2.5000	7471B	PPB	29 Mar 2018 15:42:56	S	Std
5.0 - 1	55	21369	-	5.0000	7471B	PPB	29 Mar 2018 15:45:13	S	Std
7.5 - 1	57.5	32393	-	7.5000	7471B	PPB	29 Mar 2018 15:47:30	S	Std
10 - 1	510	42659	-	10.0000	7471B	PPB	29 Mar 2018 15:49:48	S	Std
ICV67 - 1	ICV67	16111	3.7830	-	7471B	PPB	29 Mar 2018 15:52:41	U	SMPL
ICB67 - 1	ICB67	-129	0.0018	-	7471B	PPB	29 Mar 2018 15:54:56	U	SMPL
CCV96 - 1	CCV96	21028	4.9278	-	7471B	PPB	29 Mar 2018 15:57:13	U	SMPL
CCB96 - 1	CCB96	-171	-0.0079	-	7471B	PPB	29 Mar 2018 15:59:29	U	SMPL
CRA - 1	CRA	723	0.2002	-	7471B	PPB	29 Mar 2018 16:01:45	U	SMPL
HighStd - 1	HighStd	43280	10.1087	-	7471B	PPB	29 Mar 2018 16:04:01	U	SMPL
ChkStd - 1	ChkStd	28952	6.7727	-	7471B	PPB	29 Mar 2018 16:06:17	U	SMPL
PB107765BL - 1	PBS	-208	-0.0165	-	7471B	PPB	29 Mar 2018 16:08:36	U	SMPL
PB107765BS - 1	LCSS	14685	3.4510	-	7471B	PPB	29 Mar 2018 16:10:54	U	SMPL
J1749-07 - 1	BU-02-032818	1364	0.3495	-	7471B	PPB	29 Mar 2018 16:13:11	U	SMPL
J1755-05 - 1	OK-02-032818	2267	0.5597	-	7471B	PPB	29 Mar 2018 16:15:28	U	SMPL
J1761-03 - 1	TR-04-032818	1652	0.4165	-	7471B	PPB	29 Mar 2018 16:17:45	U	SMPL
J2080-07 - 1	TEC-RCA-COMP	4934	1.1807	-	7471B	PPB	29 Mar 2018 16:22:19	U	SMPL
CCV97 - 1	CCV97	20635	4.8363	-	7471B	PPB	29 Mar 2018 16:24:35	U	SMPL
CCB97 - 1	CCB97	-114	0.0053	-	7471B	PPB	29 Mar 2018 16:26:51	U	SMPL
J2091-01 - 1	FK-BPM-01-G1-G2-H1-H2	1330	0.3415	-	7471B	PPB	29 Mar 2018 16:29:08	U	SMPL
J2099-01 - 1	SB-102(1-2)	19980	4.6838	-	7471B	PPB	29 Mar 2018 16:31:24	U	SMPL
J2099-02 - 1	SB-102(11-12)	534	0.1562	-	7471B	PPB	29 Mar 2018 16:33:40	U	SMPL
J2099-03 - 1	J2099-02MS	15831	3.7178	-	7471B	PPB	29 Mar 2018 16:38:15	U	SMPL
J2099-04 - 1	J2099-02MSD	16352	3.8391	-	7471B	PPB	29 Mar 2018 16:40:32	U	SMPL
J2099-05 - 1	SB-105(2-3)	4024	0.9688	-	7471B	PPB	29 Mar 2018 16:42:49	U	SMPL
J2099-06 - 1	SB-105(13-14)	176	0.0729	-	7471B	PPB	29 Mar 2018 16:45:06	U	SMPL
J2099-07 - 1	SB-103(5-6)	18329	4.2994	-	7471B	PPB	29 Mar 2018 16:49:31	U	SMPL
J2099-08 - 1	DUP-032718	124223	28.9544	-	7471B	PPB	29 Mar 2018 16:51:47	U	SMPL
CCV98 - 1	CCV98	20527	4.8111	-	7471B	PPB	29 Mar 2018 16:54:04	U	SMPL
CCB98 - 1	CCB98	-154	-0.0040	-	7471B	PPB	29 Mar 2018 16:56:24	U	SMPL
PB107766BL - 1	PBS	19	0.0363	-	7471B	PPB	29 Mar 2018 16:58:41	U	SMPL
PB107766BS - 1	LCSS	15122	3.5527	-	7471B	PPB	29 Mar 2018 17:00:58	U	SMPL
J2100-01 - 1	TP-8-DB	87706	20.4523	-	7471B	PPB	29 Mar 2018 17:03:14	U	SMPL
J2100-04 - 1	TP-5-DB	112344	26.1887	-	7471B	PPB	29 Mar 2018 17:05:30	U	SMPL
J2103-01 - 1	LS-SOIL-1	1352	0.3467	-	7471B	PPB	29 Mar 2018 17:10:26	U	SMPL
J2103-01DUP - 1	LS-SOIL-1DUP	1194	0.3099	-	7471B	PPB	29 Mar 2018 17:12:43	U	SMPL
J2103-01MS - 1	LS-SOIL-1MS	17604	4.1306	-	7471B	PPB	29 Mar 2018 17:15:00	U	SMPL
J2103-01MSD - 1	LS-SOIL-1MSD	17579	4.1248	-	7471B	PPB	29 Mar 2018 17:17:18	U	SMPL
J2103-05 - 1	LS-SOIL-2	1760	0.4417	-	7471B	PPB	29 Mar 2018 17:19:36	U	SMPL
J2104-01 - 1	20180220-COMPOSITE-A	1159	0.3017	-	7471B	PPB	29 Mar 2018 17:21:52	U	SMPL
CCV99 - 1	CCV99	21377	5.0090	-	7471B	PPB	29 Mar 2018 17:24:08	U	SMPL
CCB99 - 1	CCB99	-93	0.0102	-	7471B	PPB	29 Mar 2018 17:26:25	U	SMPL
J2104-02 - 1	20180221-COMPOSITE-B	7848	1.8591	-	7471B	PPB	29 Mar 2018 17:28:41	U	SMPL
J2112-02 - 1	S-1-COMP	4565	1.0947	-	7471B	PPB	29 Mar 2018 17:30:57	U	SMPL
J2112-04 - 1	S-2-COMP	7755	1.8375	-	7471B	PPB	29 Mar 2018 17:33:14	U	SMPL
J2112-06 - 1	S-3-COMP	3676	0.8878	-	7471B	PPB	29 Mar 2018 17:35:30	U	SMPL
J2117-01 - 1	TPIAS	39403	9.2060	-	7471B	PPB	29 Mar 2018 17:37:47	U	SMPL
J2117-03 - 1	TPIAD	850	0.2298	-	7471B	PPB	29 Mar 2018 17:40:03	U	SMPL
J2128-01 - 1	EPE-124-2(0-5)	150624	35.1013	-	7471B	PPB	29 Mar 2018 17:42:22	U	SMPL
J2128-02 - 1	EPE-124-2(5-10)	51734	12.0770	-	7471B	PPB	29 Mar 2018 17:44:40	U	SMPL
J2099-02DUP - 1	SB-102(11-12)DUP	581	0.1672	-	7471B	PPB	29 Mar 2018 17:52:54	U	SMPL
J2072-07DLX5 - 1	COAB5	246	0.0892	-	7471B	PPB	29 Mar 2018 17:55:10	U	SMPL
CCV01 - 1	CCV01	21215	4.9713	-	7471B	PPB	29 Mar 2018 17:57:26	U	SMPL
CCB01 - 1	CCB01	-102	0.0081	-	7471B	PPB	29 Mar 2018 17:59:42	U	SMPL
J2099-02LX5 - 1		122	0.0603	-	7471B	PPB	29 Mar 2018 18:01:59	U	SMPL
J2099-02A - 1		17898	4.1990	-	7471B	PPB	29 Mar 2018 18:04:16	U	SMPL
J2103-01LX5 - 1		135	0.0633	-	7471B	PPB	29 Mar 2018 18:06:32	U	SMPL
J2103-01A - 1		17754	4.1655	-	7471B	PPB	29 Mar 2018 18:08:48	U	SMPL
J2099-08DLX5 - 1		23611	5.5292	-	7471B	PPB	29 Mar 2018 18:11:05	U	SMPL
J2100-01DLX5 - 1		16777	3.9380	-	7471B	PPB	29 Mar 2018 18:13:21	U	SMPL
J2100-04DLX5 - 1		21805	5.1087	-	7471B	PPB	29 Mar 2018 18:15:40	U	SMPL
J2128-01DLX5 - 1		29383	6.8731	-	7471B	PPB	29 Mar 2018 18:17:58	U	SMPL
J2128-02DLX5 - 1		25335	5.9306	-	7471B	PPB	29 Mar 2018 18:20:17	U	SMPL
CCV02 - 1	CCV02	21126	4.9506	-	7471B	PPB	29 Mar 2018 18:22:35	U	SMPL
CCB02 - 1	CCB02	-107	0.0070	-	7471B	PPB	29 Mar 2018 18:24:52	U	SMPL