



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94249

Temp.: 21.2 °C

BalanceID: WC SC-4

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

True Value of CCV3 = 2.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/29/2018	18:10
2	CAL2	1	Water			7.00	03/29/2018	18:11
3	CAL3	1	Water			10.02	03/29/2018	18:12
4	ICV	1	Water			7.00	03/29/2018	18:14
5	CCV1	1	Water			2.01	03/30/2018	18:15
6	J1752-04	1	Solid	20.02	20	7.10	03/29/2018	18:16
7	J1752-04DUP	1	Solid	20.05	20	7.11	03/29/2018	18:19
8	J1752-06	1	Solid	20.02	20	5.63	03/29/2018	18:20
9	J1752-08	1	Solid	20.05	20	7.12	03/29/2018	18:21
10	J1760-10	1	Solid	20.02	20	8.09	03/29/2018	18:22
11	J2137-01	1	Solid	20.03	20	8.62	03/29/2018	18:25
12	J2142-02	1	Solid	20.01	20	7.63	03/29/2018	18:26
13	J2142-04	1	Solid	20.03	20	9.96	03/29/2018	18:30
14	J2142-06	1	Solid	20.02	20	7.93	03/29/2018	18:31
15	J2142-08	1	Solid	20.03	20	8.16	03/29/2018	18:33
16	CCV2	1	Water			12.01	03/29/2018	18:35
17	J2142-10	1	Solid	20.02	20	11.02	03/29/2018	18:36
18	J2146-02	1	Solid	20.05	20	8.03	03/29/2018	18:37
19	J2146-04	1	Solid	20.02	20	8.87	03/29/2018	18:39
20	CCV3	1	Water			2.01	03/29/2018	18:40

2894249

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j2106

WorkList ID : 110203

Date : 3/29/2018 6:03:09 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/29/2018	Solid	J1752-04	Corrosivity	Cool 4 deg C	PSEG05		NB-01-032918-A	03/29/2018	9045C
03/29/2018	Solid	J1752-06	Corrosivity	Cool 4 deg C	PSEG05		NB-01-032918-B	03/29/2018	9045C
03/29/2018	Solid	J1752-08	Corrosivity	Cool 4 deg C	PSEG05		NB-01-032918-C	03/29/2018	9045C
03/29/2018	Solid	J1760-10	Corrosivity	Cool 4 deg C	PSEG05		SU-03-032918	03/29/2018	9045C
03/28/2018	Solid	J2137-01	Corrosivity	Cool 4 deg C	USAR03	K11	CB-0318-S-TCLP-64-1	03/28/2018	9045C
03/29/2018	Solid	J2142-02	Corrosivity	Cool 4 deg C	PSEG01	J63	GB-RT-2334	03/29/2018	9045C
03/29/2018	Solid	J2142-04	Corrosivity	Cool 4 deg C	PSEG01	J63	GB-RT-4632	03/29/2018	9045C
03/29/2018	Solid	J2142-06	Corrosivity	Cool 4 deg C	PSEG01	J63	GB-RT-3025	03/29/2018	9045C
03/29/2018	Solid	J2142-08	Corrosivity	Cool 4 deg C	PSEG01	J63	GB-RT-4252	03/29/2018	9045C
03/29/2018	Solid	J2142-10	Corrosivity	Cool 4 deg C	PSEG01	J63	GB-RT-2342	03/29/2018	9045C
03/29/2018	Solid	J2146-02	Corrosivity	Cool 4 deg C	PSEG01	J63	TP-CS	03/29/2018	9045C
03/29/2018	Solid	J2146-04	Corrosivity	Cool 4 deg C	PSEG01	J63	TP-CD	03/29/2018	9045C

Date/Time 03/29/18 6:03 PM
 Received by: JD
 Relinquished by: JD

Date/Time 03/29/18 7:30 PM
 Received by: JD
 Relinquished by: JD