



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Sweta

Run Number: LB94250

Temp.: 16.4 °C

SLOPE.: 97.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	03/29/2018	18:10
2	CAL2	1	Water			7.00	03/29/2018	18:11
3	CAL3	1	Water			10.02	03/29/2018	18:12
4	ICV	1	Water			7.01	03/29/2018	18:13
5	CCV1	1	Water			2.01	03/29/2018	18:15
6	J2103-10	1	Water			7.71	03/29/2018	18:16
7	J2103-10DUP	1	Water			7.72	03/29/2018	18:18
8	CCV2	1	Water			12.02	03/29/2018	18:20

1D94250

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity J2103 WorkList ID : 110204 Date : 3/29/2018 6:03:54 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/28/2018	Water	J2103-10-1A	Corrosivity	Cool 4 deg C	PSEG01	J53	LS	03/28/2018	9040C

Date/Time 03/29/18 6:00PM
Received by: SD
Relinquished by: SD

Date/Time 03/29/18 7:00PM
Received by: SD
Relinquished by: SD