


245.1

LB94252
INSTRUMENT ID: CY2

Linear

 μ Abs.:

32168



A= 0.0000e+000

B= 3.0691e-004

C= -1.2062e-002

Rho= 0.9996524

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	43	0.000	43				
0.05	0.050	0.044	-0.006	183	0.0 %	183				
0.2	0.200	0.188	-0.012	653	0.0 %	653				
2.5	2.500	2.392	-0.108	7832	0.0 %	7832				
5.0	5.000	5.152	0.152	16827	0.0 %	16827				
7.5	7.500	7.612	0.112	24841	0.0 %	24841				
10.0	10.000	9.861	-0.139	32168	0.0 %	32168				

LB94252 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 29 Mar 2018 19:11:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	43	-	0.0000	245.1	PPB	29 Mar 2018 19:11:38	S	Std
0.05 - 1	50.02	183	-	0.0500	245.1	PPB	29 Mar 2018 19:13:47	S	Std
0.2 - 1	52.2	653	-	0.2000	245.1	PPB	29 Mar 2018 19:15:55	S	Std
2.5 - 1	52.5	7832	-	2.5000	245.1	PPB	29 Mar 2018 19:18:03	S	Std
5.0 - 1	55	16827	-	5.0000	245.1	PPB	29 Mar 2018 19:20:10	S	Std
7.5 - 1	57.5	24841	-	7.5000	245.1	PPB	29 Mar 2018 19:22:17	S	Std
10.0 - 1	510	32168	-	10.0000	245.1	PPB	29 Mar 2018 19:35:31	S	Std
ICV68 - 1	ICV68	12905	3.9486	-245.1		PPB	29 Mar 2018 19:38:12	U	SMPL
ICB68 - 1	ICB68	-139	-0.0547	-245.1		PPB	29 Mar 2018 19:40:19	U	SMPL
CCV03 - 1	CCV03	16904	5.1759	-245.1		PPB	29 Mar 2018 19:42:26	U	SMPL
CCB03 - 1	CCB03	-151	-0.0584	-245.1		PPB	29 Mar 2018 19:44:33	U	SMPL
CRA - 1	CRA	213	0.0533	-245.1		PPB	29 Mar 2018 19:49:02	U	SMPL
HighStd - 1	HighStd	32518	9.9680	-245.1		PPB	29 Mar 2018 19:51:08	U	SMPL
ChkStd - 1	ChkStd	23707	7.2638	-245.1		PPB	29 Mar 2018 19:53:15	U	SMPL
PB107802BL - 1	PBW	-24	-0.0194	-245.1		PPB	29 Mar 2018 19:57:48	U	SMPL
PB107802BS - 1	LCSW	13416	4.1054	-245.1		PPB	29 Mar 2018 19:59:56	U	SMPL
J2113-01 - 1	032618-TIDO-E-D-B	-75	-0.0351	-245.1		PPB	29 Mar 2018 20:02:03	U	SMPL
J2113-02 - 1	032618-TIDO-E-D-S	44	0.0014	-245.1		PPB	29 Mar 2018 20:04:11	U	SMPL
J2113-03 - 1	032618-TIDO-E-U-B	29	-0.0032	-245.1		PPB	29 Mar 2018 20:06:19	U	SMPL
J2113-04 - 1	032618-TIDO-E-U-S	36	-0.0010	-245.1		PPB	29 Mar 2018 20:08:27	U	SMPL
J2113-05 - 1	032618-DBRI-E-D-B	51	0.0036	-245.1		PPB	29 Mar 2018 20:10:34	U	SMPL
CCV04 - 1	CCV04	16927	5.1830	-245.1		PPB	29 Mar 2018 20:12:41	U	SMPL
CCB04 - 1	CCB04	-130	-0.0520	-245.1		PPB	29 Mar 2018 20:14:48	U	SMPL
J2113-06 - 1	032618-DBRI-E-D-S	19	-0.0062	-245.1		PPB	29 Mar 2018 20:16:56	U	SMPL
J2113-07 - 1	032618-DBRI-E-U-B	48	0.0027	-245.1		PPB	29 Mar 2018 20:19:03	U	SMPL
J2113-08 - 1	032618-DBRI-E-U-S	40	0.0002	-245.1		PPB	29 Mar 2018 20:21:10	U	SMPL
J2113-08DUP - 1	032618-DBRI-E-U-SDUP	42	0.0008	-245.1		PPB	29 Mar 2018 20:23:18	U	SMPL
J2113-08MS - 1	032618-DBRI-E-U-SMS	15734	4.8168	-245.1		PPB	29 Mar 2018 20:25:25	U	SMPL
J2113-08MSD - 1	032618-DBRI-E-U-SMSD	14830	4.5394	-245.1		PPB	29 Mar 2018 20:27:32	U	SMPL
PB107803BL - 1	PBW	-25	-0.0197	-245.1		PPB	29 Mar 2018 20:32:20	U	SMPL
PB107803BS - 1	LCSW	13436	4.1116	-245.1		PPB	29 Mar 2018 20:34:28	U	SMPL
J2114-01 - 1	032718-DBRI-F-D-B	-48	-0.0268	-245.1		PPB	29 Mar 2018 20:36:37	U	SMPL
J2114-02 - 1	032718-DBRI-F-D-M	28	-0.0035	-245.1		PPB	29 Mar 2018 20:38:45	U	SMPL
CCV05 - 1	CCV05	16578	5.0759	-245.1		PPB	29 Mar 2018 20:40:51	U	SMPL
CCB05 - 1	CCB05	-126	-0.0507	-245.1		PPB	29 Mar 2018 20:42:59	U	SMPL
J2114-03 - 1	032718-DBRI-F-D-S	30	-0.0029	-245.1		PPB	29 Mar 2018 20:45:07	U	SMPL
J2114-04 - 1	032718-DBRI-F-U-B	57	0.0054	-245.1		PPB	29 Mar 2018 20:47:14	U	SMPL
J2114-05 - 1	032718-DBRI-F-U-M	56	0.0051	-245.1		PPB	29 Mar 2018 20:49:21	U	SMPL
J2114-06 - 1	032718-DBRI-F-U-S	47	0.0024	-245.1		PPB	29 Mar 2018 20:51:29	U	SMPL
J2114-07 - 1	032718-TIDO-E-D-B	67	0.0085	-245.1		PPB	29 Mar 2018 20:53:36	U	SMPL
J2114-08 - 1	032718-TIDO-E-D-S	52	0.0039	-245.1		PPB	29 Mar 2018 20:55:44	U	SMPL
J2114-09 - 1	032718-TIDO-E-U-B	61	0.0067	-245.1		PPB	29 Mar 2018 20:57:52	U	SMPL
J2114-10 - 1	032718-TIDO-E-U-S	56	0.0051	-245.1		PPB	29 Mar 2018 21:00:00	U	SMPL
J2114-11 - 1	032718-SIDI-E-D-B	71	0.0097	-245.1		PPB	29 Mar 2018 21:02:09	U	SMPL
J2114-12 - 1	032718-SIDI-E-D-M	69	0.0091	-245.1		PPB	29 Mar 2018 21:04:16	U	SMPL
CCV06 - 1	CCV06	16482	5.0464	-245.1		PPB	29 Mar 2018 21:06:23	U	SMPL
CCB06 - 1	CCB06	-126	-0.0507	-245.1		PPB	29 Mar 2018 21:08:30	U	SMPL
J2114-13 - 1	032718-SIDI-E-D-S	17	-0.0068	-245.1		PPB	29 Mar 2018 21:10:39	U	SMPL
J2114-14 - 1	032718-SIDI-E-U-B	60	0.0064	-245.1		PPB	29 Mar 2018 21:12:46	U	SMPL
J2114-15 - 1	032718-SIDI-E-U-M	59	0.0060	-245.1		PPB	29 Mar 2018 21:14:53	U	SMPL
J2114-16 - 1	032718-SIDI-E-U-S	69	0.0091	-245.1		PPB	29 Mar 2018 21:17:00	U	SMPL
J2114-16DUP - 1	032718-SIDI-E-U-SDUP	53	0.0042	-245.1		PPB	29 Mar 2018 21:19:08	U	SMPL
J2114-16MS - 1	032718-SIDI-E-U-SMS	14896	4.5597	-245.1		PPB	29 Mar 2018 21:21:15	U	SMPL
J2114-16MSD - 1	032718-SIDI-E-U-SMSD	14371	4.3985	-245.1		PPB	29 Mar 2018 21:23:24	U	SMPL
CCV07 - 1	CCV07	16376	5.0139	-245.1		PPB	29 Mar 2018 21:25:32	U	SMPL
CCB07 - 1	CCB07	-136	-0.0538	-245.1		PPB	29 Mar 2018 21:27:40	U	SMPL
J2113-08LX5 - 1		-3	-0.0130	-245.1		PPB	29 Mar 2018 21:29:49	U	SMPL
J2113-08A - 1		15384	4.7094	-245.1		PPB	29 Mar 2018 21:31:56	U	SMPL
J2114-16LX5 - 1		-92	-0.0403	-245.1		PPB	29 Mar 2018 21:34:03	U	SMPL
J2114-16A - 1		15638	4.7874	-245.1		PPB	29 Mar 2018 21:36:10	U	SMPL
CCV08 - 1	CCV08	16575	5.0750	-245.1		PPB	29 Mar 2018 21:38:17	U	SMPL
CCB08 - 1	CCB08	-144	-0.0563	-245.1		PPB	29 Mar 2018 21:40:25	U	SMPL