



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Sweta

Run Number: LB94259

Temp.: 16.4 °C

SLOPE.: 97.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	03/29/2018	18:10
2	CAL2	1	Water			7.00	03/29/2018	18:11
3	CAL3	1	Water			10.02	03/29/2018	18:12
4	ICV	1	Water			7.01	03/29/2018	18:13
5	CCV1	1	Water			2.01	03/29/2018	18:15
6	J2103-09	1	Water			7.71	03/29/2018	18:16
7	J2103-09DUP	1	Water			7.72	03/29/2018	18:19
8	J2106-03	1	Water			5.51	03/29/2018	18:20
9	J2106-04	1	Water			7.58	03/29/2018	18:21
10	J2106-05	1	Water			7.87	03/29/2018	18:22
11	J2106-06	1	Water			7.78	03/29/2018	18:25
12	J2131-10	1	Water			6.43	03/29/2018	18:26
13	CCV2	1	Water			12.02	03/29/2018	18:29

LB94259

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph J2103

WorkList ID : 110205

Date : 3/29/2018 6:04:33 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/28/2018	Water	J2103-09 -A	pH	Cool 4 deg C	PSEG01	J53	LS	03/28/2018	9040C
03/28/2018	Water	J2106-03 -A	pH	Cool 4 deg C	PSEG01	J53	19619	03/28/2018	9040C
03/28/2018	Water	J2106-04 -A	pH	Cool 4 deg C	PSEG01	J53	19622	03/28/2018	9040C
03/28/2018	Water	J2106-05 -A	pH	Cool 4 deg C	PSEG01	J53	19606	03/28/2018	9040C
03/28/2018	Water	J2106-06 -A	pH	Cool 4 deg C	PSEG01	J53	19607	03/28/2018	9040C
03/28/2018	Water	J2131-10 -A	pH	Cool 4 deg C	PSEG01	J63	4507-4543	03/28/2018	9040C

Date/Time 03/29/18 6:00 PM
 Received by: SN
 Relinquished by: Y

Date/Time 03/29/18 7:30 PM
 Received by: CP
 Relinquished by: SN