


7470A

LB94262
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34178



A= 0.0000e+000

B= 2.8900e-004

C= 7.2631e-003

Rho= 0.9996766

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	6	0.000	6				
0.2	0.200	0.181	-0.019	602	0.0 %	602				
2.5	2.500	2.395	-0.105	8262	0.0 %	8262				
5.0	5.000	5.144	0.144	17773	0.0 %	17773				
7.5	7.500	7.586	0.086	26225	0.0 %	26225				
10.0	10.000	9.885	-0.115	34178	0.0 %	34178				

LB94262

Method: 7470A

Operator: Admin

Date of Analysis: 30 Mar 2018 10:05:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	6	-	0.0000	7470A	PPB	30 Mar 2018 10:20:10	S	Std
0.2 - 1	50.2	602	-	0.2000	7470A	PPB	30 Mar 2018 10:22:18	S	Std
2.5 - 1	52.5	8262	-	2.5000	7470A	PPB	30 Mar 2018 10:24:26	S	Std
5.0 - 1	55	17773	-	5.0000	7470A	PPB	30 Mar 2018 10:26:34	S	Std
7.5 - 1	57.5	26225	-	7.5000	7470A	PPB	30 Mar 2018 10:28:43	S	Std
10.0 - 1	510	34178	-	10.0000	7470A	PPB	30 Mar 2018 10:33:17	S	Std
ICV69 - 1	ICV69	12963	3.7536	-	7470A	PPB	30 Mar 2018 10:35:50	U	SMPL
ICB69 - 1	ICB69	-254	-0.0661	-	7470A	PPB	30 Mar 2018 10:37:58	U	SMPL
CCV09 - 1	CCV09	17427	5.0437	-	7470A	PPB	30 Mar 2018 10:40:06	U	SMPL
CCB09 - 1	CCB09	-219	-0.0560	-	7470A	PPB	30 Mar 2018 10:42:12	U	SMPL
CRA - 1	CRA	616	0.1853	-	7470A	PPB	30 Mar 2018 10:44:20	U	SMPL
HighStd - 1	HighStd	34517	9.9827	-	7470A	PPB	30 Mar 2018 10:46:26	U	SMPL
ChkStd - 1	ChkStd	24354	7.0456	-	7470A	PPB	30 Mar 2018 10:48:33	U	SMPL
PB107790BL - 1	PBW	-281	-0.0739	-	7470A	PPB	30 Mar 2018 10:50:43	U	SMPL
PB107790BS - 1	LCSW	13552	3.9238	-	7470A	PPB	30 Mar 2018 10:52:52	U	SMPL
J2072-01 - 1	C0AA1	-145	-0.0346	-	7470A	PPB	30 Mar 2018 10:54:59	U	SMPL
J2072-02 - 1	C0AB1	12	0.0107	-	7470A	PPB	30 Mar 2018 10:57:08	U	SMPL
J2072-03 - 1	C0AB2	107	0.0382	-	7470A	PPB	30 Mar 2018 10:59:16	U	SMPL
J2072-04 - 1	C0AB3	-24	0.0003	-	7470A	PPB	30 Mar 2018 11:01:24	U	SMPL
J2072-05 - 1	C0AB4	-13	0.0035	-	7470A	PPB	30 Mar 2018 11:03:30	U	SMPL
CCV10 - 1	CCV10	17553	5.0801	-	7470A	PPB	30 Mar 2018 11:05:37	U	SMPL
CCB10 - 1	CCB10	-200	-0.0505	-	7470A	PPB	30 Mar 2018 11:07:43	U	SMPL
J2101-02 - 1	FB-032718	12	0.0107	-	7470A	PPB	30 Mar 2018 11:09:51	U	SMPL
J2103-09 - 1	LS	50	0.0217	-	7470A	PPB	30 Mar 2018 11:11:58	U	SMPL
J2103-09DUP - 1	LS DUP	52	0.0223	-	7470A	PPB	30 Mar 2018 11:14:05	U	SMPL
J2103-09MS - 1	LSMS	13361	3.8686	-	7470A	PPB	30 Mar 2018 11:16:12	U	SMPL
J2103-09MSD - 1	LSMSD	13540	3.9203	-	7470A	PPB	30 Mar 2018 11:18:20	U	SMPL
J2104-03 - 1	20180219-FIELD-BLANK	-153	-0.0370	-	7470A	PPB	30 Mar 2018 11:20:28	U	SMPL
PB107819BL - 1	PBW	2	0.0078	-	7470A	PPB	30 Mar 2018 11:22:36	U	SMPL
PB107819BS - 1	LCSW	13078	3.7868	-	7470A	PPB	30 Mar 2018 11:24:45	U	SMPL
PB107819TB - 1	PB107758TB	-130	-0.0303	-	7470A	PPB	30 Mar 2018 11:26:53	U	SMPL
J1749-08 - 1	BU-02-032818	-16	0.0026	-	7470A	PPB	30 Mar 2018 11:29:01	U	SMPL
CCV11 - 1	CCV11	17593	5.0917	-	7470A	PPB	30 Mar 2018 11:31:08	U	SMPL
CCB11 - 1	CCB11	-202	-0.0511	-	7470A	PPB	30 Mar 2018 11:33:15	U	SMPL
J1755-06 - 1	OK-02-032818	-10	0.0044	-	7470A	PPB	30 Mar 2018 11:35:23	U	SMPL
J1761-04 - 1	TR-04-032818	6	0.0090	-	7470A	PPB	30 Mar 2018 11:37:30	U	SMPL
J2103-04 - 1	LS-SOIL-1	4	0.0084	-	7470A	PPB	30 Mar 2018 11:39:37	U	SMPL
J2103-08 - 1	LS-SOIL-2	7	0.0093	-	7470A	PPB	30 Mar 2018 11:41:44	U	SMPL
J2117-02 - 1	TPIAS	27	0.0151	-	7470A	PPB	30 Mar 2018 11:43:51	U	SMPL
J2117-04 - 1	TPIAD	44	0.0200	-	7470A	PPB	30 Mar 2018 11:46:00	U	SMPL
J2117-04DUP - 1	TPIADDUP	44	0.0200	-	7470A	PPB	30 Mar 2018 11:48:08	U	SMPL
J2117-04MS - 1	TPIADMS	13484	3.9042	-	7470A	PPB	30 Mar 2018 11:50:16	U	SMPL
J2117-04MSD - 1	TPIADMSD	14553	4.2131	-	7470A	PPB	30 Mar 2018 11:52:25	U	SMPL
PB107820BL - 1	PBW	-127	-0.0294	-	7470A	PPB	30 Mar 2018 11:54:32	U	SMPL
CCV12 - 1	CCV12	17332	5.0162	-	7470A	PPB	30 Mar 2018 11:56:40	U	SMPL
CCB12 - 1	CCB12	-183	-0.0456	-	7470A	PPB	30 Mar 2018 11:58:47	U	SMPL
PB107820BS - 1	LCSW	13444	3.8926	-	7470A	PPB	30 Mar 2018 12:00:55	U	SMPL
PB107820TB - 1	PB107764TB	-141	-0.0335	-	7470A	PPB	30 Mar 2018 12:03:03	U	SMPL
J2103-10 - 1	LS	-17	0.0024	-	7470A	PPB	30 Mar 2018 12:05:11	U	SMPL
J2103-10DUP - 1	LS DUP	-3	0.0064	-	7470A	PPB	30 Mar 2018 12:07:18	U	SMPL
J2103-10MS - 1	LSMS	14510	4.2007	-	7470A	PPB	30 Mar 2018 12:09:26	U	SMPL
J2103-10MSD - 1	LSMSD	13828	4.0036	-	7470A	PPB	30 Mar 2018 12:11:34	U	SMPL
CCV13 - 1	CCV13	17578	5.0873	-	7470A	PPB	30 Mar 2018 12:13:43	U	SMPL
CCB13 - 1	CCB13	-181	-0.0450	-	7470A	PPB	30 Mar 2018 12:15:51	U	SMPL
J2103-09LX5 - 1		7	0.0093	-	7470A	PPB	30 Mar 2018 12:18:01	U	SMPL
J2103-09A - 1		14039	4.0646	-	7470A	PPB	30 Mar 2018 12:20:08	U	SMPL
J2117-04LX5 - 1		-112	-0.0251	-	7470A	PPB	30 Mar 2018 12:22:15	U	SMPL
J2117-04A - 1		13724	3.9735	-	7470A	PPB	30 Mar 2018 12:24:23	U	SMPL
J2103-10LX5 - 1		0	0.0073	-	7470A	PPB	30 Mar 2018 12:26:57	U	SMPL
J2103-10A - 1		13333	3.8605	-	7470A	PPB	30 Mar 2018 12:29:04	U	SMPL
CCV14 - 1	CCV14	17582	5.0885	-	7470A	PPB	30 Mar 2018 12:41:11	U	SMPL
CCB14 - 1	CCB14	-205	-0.0520	-	7470A	PPB	30 Mar 2018 12:43:19	U	SMPL