

7471B


LB 94276

INSTRUMENT ID: CV2

Linear

 μ Abs.:

41815



A= 0.0000e+000

B= 2.3616e-004

C= 2.1248e-002

Rho= 0.9997239

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.026	0.026	22	0.000	22				
0.2	0.200	0.202	0.002	766	0.0 %	766				
2.5	2.500	2.387	-0.113	10018	0.0 %	10018				
5.0	5.000	5.054	0.054	21313	0.0 %	21313				
7.5	7.500	7.634	0.134	32235	0.0 %	32235				
10	10.000	9.896	-0.104	41815	0.0 %	41815				

LB94276 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Mar 2018 14:38:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	22	-	0.0000	7471B	PPB	30 Mar 2018 14:40:33	S	Std
0.2 - 1	50.2	766	-	0.2000	7471B	PPB	30 Mar 2018 14:42:51	S	Std
2.5 - 1	52.5	10018	-	2.5000	7471B	PPB	30 Mar 2018 14:45:09	S	Std
5.0 - 1	55	21313	-	5.0000	7471B	PPB	30 Mar 2018 14:47:26	S	Std
7.5 - 1	57.5	32235	-	7.5000	7471B	PPB	30 Mar 2018 14:49:44	S	Std
10 - 1	510	41815	-	10.0000	7471B	PPB	30 Mar 2018 15:02:19	S	Std
ICV70 - 1	ICV70	15777	3.7471	-	7471B	PPB	30 Mar 2018 15:05:02	U	SMPL
ICB70 - 1	ICB70	-170	-0.0189	-	7471B	PPB	30 Mar 2018 15:07:18	U	SMPL
CCV15 - 1	CCV15	22049	5.2283	-	7471B	PPB	30 Mar 2018 15:09:35	U	SMPL
CCB15 - 1	CCB15	-179	-0.0210	-	7471B	PPB	30 Mar 2018 15:11:51	U	SMPL
CRA - 1	CRA	810	0.2125	-	7471B	PPB	30 Mar 2018 15:14:08	U	SMPL
HighStd - 1	HighStd	41848	9.9039	-	7471B	PPB	30 Mar 2018 15:16:23	U	SMPL
ChkStd - 1	ChkStd	28799	6.8223	-	7471B	PPB	30 Mar 2018 15:18:40	U	SMPL
PB107826BL - 1	PBS	-279	-0.0446	-	7471B	PPB	30 Mar 2018 15:20:58	U	SMPL
PB107826BS - 1	LCSS	14589	3.4665	-	7471B	PPB	30 Mar 2018 15:23:16	U	SMPL
J1752-03 - 1	NB-01-032918-A	3180	0.7722	-	7471B	PPB	30 Mar 2018 15:25:33	U	SMPL
J1752-05 - 1	NB-01-032918-B	1518	0.3797	-	7471B	PPB	30 Mar 2018 15:27:50	U	SMPL
J1752-07 - 1	NB-01-032918-C	2727	0.6652	-	7471B	PPB	30 Mar 2018 15:30:07	U	SMPL
J1760-09 - 1	SU-03-032918	953	0.2463	-	7471B	PPB	30 Mar 2018 15:32:26	U	SMPL
J2139-10 - 1	PAD-1	65337	15.4510	-	7471B	PPB	30 Mar 2018 15:34:42	U	SMPL
CCV16 - 1	CCV16	21260	5.0419	-	7471B	PPB	30 Mar 2018 15:36:58	U	SMPL
CCB16 - 1	CCB16	-182	-0.0217	-	7471B	PPB	30 Mar 2018 15:39:15	U	SMPL
J2139-11 - 1	PAD-2	33410	7.9112	-	7471B	PPB	30 Mar 2018 15:41:32	U	SMPL
J2142-01 - 1	GB-RT-2334	3888	0.9394	-	7471B	PPB	30 Mar 2018 15:43:48	U	SMPL
J2142-03 - 1	GB-RT-4632	1584	0.3953	-	7471B	PPB	30 Mar 2018 15:46:06	U	SMPL
J2142-05 - 1	GB-RT-3025	4651	1.1196	-	7471B	PPB	30 Mar 2018 15:48:22	U	SMPL
J2142-07 - 1	GB-RT-4252	5151	1.2377	-	7471B	PPB	30 Mar 2018 15:50:38	U	SMPL
J2142-09 - 1	GB-RT-2342	501	0.1396	-	7471B	PPB	30 Mar 2018 15:52:55	U	SMPL
J2144-01 - 1	RW-3	425	0.1216	-	7471B	PPB	30 Mar 2018 15:55:12	U	SMPL
J2144-01DUP - 1	RW-3DUP	472	0.1327	-	7471B	PPB	30 Mar 2018 15:57:30	U	SMPL
J2144-01MS - 1	RW-3MS	17261	4.0975	-	7471B	PPB	30 Mar 2018 15:59:47	U	SMPL
J2144-01MSD - 1	RW-3MSD	17394	4.1290	-	7471B	PPB	30 Mar 2018 16:02:03	U	SMPL
CCV17 - 1	CCV17	21859	5.1834	-	7471B	PPB	30 Mar 2018 16:04:20	U	SMPL
CCB17 - 1	CCB17	-154	-0.0151	-	7471B	PPB	30 Mar 2018 16:06:36	U	SMPL
J2144-02 - 1	RW-4	1877	0.4645	-	7471B	PPB	30 Mar 2018 16:08:53	U	SMPL
J2146-01 - 1	TP-CS	284	0.0883	-	7471B	PPB	30 Mar 2018 16:11:09	U	SMPL
J2146-03 - 1	TP-CD	436	0.1242	-	7471B	PPB	30 Mar 2018 16:13:25	U	SMPL
J2144-01LX5 - 1		58	0.0349	-	7471B	PPB	30 Mar 2018 16:15:42	U	SMPL
J2144-01A - 1		17421	4.1353	-	7471B	PPB	30 Mar 2018 16:17:59	U	SMPL
J2139-010DLX2 - 1		32580	7.7152	-	7471B	PPB	30 Mar 2018 16:20:16	U	SMPL
CCV18 - 1	CCV18	21825	5.1754	-	7471B	PPB	30 Mar 2018 16:22:33	U	SMPL
CCB18 - 1	CCB18	-139	-0.0116	-	7471B	PPB	30 Mar 2018 16:24:52	U	SMPL