



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB94282

Temp.: 20.2 °C

SLOPE.: 99.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/30/2018	15:15
2	CAL2	1	Water			7.00	03/30/2018	15:16
3	CAL3	1	Water			10.03	03/30/2018	15:17
4	ICV	1	Water			7.00	03/30/2018	15:18
5	CCV1	1	Water			2.00	03/30/2018	15:19
6	J2172-02	1	Water			5.68	03/30/2018	15:22
7	J2172-02DUP	1	Water			5.67	03/30/2018	15:23
8	J2172-03	1	Water			5.15	03/30/2018	15:24
9	J2172-04	1	Water			5.09	03/30/2018	15:25
10	J2172-05	1	Water			5.08	03/30/2018	15:26
11	J2172-06	1	Water			4.99	03/30/2018	15:27
12	J2172-07	1	Water			5.02	03/30/2018	15:28
13	J2172-08	1	Water			5.07	03/30/2018	15:29
14	J2172-09	1	Water			5.18	03/30/2018	15:30
15	J2172-11	1	Water			5.18	03/30/2018	15:31
16	CCV2	1	Water			12.00	03/30/2018	15:33
17	J2172-12	1	Water			5.97	03/30/2018	15:34
18	J2172-13	1	Water			6.17	03/30/2018	15:36
19	J2172-14	1	Water			6.16	03/30/2018	15:37
20	J2172-15	1	Water			5.90	03/30/2018	15:39

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	J2172-16	1	Water			6.20	03/30/2018	15:41
22	J2172-17	1	Water			7.44	03/30/2018	15:43
23	J2172-18	1	Water			7.84	03/30/2018	15:45
24	J2172-18DUP	1	Water			7.85	03/30/2018	15:47
25	CCV3	1	Water			2.00	03/30/2018	15:50

WORKLIST(Hardcopy Internal Chain)

1304282

WorkList Name : PH J2172

WorkList ID : 110255

Date : 3/30/2018 3:11:46 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/30/2018	Water	J2172-02 -1A	pH	Cool 4 deg C	PSEG01	J63	22043	03/30/2018	9040C
03/30/2018	Water	J2172-03 -1A	pH	Cool 4 deg C	PSEG01	J63	22044	03/30/2018	9040C
03/30/2018	Water	J2172-04 -1A	pH	Cool 4 deg C	PSEG01	J63	22045	03/30/2018	9040C
03/30/2018	Water	J2172-05 -1A	pH	Cool 4 deg C	PSEG01	J63	22132	03/30/2018	9040C
03/30/2018	Water	J2172-06 -1A	pH	Cool 4 deg C	PSEG01	J63	22133	03/30/2018	9040C
03/30/2018	Water	J2172-07 -1A	pH	Cool 4 deg C	PSEG01	J63	22134	03/30/2018	9040C
03/30/2018	Water	J2172-08 -1A	pH	Cool 4 deg C	PSEG01	J63	22178	03/30/2018	9040C
03/30/2018	Water	J2172-09 -1A	pH	Cool 4 deg C	PSEG01	J63	22180	03/30/2018	9040C
03/30/2018	Water	J2172-11 -1A	pH	Cool 4 deg C	PSEG01	J63	22181	03/30/2018	9040C
03/30/2018	Water	J2172-12 -1A	pH	Cool 4 deg C	PSEG01	J63	22182	03/30/2018	9040C
03/30/2018	Water	J2172-13 -1A	pH	Cool 4 deg C	PSEG01	J63	22183	03/30/2018	9040C
03/30/2018	Water	J2172-14 -1A	pH	Cool 4 deg C	PSEG01	J63	22184	03/30/2018	9040C
03/30/2018	Water	J2172-15 -1A	pH	Cool 4 deg C	PSEG01	J63	22185	03/30/2018	9040C
03/30/2018	Water	J2172-16 -1A	pH	Cool 4 deg C	PSEG01	J63	22186	03/30/2018	9040C
03/30/2018	Water	J2172-17 -1A	pH	Cool 4 deg C	PSEG01	J63	22187	03/30/2018	9040C
03/30/2018	Water	J2172-18 -1A	pH	Cool 4 deg C	PSEG01	J63	22188	03/30/2018	9040C

Date/Time 03/30/18 3:45 PM
 Received by: JP
 Relinquished by: CB

Date/Time 03/30/18 5:00 AM
 Received by: JP
 Relinquished by: CB