



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB94287

Ambient Barometric Pressure (mmHg): 755

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	82.3	03/30/2018	13:20
2	J2103-10		1	>100.0	>212.0	>212.0	03/30/2018	13:40
3	J2103-10DUP		1	>100.0	>212.0	>212.0	03/30/2018	14:02
4	J2106-03		1	>100.0	>212.0	>212.0	03/30/2018	14:32
5	J2106-04		1	>100.0	>212.0	>212.0	03/30/2018	14:50
6	J2106-05		1	>100.0	>212.0	>212.0	03/30/2018	15:15
7	J2106-06		1	>100.0	>212.0	>212.0	03/30/2018	15:45
8	J2119-09		1	>100.0	>212.0	>212.0	03/30/2018	16:10
9	J2119-10		1	>100.0	>212.0	>212.0	03/30/2018	16:32
10	J2131-10		1	>100.0	>212.0	>212.0	03/30/2018	16:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP2-33018

WorkList ID : 110249

Date : 3/30/2018 12:50:45 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/01/2018	Water	J2103-10	Flash Point	Cool 4 deg C	PSEG01	J53	LS	03/28/2018	1010A
04/04/2018	Water	J2106-03	Flash Point	Cool 4 deg C	PSEG01	J53	19619	03/28/2018	1010A
04/04/2018	Water	J2106-04	Flash Point	Cool 4 deg C	PSEG01	J53	19622	03/28/2018	1010A
04/04/2018	Water	J2106-05	Flash Point	Cool 4 deg C	PSEG01	J53	19606	03/28/2018	1010A
04/04/2018	Water	J2106-06	Flash Point	Cool 4 deg C	PSEG01	J53	19607	03/28/2018	1010A
04/04/2018	Water	J2119-09	Flash Point	Cool 4 deg C	PSEG01	J63	15284	03/28/2018	1010A
04/04/2018	Water	J2119-10	Flash Point	Cool 4 deg C	PSEG01	J63	15291	03/28/2018	1010A
04/04/2018	Water	J2131-10	Flash Point	Cool 4 deg C	PSEG01	J63	4507-4543	03/28/2018	1010A

Date/Time

Received by:

Relinquished by:

Date/Time

Received by:

Relinquished by: