



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB94290

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	03/30/2018	17:22
2	CAL2	1	Water			7.01	03/30/2018	17:23
3	CAL3	1	Water			10.02	03/30/2018	17:25
4	ICV	1	Water			7.00	03/30/2018	17:26
5	CCV1	1	Water			2.01	03/30/2018	17:28
6	J1757-08	1	Solid	20.02	20	7.85	03/30/2018	17:30
7	J1757-08DUP	1	Solid	20.03	20	7.86	03/30/2018	17:31
8	J1757-10	1	Solid	20.05	20	8.05	03/30/2018	17:32
9	J1757-12	1	Solid	20.02	20	9.32	03/30/2018	17:33
10	J2175-03	1	Solid	20.03	20	7.81	03/30/2018	17:34
11	J2178-04	1	Solid	20.02	20	7.27	03/30/2018	17:35
12	CCV2	1	Water			12.01	03/30/2018	17:36

1394290

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity j2162

WorkList ID : 110263

Date : 3/30/2018 4:58:24 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/30/2018	Solid	J1757-08	Corrosivity	Cool 4 deg C	PSEG05		EO-03-033018-A	03/30/2018	9045C
03/30/2018	Solid	J1757-10	Corrosivity	Cool 4 deg C	PSEG05		EO-03-033018-B	03/30/2018	9045C
03/30/2018	Solid	J1757-12	Corrosivity	Cool 4 deg C	PSEG05		EO-03-033018-C	03/30/2018	9045C
03/30/2018	Solid	J2175-03	Corrosivity	Cool 4 deg C	PSEG01	J63	TP-3	03/30/2018	9045C
03/30/2018	Solid	J2178-04	Corrosivity	Cool 4 deg C	PSEG01	J63	S6-TP	03/30/2018	9045C

Date/Time 03/30/18 5:00 PM
Received by: JH
Relinquished by: JH

Date/Time 03/30/18 7:00 PM
Received by: JH
Relinquished by: JH