

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB94292

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP61786
calibration std. phosphate 1 ppm	WP63186
calibration std. phosphate 0.5 ppm	WP63185
calibration std. phosphate 0.3 ppm	WP63184
calibration std. 0 ppm	WP63181
phosphate ICV std.	WP63187
phosphate CCV std.	WP63188
calibration std. phosphate 0.1 ppm	WP63183
calibration std. phosphate 0.05 ppm	WP63182
Combined reagent	WP63191
Phenolphthalein indicator	WP61777
Sodium hydroxide, 1N	WP61776

Intercept: -0.0009Slope: 0.6504Regression: 0.999829

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	03/30/2018	10:56
2	CAL2	0.05	1	50	50	0.033	03/30/2018	10:56
3	CAL3	0.10	1	50	50	0.065	03/30/2018	10:57
4	CAL4	0.30	1	50	50	0.186	03/30/2018	10:57
5	CAL5	0.50	1	50	50	0.330	03/30/2018	10:58
6	CAL6	1.00	1	50	50	0.649	03/30/2018	10:58

Analysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB94292

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.325	0.501	03/30/2018	10:59
2	ICB		1	50	50	0.000	0.001	03/30/2018	11:00
3	CCV1	0.50	1	50	50	0.328	0.506	03/30/2018	11:01
4	CCB1		1	50	50	0.000	0.001	03/30/2018	11:01
5	PB107839BL		1	50	50	0.000	0.001	03/30/2018	11:02
6	PB107839BS	0.50	1	50	50	0.324	0.500	03/30/2018	11:02
7	J2110-01		1	50	50	0.032	0.051	03/30/2018	11:03
8	J2110-01DUP		1	50	50	0.032	0.051	03/30/2018	11:03
9	J2110-01MS	0.50	1	50	50	0.350	0.540	03/30/2018	11:04
10	J2110-01MSD	0.50	1	50	50	0.349	0.538	03/30/2018	11:04
11	J2110-02		1	50	50	0.031	0.049	03/30/2018	11:05
12	J2130-01		1	50	50	2.149	3.306	03/30/2018	11:05
13	J2133-07		1	50	50	0.011	0.018	03/30/2018	11:06
14	J2133-08		1	50	50	0.045	0.071	03/30/2018	11:06
15	CCV2	0.50	1	50	50	0.328	0.506	03/30/2018	11:07
16	CCB2		1	50	50	0.000	0.001	03/30/2018	11:07
17	J2133-09		1	50	50	0.012	0.020	03/30/2018	11:08
18	J2130-01		10	50	50	0.589	0.907	03/30/2018	11:17
19	CCV3	0.50	1	50	50	0.327	0.504	03/30/2018	11:19
20	CCB3		1	50	50	0.000	0.001	03/30/2018	11:19