


7470A

LB94297
INSTRUMENT ID: CV 1

Linear

 μ Abs.:

35181



A= 0.0000e+000

B= 2.8355e-004

C= -1.2760e-002

Rho= 0.9999270

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	5	0.000	5				
0.2	0.200	0.197	-0.003	740	0.0 %	740				
2.5	2.500	2.535	0.035	8984	0.0 %	8984				
5.0	5.000	4.941	-0.059	17470	0.0 %	17470				
7.5	7.500	7.576	0.076	26762	0.0 %	26762				
10.0	10.000	9.963	-0.037	35181	0.0 %	35181				

LB94297 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 02 Apr 2018 09:52:37				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	5	-	0.00007470A	PPB		02 Apr 2018 09:53:10	S Std
0.2 - 1	50.2	740	-	0.20007470A	PPB		02 Apr 2018 09:55:19	S Std
2.5 - 1	52.5	8984	-	2.50007470A	PPB		02 Apr 2018 09:57:27	S Std
5.0 - 1	55	17470	-	5.00007470A	PPB		02 Apr 2018 09:59:35	S Std
7.5 - 1	57.5	26762	-	7.50007470A	PPB		02 Apr 2018 10:01:43	S Std
10.0 - 1	510	35181	-	10.00007470A	PPB		02 Apr 2018 10:03:51	S Std
ICV71 - 1	ICV71	13574	3.8362	-7470A	PPB		02 Apr 2018 10:06:39	U SMPL
ICB71 - 1	ICB71	-85	-0.0369	-7470A	PPB		02 Apr 2018 10:08:45	U SMPL
CCV19 - 1	CCV19	18370	5.1961	-7470A	PPB		02 Apr 2018 10:10:52	U SMPL
CCB19 - 1	CCB19	-124	-0.0479	-7470A	PPB		02 Apr 2018 10:12:59	U SMPL
CRA - 1	CRA	720	0.1914	-7470A	PPB		02 Apr 2018 10:15:06	U SMPL
HighStd - 1	HighStd	36440	10.3199	-7470A	PPB		02 Apr 2018 10:17:13	U SMPL
ChkStd - 1	ChkStd	26427	7.4807	-7470A	PPB		02 Apr 2018 10:19:20	U SMPL
PB107860BL - 1	PBW	-15	-0.0170	-7470A	PPB		02 Apr 2018 10:24:11	U SMPL
PB107860BS - 1	LCSW	13156	3.7177	-7470A	PPB		02 Apr 2018 10:26:18	U SMPL
PB107860TB - 1	PB107825T B	-91	-0.0386	-7470A	PPB		02 Apr 2018 10:28:26	U SMPL
J2142-02 - 1	GB-RT-2334	-5	-0.0142	-7470A	PPB		02 Apr 2018 10:30:34	U SMPL
J2142-02DUP - 1	GB-RT-2334DUP	18	-0.0077	-7470A	PPB		02 Apr 2018 10:32:43	U SMPL
J2142-02MS - 1	GB-RT-2334MS	13771	3.8921	-7470A	PPB		02 Apr 2018 10:34:51	U SMPL
J2142-02MSD - 1	GB-RT-2334MSD	13371	3.7786	-7470A	PPB		02 Apr 2018 10:36:58	U SMPL
CCV20 - 1	CCV20	18710	5.2925	-7470A	PPB		02 Apr 2018 10:39:04	U SMPL
CCB20 - 1	CCB20	-127	-0.0488	-7470A	PPB		02 Apr 2018 10:41:11	U SMPL
J2142-04 - 1	GB-RT-4632	6	-0.0111	-7470A	PPB		02 Apr 2018 10:43:19	U SMPL
J2142-06 - 1	GB-RT-3025	14	-0.0088	-7470A	PPB		02 Apr 2018 10:45:25	U SMPL
J2142-08 - 1	GB-RT-4252	37	-0.0023	-7470A	PPB		02 Apr 2018 10:47:32	U SMPL
J2142-10 - 1	GB-RT-2342	23	-0.0062	-7470A	PPB		02 Apr 2018 10:49:40	U SMPL
J2146-02 - 1	TP-CS	19	-0.0074	-7470A	PPB		02 Apr 2018 10:51:47	U SMPL
J2146-04 - 1	TP-CD	34	-0.0031	-7470A	PPB		02 Apr 2018 10:53:55	U SMPL
PB107861BL - 1	PBW	18	-0.0077	-7470A	PPB		02 Apr 2018 10:56:03	U SMPL
PB107861BS - 1	LCSW	12768	3.6077	-7470A	PPB		02 Apr 2018 10:58:12	U SMPL
PB107861TB - 1	PB107813TB	-87	-0.0374	-7470A	PPB		02 Apr 2018 11:00:20	U SMPL
J2121-21 - 1	PA-01-COMP	-13	-0.0164	-7470A	PPB		02 Apr 2018 11:02:27	U SMPL
CCV21 - 1	CCV21	18262	5.1655	-7470A	PPB		02 Apr 2018 11:04:34	U SMPL
CCB21 - 1	CCB21	-124	-0.0479	-7470A	PPB		02 Apr 2018 11:06:41	U SMPL
J2121-21DUP - 1	PA-01-COMP DUP	8	-0.0105	-7470A	PPB		02 Apr 2018 11:08:48	U SMPL
J2121-21MS - 1	PA-01-COMPMS	13699	3.8716	-7470A	PPB		02 Apr 2018 11:13:41	U SMPL
J2121-21MSD - 1	PA-01-COMPMSD	13432	3.7959	-7470A	PPB		02 Apr 2018 11:15:47	U SMPL
J2121-22 - 1	PA-02-COMP	-93	-0.0391	-7470A	PPB		02 Apr 2018 11:17:54	U SMPL
J2121-23 - 1	PA-03-COMP	22	-0.0065	-7470A	PPB		02 Apr 2018 11:20:02	U SMPL
J2121-24 - 1	PA-05-COMP	51	0.0017	-7470A	PPB		02 Apr 2018 11:22:10	U SMPL
J2121-25 - 1	PA-06-COMP	8	-0.0105	-7470A	PPB		02 Apr 2018 11:24:18	U SMPL
J2121-26 - 1	PA-07-COMP	12	-0.0094	-7470A	PPB		02 Apr 2018 11:26:26	U SMPL
J2121-27 - 1	PA-10-COMP	35	-0.0028	-7470A	PPB		02 Apr 2018 11:28:35	U SMPL
J2121-28 - 1	PA-04-COMP	72	0.0077	-7470A	PPB		02 Apr 2018 11:30:42	U SMPL
CCV22 - 1	CCV22	18051	5.1057	-7470A	PPB		02 Apr 2018 11:32:49	U SMPL
CCB22 - 1	CCB22	-156	-0.0570	-7470A	PPB		02 Apr 2018 11:34:57	U SMPL
J2121-29 - 1	PA-08-COMP	22	-0.0065	-7470A	PPB		02 Apr 2018 11:37:05	U SMPL
J2121-30 - 1	PA-09-COMP	14	-0.0088	-7470A	PPB		02 Apr 2018 11:39:12	U SMPL
J2122-13 - 1	PB-12-COMP	21	-0.0068	-7470A	PPB		02 Apr 2018 11:41:19	U SMPL
J2122-14 - 1	PB-11-COMP	16	-0.0082	-7470A	PPB		02 Apr 2018 11:43:26	U SMPL
J2122-15 - 1	PB-14-COMP	-14	-0.0167	-7470A	PPB		02 Apr 2018 11:45:33	U SMPL
J2122-16 - 1	PB-13-COMP	29	-0.0045	-7470A	PPB		02 Apr 2018 11:47:41	U SMPL
J2122-17 - 1	PB-16-COMP	67	0.0062	-7470A	PPB		02 Apr 2018 11:49:49	U SMPL
J2122-18 - 1	PB-15-COMP	29	-0.0045	-7470A	PPB		02 Apr 2018 11:51:57	U SMPL
PB107862BL - 1	PBW	8	-0.0105	-7470A	PPB		02 Apr 2018 11:54:05	U SMPL
PB107862BS - 1	LCSW	13744	3.8844	-7470A	PPB		02 Apr 2018 11:56:12	U SMPL
CCV23 - 1	CCV23	17893	5.0609	-7470A	PPB		02 Apr 2018 11:58:19	U SMPL
CCB23 - 1	CCB23	-147	-0.0544	-7470A	PPB		02 Apr 2018 12:00:26	U SMPL
PB107862TB - 1	PB107823TB	-4	-0.0139	-7470A	PPB		02 Apr 2018 12:02:34	U SMPL
J2106-03 - 1	19619	20	-0.0071	-7470A	PPB		02 Apr 2018 12:04:41	U SMPL
J2106-03DUP - 1	19619DUP	19	-0.0074	-7470A	PPB		02 Apr 2018 12:06:48	U SMPL
J2106-03MS - 1	19619MS	13641	3.8552	-7470A	PPB		02 Apr 2018 12:08:56	U SMPL
J2106-03MSD - 1	19619MSD	13406	3.7886	-7470A	PPB		02 Apr 2018 12:11:03	U SMPL
J2106-04 - 1	19622	-90	-0.0383	-7470A	PPB		02 Apr 2018 12:13:11	U SMPL
J2106-05 - 1	19606	-32	-0.0218	-7470A	PPB		02 Apr 2018 12:15:20	U SMPL
J2119-09 - 1	15284	24	-0.0060	-7470A	PPB		02 Apr 2018 12:17:28	U SMPL
J2119-10 - 1	15291	51	0.0017	-7470A	PPB		02 Apr 2018 12:19:36	U SMPL
J2131-10 - 1	4507-4543	11	-0.0096	-7470A	PPB		02 Apr 2018 12:21:43	U SMPL
CCV24 - 1	CCV24	17784	5.0300	-7470A	PPB		02 Apr 2018 12:23:51	U SMPL
CCB24 - 1	CCB24	-188	-0.0661	-7470A	PPB		02 Apr 2018 12:25:58	U SMPL
PB107866BL - 1	PBW	-25	-0.0198	-7470A	PPB		02 Apr 2018 12:28:06	U SMPL
PB107866BS - 1	LCSW	13048	3.6871	-7470A	PPB		02 Apr 2018 12:30:13	U SMPL
PB107866TB - 1	PB107821TB	-52	-0.0275	-7470A	PPB		02 Apr 2018 12:32:21	U SMPL
J1752-04 - 1	NB-01-032918-A	26	-0.0054	-7470A	PPB		02 Apr 2018 12:34:28	U SMPL
J1752-06 - 1	NB-01-032918-B	29	-0.0045	-7470A	PPB		02 Apr 2018 12:36:35	U SMPL
J1752-08 - 1	NB-01-032918-C	123	0.0221	-7470A	PPB		02 Apr 2018 12:38:43	U SMPL
J1760-10 - 1	SU-03-032918	35	-0.0028	-7470A	PPB		02 Apr 2018 12:40:50	U SMPL
J1760-10DUP - 1	SU-03-032918DUP	26	-0.0054	-7470A	PPB		02 Apr 2018 12:42:59	U SMPL
J1760-10MS - 1	SU-03-032918MS	13554	3.8305	-7470A	PPB		02 Apr 2018 12:45:07	U SMPL
J1760-10MSD - 1	SU-03-032918MSD	13247	3.7435	-7470A	PPB		02 Apr 2018 12:47:15	U SMPL
CCV25 - 1	CCV25	17798	5.0339	-7470A	PPB		02 Apr 2018 12:49:22	U SMPL