



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB94303

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP63276
Residual chlorine std, 0PPM	WP63272
Residual chlorine std, 0.1PPM	WP63273
Residual chlorine LOQ, 0.2PPM	WP63280
Residual chlorine std, 0.2PPM	WP63274
Residual chlorine std, and CCV , 0.4PPM	WP63275
Residual chlorine std, 1.6PPM	WP63277
Residual chlorine ICV-LCS, 0.4PPM	WP63278
Residual chlorine LOD, 0.05PPM	WP63279
Total Chlorine Powder Pillows	W1960

Intercept: 0.0029

Slope: 0.9977

Regression: 0.999980

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	04/02/2018	11:10
2	CAL2	0.1	1	0.000	0.110	0.110	04/02/2018	11:10
3	CAL3	0.2	1	0.000	0.200	0.200	04/02/2018	11:11
4	CAL4	0.4	1	0.000	0.400	0.400	04/02/2018	11:11
5	CAL5	0.8	1	0.000	0.800	0.800	04/02/2018	11:12
6	CAL6	1.6	1	0.000	1.600	1.600	04/02/2018	11:12



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB94303

Seq	Lab ID	Initial Weight	Final Vol	True Value	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.000	0.390	0.390	0.390	04/02/2018	11:13
2	ICB				1	0.000	0.000	0.000	0.000	04/02/2018	11:13
3	CCV1			0.4	1	0.000	0.410	0.410	0.410	04/02/2018	11:14
4	CCB1				1	0.000	0.000	0.000	0.000	04/02/2018	11:14
5	LB94303BL	50	50		1	0.000	0.000	0.000	0.000	04/02/2018	11:15
6	LB94303BS	50	50	0.4	1	0.000	0.390	0.390	0.390	04/02/2018	11:15
7	J2133-07	50	50		1	0.000	0.050	0.050	0.050	04/02/2018	11:16
8	J2133-08	50	50		1	0.000	0.210	0.210	0.208	04/02/2018	11:16
9	J2133-09	50	50		1	0.000	0.060	0.060	0.060	04/02/2018	11:17
10	CCV2			0.4	1	0.000	0.410	0.410	0.410	04/02/2018	11:17
11	CCB2				1	0.000	0.000	0.000	0.000	04/02/2018	11:18